

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)



Laboratory Report

for

Arizona Department of Environmental Quality
1110 West Washington Street
Phoenix, AZ 85007
Attention: David Burchard

Date of Issue
06/14/2017



Eurofins Eaton
Analytical

TDF: Thomas.D.French
Project Manager

Report:665876
Project:CHINLE-UD24
ADHS License #:AZ0778
Group:Many Farms Elementary School
PO#:PO#: ADEQ16-116686:3



* Accredited in accordance with TNI 2009 and ISO/IEC 17025:2005.

* Laboratory certifies that the test results meet all **TNI 2009 and ISO/IEC 17025:2005** requirements unless noted under the individual analysis.

* Following the cover page are State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.

* Test results relate only to the sample(s) tested.

* This report shall not be reproduced except in full, without the written approval of the laboratory.

STATE CERTIFICATION LIST

| State | Certification Number | State | Certification Number |
|---------------------------------------|----------------------|---|----------------------|
| Alabama | 41060 | Mississippi | Certified |
| Arizona | AZ0778 | Montana | Cert 0035 |
| Arkansas | Certified | Nebraska | Certified |
| California-Monrovia-ELAP | 2813 | Nevada | CA00006-2016 |
| California-Colton- ELAP | 2812 | New Hampshire * | 2959 |
| California-Folsom- ELAP | 2820 | New Jersey * | CA 008 |
| California-Fresno- ELAP | 2966 | New Mexico | Certified |
| Colorado | Certified | New York * | 11320 |
| Connecticut | PH-0107 | North Carolina | 06701 |
| Delaware | CA 006 | North Dakota | R-009 |
| Florida * | E871024 | Oregon (Primary AB) * | ORELAP 4034 |
| Georgia | 947 | Pennsylvania * | 68-565 |
| Guam | 16-003r | Puerto Rico | Certified |
| Hawaii | Certified | Rhode Island | LAO00326 |
| Idaho | Certified | South Carolina | 87016 |
| Illinois * | 200033 | South Dakota | Certified |
| Indiana | C-CA-01 | Tennessee | TN02839 |
| Kansas * | E-10268 | Texas * | T104704230-15-9 |
| Kentucky | 90107 | Utah * | CA000062016-10 |
| Louisiana * | LA16003 | Vermont | VT0114 |
| Maine | CA0006 | Virginia * | 460260 |
| Maryland | 224 | Washington | C838 |
| Commonwealth of Northern Marianas Is. | MP0004 | Wyoming | Certified |
| Massachusetts | M-CA006 | EPA Region 5 | Certified |
| Michigan | 9906 | Los Angeles County Sanitation Districts | 10264 |

* NELAP/TNI Recognized Accreditation Bodies

ISO 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO 17025 as verified by the ANSI-ASQ National Accreditation Board/ANAB.

Refer to Certificate and scope of accreditation (AT 1807) found at: <http://www.eatonanalytical.com>

| SPECIFIC TESTS | METHOD OR TECHNIQUE USED | Environmental (Drinking Water) | Environmental (Waste Water) | Water as a Component of Food and Bev/Bev/ Bottled Water |
|--|----------------------------|--------------------------------|-----------------------------|---|
| 1,4-Dioxane | EPA 522 | x | | x |
| 2,3,7,8-TCDD | Modified EPA 1613B | x | | x |
| Acrylamide | In House Method (2440) | x | | x |
| Alkalinity | SM 2320B | x | x | x |
| Ammonia | EPA 350.1 | | x | x |
| Ammonia | SM 4500-NH3 H | | x | x |
| Anions and DBPs by IC | EPA 300.0 | x | x | x |
| Anions and DBPs by IC | EPA 300.1 | x | | x |
| Asbestos | EPA 100.2 | x | x | |
| Bicarbonate Alkalinity as HCO ₃ | SM 2320B | x | x | x |
| BOD / CBOD | SM 5210B | | x | x |
| Bromate | In House Method (2447) | x | | x |
| Carbamates | EPA 531.2 | x | | x |
| Carbonate as CO ₃ | SM 2330B | x | x | x |
| Carbonyls | EPA 556 | x | | x |
| COD | EPA 410.4 / SM 5220D | | x | |
| Chloramines | SM 4500-CL G | x | x | x |
| Chlorinated Acids | EPA 515.4 | x | | x |
| Chlorinated Acids | EPA 555 | x | | x |
| Chlorine Dioxide | SM 4500-CLO ₂ D | x | | x |
| Chlorine -Total/Free/ Combined Residual | SM 4500-Cl G | x | x | x |
| Conductivity | EPA 120.1 | | x | |
| Conductivity | SM 2510B | x | x | x |
| Corrosivity (Langelier Index) | SM 2330B | x | | x |
| Cryptosporidium | EPA 1623 | x | | x |
| Cyanide, Amenable | SM 4500-CN G | x | x | |
| Cyanide, Free | SM 4500CN F | x | x | x |
| Cyanide, Total | EPA 335.4 | x | x | x |
| Cyanogen Chloride (screen) | In House Method (2470) | x | | x |
| Diquat and Paraquat | EPA 549.2 | x | | x |
| DBP/HAA | SM 6251B | x | | x |
| Dissolved Oxygen | SM 4500-O G | | x | x |
| DOC | SM 5310C | x | | x |
| E. Coli (MTF/EC+MUG) | | x | | x |
| E. Coli | CFR 141.21(f)(6)(i) | x | | x |
| E. Coli | SM 9223 | | x | |
| E. Coli (Enumeration) | SM 9221B.1/ SM 9221F | x | | x |
| E. Coli (Enumeration) | SM 9223B | x | | x |
| EDB/DCBP | EPA 504.1 | x | | |
| EDB/DCBP and DBP | EPA 551.1 | x | | x |
| EDTA and NTA | In House Method (2454) | x | | x |
| Endothall | EPA 548.1 | x | | x |
| Endothall | In-house Method (2445) | x | | x |
| Enterococci | SM 9230B | x | x | |
| Fecal Coliform | SM 9221 E (MTF/EC) | x | | |
| Fecal Coliform | SM 9221C, E (MTF/EC) | | x | |
| Fecal Coliform (Enumeration) | SM 9221E (MTF/EC) | x | | x |
| Fecal Coliform with Chlorine Present | SM 9221E | | x | |
| Fecal Streptococci | SM 9230B | x | x | |
| Fluoride | SM 4500-F C | x | x | x |
| Giardia | EPA 1623 | x | | x |
| Glyphosate | EPA 547 | x | | x |
| Gross Alpha/Beta | EPA 900.0 | x | x | x |
| Gross Alpha Coprecipitation | SM 7110 C | x | x | x |
| Hardness | SM 2340B | x | x | x |
| Heterotrophic Bacteria | In House Method (2439) | x | | x |
| Heterotrophic Bacteria | SM 9215 B | x | | x |
| Hexavalent Chromium | EPA 218.6 | x | x | x |

| SPECIFIC TESTS | METHOD OR TECHNIQUE USED | Environmental (Drinking Water) | Environmental (Waste Water) | Water as a Component of Food and Bev/Bev/ Bottled Water |
|---|--|--------------------------------|-----------------------------|---|
| Hexavalent Chromium | EPA 218.7 | x | | x |
| Hexavalent Chromium | SM 3500-Cr B | | x | |
| Hormones | EPA 539 | x | | x |
| Hydroxide as OH Calc. | SM 2330B | x | | x |
| Kjeldahl Nitrogen | EPA 351.2 | | x | |
| Legionella | CDC Legionella | x | | x |
| Mercury | EPA 245.1 | x | x | x |
| Metals | EPA 200.7 / 200.8 | x | x | x |
| Microcystin LR | ELISA (2360) | x | | x |
| NDMA | EPA 521 | x | | x |
| NDMA | TQ In house method based on EPA 521 (2425) | x | | x |
| Nitrate/Nitrite Nitrogen | EPA 353.2 | x | x | x |
| OCL, Pesticides/PCB | EPA 505 | x | | x |
| Ortho Phosphate | EPA 365.1 | x | x | x |
| Ortho Phosphate | SM 4500P E | | | x |
| Ortho Phosphorous | SM 4500P E | x | | |
| Oxyhalides Disinfection Byproducts | EPA 317.0 | x | | x |
| Perchlorate | EPA 331.0 | x | | x |
| Perchlorate (low and high) | EPA 314.0 | x | | x |
| Perfluorinated Alkyl Acids | EPA 537 | x | | x |
| pH | EPA 150.1 | x | | |
| pH | SM 4500-H+B | x | x | x |
| Phenylurea Pesticides/ Herbicides | In House Method, based on EPA 532 (2448) | x | | x |
| Pseudomonas | IDEXX Pseudalert (2461) | x | | x |
| Radium-226 | GA Institute of Tech | x | | x |
| Radium-228 | GA Institute of Tech | x | | x |
| Radon-222 | SM 7500RN | x | | x |
| Residue, Filterable | SM 2540C | x | x | x |
| Residue, Non-filterable | SM 2540D | | x | |
| Residue, Total | SM 2540B | | x | x |
| Residue, Volatile | EPA 160.4 | | x | |
| Semi-VOC | EPA 525.2 | x | | x |
| Semi-VOC | EPA 625 | | x | x |
| Silica | SM 4500-Si D | x | x | |
| Silica | SM 4500-SiO ₂ C | x | x | |
| Sulfide | SM 4500-S ⁻ D | | x | |
| Sulfite | SM 4500-SO ₃ B | x | x | x |
| Surfactants | SM 5540C | x | x | x |
| Taste and Odor Analytes | SM 6040E | x | | x |
| Total Coliform (P/A) | SM 9221 A, B | x | | x |
| Total Coliform (Enumeration) | SM 9221 A, B, C | x | | x |
| Total Coliform / E. coli | Colisure SM 9223 | x | | x |
| Total Coliform | SM 9221B | | x | |
| Total Coliform with Chlorine Present | SM 9221B | | x | |
| Total Coliform / E.coli (P/A and Enumeration) | SM 9223 | x | | x |
| TOC | SM 5310C | x | x | x |
| TOX | SM 5320B | | x | |
| Total Phenols | EPA 420.1 | | x | |
| Total Phenols | EPA 420.4 | x | x | x |
| Total Phosphorous | SM 4500 P E | | x | |
| Turbidity | EPA 180.1 | x | x | x |
| Turbidity | SM 2130B | x | x | |
| Uranium by ICP/MS | EPA 200.8 | x | | x |
| UV 254 | SM 5910B | x | | |
| VOC | EPA 524.2/EPA 524.3 | x | | x |
| VOC | EPA 624 | | x | x |
| VOC | EPA SW 846 8260 | x | | x |
| VOC | In House Method (2411) | x | | x |
| Yeast and Mold | SM 9610 | x | | x |

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 665876
Project: CHINLE-UD24
Sample Group: Many Farms Elementary School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Dwayne O'Daniel

The following samples were received from you on **June 09, 2017 at 1041**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

| Sample # | Sample ID | Sample Date |
|---------------------|---|--------------------|
| <u>201706120442</u> | 005687 | 06/06/2017 0630 |
| | Sample Type: Fountain Facility ID: New Gym Sample Point ID: East Foyer | |
| | @ICPMS | Freight - Outbound |
| | | Freight - Return |
| <u>201706120443</u> | 005635 | 06/06/2017 0630 |
| | Sample Type: Tap Facility ID: Multi-Purpose (Kitchen) Sample Point ID: 2-Comp Sink | |
| | @ICPMS | |
| <u>201706120444</u> | 005648 | 06/06/2017 0630 |
| | Sample Type: Fountain Facility ID: Building B Sample Point ID: Corridor | |
| | @ICPMS | |
| <u>201706120445</u> | 005716 | 06/06/2017 0630 |
| | Sample Type: Fountain Facility ID: Building C Sample Point ID: Corridor | |
| | @ICPMS | |
| <u>201706120446</u> | 005644 | 06/06/2017 0630 |
| | Sample Type: Fountain Facility ID: Hollistic Sample Point ID: Corridor | |
| | @ICPMS | |
| <u>201706120447</u> | 005627 | 06/06/2017 0630 |
| | Sample Type: Tap Facility ID: Old Main (Jr. High) Sample Point ID: Rm A-113 | |
| | @ICPMS | |
| <u>201706120448</u> | 005649 | 06/06/2017 0630 |
| | Sample Type: Drinking Fountain Facility ID: Old Main (Jr. High) Sample Point ID: Corridor | |
| | @ICPMS | |
| <u>201706120449</u> | 005623 | 06/06/2017 0630 |

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 665876
Project: CHINLE-UD24
Sample Group: Many Farms Elementary School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Dwayne O'Daniel

The following samples were received from you on **June 09, 2017 at 1041**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

| Sample # | Sample ID | Sample Date |
|---------------------|---|-----------------|
| | Sample Type: Tap Facility ID: Pump House Sample Point ID: Main @ICPMS | |
| <u>201706120450</u> | 005632 | 06/06/2017 0630 |
| | Sample Type: Fountain Facility ID: New Gym Sample Point ID: Gym @ICPMS | |

Test Description

@ICPMS -- ICPMS Metals

665876

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | New Gym |
| Type of Fixture (tap, drinking fountain etc.) | Fountain East Foyer |
| Location of Fixture (example, room number) | East Foyer |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005687 New Gym E. |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 a.m. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|-------------|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature | |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | Multi-Purpose (Kitchen) |
| Type of Fixture (tap, drinking fountain etc.) | Tap |
| Location of Fixture (example, room number) | 2-comp sink |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005635 MPR |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 am. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|---|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|--------------------------------|
| Name of School District | Chinle Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | Building "B" |
| Type of Fixture (tap, drinking fountain etc.) | Fountain |
| Location of Fixture (example, room number) | Corridor |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 0056418 B |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|---|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Manly Farms Elementary School |
| Building (name/number) | Building "C" |
| Type of Fixture (tap, drinking fountain etc.) | Fountain |
| Location of Fixture (example, room number) | Corridor |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005716 C-1 |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 am, |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| For Lab use only | |
|---|---|
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|--------------------------------|
| Name of School District | Chinle Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | Hollistic |
| Type of Fixture (tap, drinking fountain etc.) | Fountain |
| Location of Fixture (example, room number) | Corridor |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005644 Hollistic |
| Date of Collection | 6-6-2017 |
| Time of Collection | 8:30 am. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|---|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

ADEQ Public School Drinking Water Lead Screening Program
Sampling Plan & Collection Log

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Mary Farms Elementary School |
| Building (name/number) | Old Main (Jr. High) |
| Type of Fixture (tap, drinking fountain etc.) | Tap A-2 |
| Location of Fixture (example, room number) | Rm. A-113 |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005627 |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 am. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| For Lab use only | |
|---|---|
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-----------------------------------|
| Name of School District | Chino Unified School District #24 |
| School Name | Many Farms Elementary School |
| Building (name/number) | Old Main (Jr. High) |
| Type of Fixture (tap, drinking fountain etc.) | Drinking Fountain |
| Location of Fixture (example, room number) | Corridor |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005649 A-1 |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 A.M. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|---|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

TRK# **7324 9820 5528**

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | Pump house |
| Type of Fixture (tap, drinking fountain etc.) | Tap |
| Location of Fixture (example, room number) | Main |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005632 005623 Pump house |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 a.m. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |

Notes Sample collector:

| | |
|---|-------------|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature | |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

| | |
|--|-------------------------------|
| Name of School District | Chino Unified School District |
| School Name | Many Farms Elementary School |
| Building (name/number) | New Gym |
| Type of Fixture (tap, drinking fountain etc.) | Fountain |
| Location of Fixture (example, room number) | Gym |
| Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>) | 005632 New Gym-1 |
| Date of Collection | 6-6-2017 |
| Time of Collection | 6:30 am. |
| Printed Name of Sample Collector | Dwayne O'Daniel |
| Signature Sample Collector | Dwayne O'Daniel |


Notes Sample collector:

| | |
|---|---|
| For Lab use only | |
| Analyze this drinking water sample for lead | |
| Date and Time Lab received | 6/9/17 1041 |
| Signature |  |
| Notes: | |

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number:

665876

SAMPLE TEMP RECEIVED:

SAMPLES REC'D DAY OF COLLECTION? ☐

IR Gun ID = 561A (Observation = 23.0 °C) (Corr. Factor = -0.2 °C) (Final = 22.8 °C)

CONDITION OF ICE: Frozen Partially Frozen Thawed N/A
 TYPE OF ICE: Real Synthetic No Ice ON TRAC DCS

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx UPS / DHL / Area Fast / Top Line / Other:

Compliance Acceptance Criteria:

- 1) Chemistry: >0; ≤ 6°C, not frozen (NELAP) (If received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen. (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

| | |
|---|---|
| 1 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C) | 2 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C) |
| 3 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C) | 4 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C) |

- 4) UCMR3: 524.3: (Observation = °C) (Corr. Factor = °C) (Final = °C)
 (non-GLEC)
 522: (Observation = °C) (Corr. Factor = °C) (Final = °C)

≤ 10°C if received within 48 hours of sample collection (not the same business day); ≤ 6°C if received after 48 hours of sample collection. Measure temperature for each method above.

- 5) LT2: Giardia/Cryptosporidium: < 20 °C, not frozen (received after 8 hours of sample collection)

E. Coll: < 10°C, not frozen (if received after 2 hours of sample collection)

Giardia/Crypto: (Observation = °C) (Corr. Factor = °C) (Final = °C)

E. Coll: (Observation = °C) (Corr. Factor = °C) (Final = °C)

- 6) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

Note: If samples are out of temperature range, let the ASMR know, ASMR will determine whether to proceed with analysis or not

| | | | | |
|-----------|------------|---------------------------|------|------|
| SIGNATURE | PRINT NAME | COMPANY/TITLE | DATE | TIME |
| | | Eurofins Eaton Analytical | | 1041 |

RECEIVED BY:

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/09/2017 1041

| Analyzed | Analyte | Sample ID | Result | Federal MCL | Units | MRL |
|------------------|--------------------|---|--------|-------------|-------|-----|
| 06/14/2017 12:06 | Lead Total ICAP/MS | 201706120443 <u>005635</u> | 9.5 | 15 | ug/L | 0.5 |
| 06/14/2017 12:11 | Lead Total ICAP/MS | 201706120447 <u>005627</u> | 3.0 | 15 | ug/L | 0.5 |
| 06/14/2017 12:13 | Lead Total ICAP/MS | 201706120449 <u>005623</u> | 2.8 | 15 | ug/L | 0.5 |

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/09/2017 1041

| Prepped | Analyzed | Prep Batch | Analytical Batch | Method | Analyte | Result | Units | MRL | Dilution |
|---|----------------|------------|------------------|-------------|--------------------|-----------------------------------|-------|-----|----------|
| 005687 (201706120442) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Fountain Facility ID: New Gym Sample Point ID: East Foyer | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:06 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |
| 005635 (201706120443) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Tap Facility ID: Multi-Purpose (Kitchen) Sample Point ID: 2-Comp Sink | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:06 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | 9.5 | ug/L | 0.5 | 1 |
| 005648 (201706120444) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Fountain Facility ID: Building B Sample Point ID: Corridor | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:07 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |
| 005716 (201706120445) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Fountain Facility ID: Building C Sample Point ID: Corridor | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:10 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |
| 005644 (201706120446) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Fountain Facility ID: Hollistic Sample Point ID: Corridor | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:10 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |
| 005627 (201706120447) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Tap Facility ID: Old Main (Jr. High) Sample Point ID: Rm A-113 | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:11 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | 3.0 | ug/L | 0.5 | 1 |
| 005649 (201706120448) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Drinking Fountain Facility ID: Old Main (Jr. High) Sample Point ID: Corridor | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:12 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |
| 005623 (201706120449) | | | | | | Sampled on 06/06/2017 0630 | | | |

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/09/2017 1041

| Prepped | Analyzed | Prep Batch | Analytical Batch | Method | Analyte | Result | Units | MRL | Dilution |
|--------------------------|----------------|------------|------------------|-------------|--------------------|----------------------------|-------|-----|----------|
| Sample Type: Tap | | | | | | | | | |
| Facility ID: Pump House | | | | | | | | | |
| Sample Point ID: Main | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:13 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | 2.8 | ug/L | 0.5 | 1 |
| 005632 (201706120450) | | | | | | Sampled on 06/06/2017 0630 | | | |
| Sample Type: Fountain | | | | | | | | | |
| Facility ID: New Gym | | | | | | | | | |
| Sample Point ID: Gym | | | | | | | | | |
| EPA 200.8 - ICPMS Metals | | | | | | | | | |
| 06/12/17 | 06/14/17 12:13 | 1002281 | 1002890 | (EPA 200.8) | Lead Total ICAP/MS | ND | ug/L | 0.5 | 1 |

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory QC Summary

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality

ICPMS Metals

Prep Batch: 1002281 Analytical Batch: 1002890

| | |
|--------------|--------|
| 201706120442 | 005687 |
| 201706120443 | 005635 |
| 201706120444 | 005648 |
| 201706120445 | 005716 |
| 201706120446 | 005644 |
| 201706120447 | 005627 |
| 201706120448 | 005649 |
| 201706120449 | 005623 |
| 201706120450 | 005632 |

Analysis Date: 06/14/2017

Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD
Analyzed by: RPD

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Report: 665876
Project: CHINLE-UD24
Group: Many Farms Elementary School

Arizona Department of Environmental Quality

| QC Type | Analyte | Native | Spiked | Recovered | Units | Yield (%) | Limits (%) | RPDLimit (%) | RPD% |
|----------------------------------|--------------------|--------|--------|-----------|----------------------------------|-----------|------------|--------------|------|
| ICPMS Metals by EPA 200.8 | | | | | | | | | |
| Analytical Batch: 1002890 | | | | | Analysis Date: 06/14/2017 | | | | |
| LCS1 | Lead Total ICAP/MS | | 20 | 19.2 | ug/L | 96 | (85-115) | | |
| LCS2 | Lead Total ICAP/MS | | 20 | 19.2 | ug/L | 96 | (85-115) | 20 | 0.0 |
| MBLK | Lead Total ICAP/MS | | | <0.25 | ug/L | | | | |
| MRL_CHK | Lead Total ICAP/MS | | 0.5 | 0.520 | ug/L | 104 | (50-150) | | |
| MS_201706120389 | Lead Total ICAP/MS | ND | 20 | 20.8 | ug/L | 104 | (70-130) | | |
| MS2_201706120529 | Lead Total ICAP/MS | ND | 20 | 20.6 | ug/L | 103 | (70-130) | | |
| MSD_201706120389 | Lead Total ICAP/MS | ND | 20 | 20.8 | ug/L | 104 | (70-130) | 20 | 0.0 |
| MSD2_201706120529 | Lead Total ICAP/MS | ND | 20 | 21.0 | ug/L | 105 | (70-130) | 20 | 1.4 |

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.