

# TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

## ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Phoenix

4625 East Cotton Ctr Blvd

Suite 189

Phoenix, AZ 85040

Tel: (602)437-3340

TestAmerica Job ID: 550-80679-1

TestAmerica Sample Delivery Group: Prescott High School

Client Project/Site: Pb in DW / Prescott Unified District

Revision: 1

For:

Arizona Dept of Environmental Quality

1110 W. Washington St., MC5415B-1

Phoenix, Arizona 85007

Attn: David A Burchard



Authorized for release by:

4/24/2017 4:18:13 PM

Susan Armijo, Project Manager II

(602)659-7664

[susan.armijo@testamericainc.com](mailto:susan.armijo@testamericainc.com)

### LINKS

Review your project  
results through

TotalAccess

Have a Question?



Visit us at:

[www.testamericainc.com](http://www.testamericainc.com)

*This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.*

*Results relate only to the items tested and the sample(s) as received by the laboratory.*

# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Sample Summary . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	8
QC Sample Results . . . . .	11
QC Association Summary . . . . .	13
Lab Chronicle . . . . .	15
Certification Summary . . . . .	19
Method Summary . . . . .	20
Chain of Custody . . . . .	21
Receipt Checklists . . . . .	42



## Definitions/Glossary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

### Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

## Case Narrative

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

**Job ID: 550-80679-1**

**Laboratory: TestAmerica Phoenix**

### Narrative

#### Job Narrative 550-80679-1

### Comments

No additional comments.

### Receipt

The samples were received on 4/10/2017 11:30 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 21.3° C.

### Receipt Exceptions

The following samples for metals were received unpreserved and were preserved upon receipt to the laboratory: 1001-Main (550-80679-1), 1001-Main (550-80679-2), 1002-Math (550-80679-3), 1002-Math (550-80679-4), 1003-Gym A (550-80679-5), 1004-Gym B (550-80679-6), 1005-W/R (550-80679-7), 1005-W/R (550-80679-8), 1006-Fine Arts (550-80679-9), 1007-VOC (550-80679-10), 1007-VOC (550-80679-11), 1008-Portables (550-80679-12), 1008-Portables (550-80679-13), 1009-JTED (550-80679-14), 1009-JTED (550-80679-15), 1011-MPR (550-80679-16), 1012-Book (550-80679-17), 1013-Music (550-80679-18), 1014-New R/R (550-80679-19), 1003-Gym A (550-80679-20) and 1004-Gym B (550-80679-21). Regulatory documents require a 24-hour waiting period from the time of the addition of the acid preservative to the time of digestion.

### Metals

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

## Sample Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
550-80679-1	1001-Main	Drinking Water	04/05/17 07:01	04/10/17 11:30
550-80679-2	1001-Main	Drinking Water	04/05/17 07:02	04/10/17 11:30
550-80679-3	1002-Math	Drinking Water	04/05/17 07:05	04/10/17 11:30
550-80679-4	1002-Math	Drinking Water	04/05/17 07:06	04/10/17 11:30
550-80679-5	1003-Gym A	Drinking Water	04/05/17 07:08	04/10/17 11:30
550-80679-6	1004-Gym B	Drinking Water	04/05/17 07:12	04/10/17 11:30
550-80679-7	1005-W/R	Drinking Water	04/05/17 07:18	04/10/17 11:30
550-80679-8	1005-W/R	Drinking Water	04/05/17 07:20	04/10/17 11:30
550-80679-9	1006-Fine Arts	Drinking Water	04/05/17 06:55	04/10/17 11:30
550-80679-10	1007-VOC	Drinking Water	04/05/17 07:30	04/10/17 11:30
550-80679-11	1007-VOC	Drinking Water	04/05/17 07:31	04/10/17 11:30
550-80679-12	1008-Portables	Drinking Water	04/05/17 07:25	04/10/17 11:30
550-80679-13	1008-Portables	Drinking Water	04/05/17 07:26	04/10/17 11:30
550-80679-14	1009-JTED	Drinking Water	04/05/17 07:34	04/10/17 11:30
550-80679-15	1009-JTED	Drinking Water	04/05/17 07:35	04/10/17 11:30
550-80679-16	1011-MPR	Drinking Water	04/05/17 06:46	04/10/17 11:30
550-80679-17	1012-Book	Drinking Water	04/05/17 06:57	04/10/17 11:30
550-80679-18	1013-Music	Drinking Water	04/05/17 06:58	04/10/17 11:30
550-80679-19	1014-New R/R	Drinking Water	04/05/17 06:54	04/10/17 11:30
550-80679-20	1003-Gym A	Drinking Water	04/05/17 07:13	04/10/17 11:30
550-80679-21	1004-Gym B	Drinking Water	04/05/17 07:12	04/10/17 11:30

## Detection Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

**Client Sample ID: 1001-Main**

**Lab Sample ID: 550-80679-1**

☐ No Detections.

**Client Sample ID: 1001-Main**

**Lab Sample ID: 550-80679-2**

☐ No Detections.

**Client Sample ID: 1002-Math**

**Lab Sample ID: 550-80679-3**

☐ No Detections.

**Client Sample ID: 1002-Math**

**Lab Sample ID: 550-80679-4**

☐ No Detections.

**Client Sample ID: 1003-Gym A**

**Lab Sample ID: 550-80679-5**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.68		0.50	ppb	1		200.8	Total/NA

**Client Sample ID: 1004-Gym B**

**Lab Sample ID: 550-80679-6**

☐ No Detections.

**Client Sample ID: 1005-W/R**

**Lab Sample ID: 550-80679-7**

☐ No Detections.

**Client Sample ID: 1005-W/R**

**Lab Sample ID: 550-80679-8**

☐ No Detections.

**Client Sample ID: 1006-Fine Arts**

**Lab Sample ID: 550-80679-9**

☐ No Detections.

**Client Sample ID: 1007-VOC**

**Lab Sample ID: 550-80679-10**

☐ No Detections.

**Client Sample ID: 1007-VOC**

**Lab Sample ID: 550-80679-11**

☐ No Detections.

**Client Sample ID: 1008-Portables**

**Lab Sample ID: 550-80679-12**

☐ No Detections.

**Client Sample ID: 1008-Portables**

**Lab Sample ID: 550-80679-13**

☐ No Detections.

**Client Sample ID: 1009-JTED**

**Lab Sample ID: 550-80679-14**

☐ No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

## Detection Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

### Client Sample ID: 1009-JTED

### Lab Sample ID: 550-80679-15

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.59		0.50	ppb	1		200.8	Total/NA

### Client Sample ID: 1011-MPR

### Lab Sample ID: 550-80679-16

No Detections.

### Client Sample ID: 1012-Book

### Lab Sample ID: 550-80679-17

No Detections.

### Client Sample ID: 1013-Music

### Lab Sample ID: 550-80679-18

No Detections.

### Client Sample ID: 1014-New R/R

### Lab Sample ID: 550-80679-19

No Detections.

### Client Sample ID: 1003-Gym A

### Lab Sample ID: 550-80679-20

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.0		0.50	ppb	1		200.8	Total/NA

### Client Sample ID: 1004-Gym B

### Lab Sample ID: 550-80679-21

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

# Client Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Client Sample ID: 1001-Main

Date Collected: 04/05/17 07:01

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-1

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:19	1

## Client Sample ID: 1001-Main

Date Collected: 04/05/17 07:02

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-2

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:21	1

## Client Sample ID: 1002-Math

Date Collected: 04/05/17 07:05

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-3

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:22	1

## Client Sample ID: 1002-Math

Date Collected: 04/05/17 07:06

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-4

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:24	1

## Client Sample ID: 1003-Gym A

Date Collected: 04/05/17 07:08

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-5

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.68		0.50	ppb		04/13/17 07:09	04/17/17 19:26	1

## Client Sample ID: 1004-Gym B

Date Collected: 04/05/17 07:12

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-6

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:28	1

## Client Sample ID: 1005-W/R

Date Collected: 04/05/17 07:18

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-7

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 19:30	1

TestAmerica Phoenix



## Client Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

### Client Sample ID: 1005-W/R

Date Collected: 04/05/17 07:20

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-8

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:09	04/17/17 19:32	1

### Client Sample ID: 1006-Fine Arts

Date Collected: 04/05/17 06:55

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-9

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:09	04/17/17 19:34	1

### Client Sample ID: 1007-VOC

Date Collected: 04/05/17 07:30

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-10

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:09	04/17/17 19:36	1

### Client Sample ID: 1007-VOC

Date Collected: 04/05/17 07:31

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-11

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:14	04/22/17 19:08	1

### Client Sample ID: 1008-Portables

Date Collected: 04/05/17 07:25

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-12

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:14	04/22/17 19:14	1

### Client Sample ID: 1008-Portables

Date Collected: 04/05/17 07:26

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-13

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:14	04/22/17 19:16	1

### Client Sample ID: 1009-JTED

Date Collected: 04/05/17 07:34

Date Received: 04/10/17 11:30

### Lab Sample ID: 550-80679-14

Matrix: Drinking Water

#### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb	—	04/13/17 07:14	04/22/17 19:17	1

TestAmerica Phoenix

# Client Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Client Sample ID: 1009-JTED

Date Collected: 04/05/17 07:35

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-15

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.59		0.50	ppb		04/13/17 07:14	04/22/17 19:19	1

## Client Sample ID: 1011-MPR

Date Collected: 04/05/17 06:46

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-16

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:21	1

## Client Sample ID: 1012-Book

Date Collected: 04/05/17 06:57

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-17

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:23	1

## Client Sample ID: 1013-Music

Date Collected: 04/05/17 06:58

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-18

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:25	1

## Client Sample ID: 1014-New R/R

Date Collected: 04/05/17 06:54

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-19

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:27	1

## Client Sample ID: 1003-Gym A

Date Collected: 04/05/17 07:13

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-20

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.0		0.50	ppb		04/13/17 07:14	04/22/17 19:29	1

## Client Sample ID: 1004-Gym B

Date Collected: 04/05/17 07:12

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-21

Matrix: Drinking Water

### Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:35	1

TestAmerica Phoenix

# QC Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 550-114756/1-A

Matrix: Water

Analysis Batch: 115117

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 114756

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:09	04/17/17 18:45	1

Lab Sample ID: LCS 550-114756/3-A

Matrix: Water

Analysis Batch: 115117

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 114756

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	93.9		ppb		94	85 - 115

Lab Sample ID: LCSD 550-114756/4-A

Matrix: Water

Analysis Batch: 115117

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 114756

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	100	93.8		ppb		94	85 - 115	0	20

Lab Sample ID: 550-80614-A-3-B MS

Matrix: Water

Analysis Batch: 115117

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 114756

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	1.6		100	92.4		ppb		91	70 - 130

Lab Sample ID: 550-80614-A-3-C MSD

Matrix: Water

Analysis Batch: 115117

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 114756

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	1.6		100	93.2		ppb		92	70 - 130	1	20

Lab Sample ID: MB 550-114758/1-A

Matrix: Water

Analysis Batch: 115521

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 114758

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		04/13/17 07:14	04/22/17 19:01	1

Lab Sample ID: LCS 550-114758/3-A

Matrix: Water

Analysis Batch: 115532

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 114758

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	98.3		ppb		98	85 - 115

Lab Sample ID: LCSD 550-114758/4-A

Matrix: Water

Analysis Batch: 115521

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 114758

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	100	85.6		ppb		86	85 - 115	2	20

TestAmerica Phoenix

## QC Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

Lab Sample ID: 550-80679-11 MS  
Matrix: Drinking Water  
Analysis Batch: 115521

Client Sample ID: 1007-VOC  
Prep Type: Total/NA  
Prep Batch: 114758  
%Rec.

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Lead	ND		100	88.9		ppb		89	70 - 130

Lab Sample ID: 550-80679-11 MSD  
Matrix: Drinking Water  
Analysis Batch: 115521

Client Sample ID: 1007-VOC  
Prep Type: Total/NA  
Prep Batch: 114758  
%Rec. RPD

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	Limits	RPD	Limit
Lead	ND		100	82.8		ppb		83	70 - 130	7	20

# QC Association Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Metals

### Prep Batch: 114756

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-80679-1	1001-Main	Total/NA	Drinking Water	200.8	
550-80679-2	1001-Main	Total/NA	Drinking Water	200.8	
550-80679-3	1002-Math	Total/NA	Drinking Water	200.8	
550-80679-4	1002-Math	Total/NA	Drinking Water	200.8	
550-80679-5	1003-Gym A	Total/NA	Drinking Water	200.8	
550-80679-6	1004-Gym B	Total/NA	Drinking Water	200.8	
550-80679-7	1005-W/R	Total/NA	Drinking Water	200.8	
550-80679-8	1005-W/R	Total/NA	Drinking Water	200.8	
550-80679-9	1006-Fine Arts	Total/NA	Drinking Water	200.8	
550-80679-10	1007-VOC	Total/NA	Drinking Water	200.8	
MB 550-114756/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-114756/3-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-114756/4-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-80614-A-3-B MS	Matrix Spike	Total/NA	Water	200.8	
550-80614-A-3-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	

### Prep Batch: 114758

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-80679-11	1007-VOC	Total/NA	Drinking Water	200.8	
550-80679-12	1008-Portables	Total/NA	Drinking Water	200.8	
550-80679-13	1008-Portables	Total/NA	Drinking Water	200.8	
550-80679-14	1009-JTED	Total/NA	Drinking Water	200.8	
550-80679-15	1009-JTED	Total/NA	Drinking Water	200.8	
550-80679-16	1011-MPR	Total/NA	Drinking Water	200.8	
550-80679-17	1012-Book	Total/NA	Drinking Water	200.8	
550-80679-18	1013-Music	Total/NA	Drinking Water	200.8	
550-80679-19	1014-New R/R	Total/NA	Drinking Water	200.8	
550-80679-20	1003-Gym A	Total/NA	Drinking Water	200.8	
550-80679-21	1004-Gym B	Total/NA	Drinking Water	200.8	
MB 550-114758/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-114758/3-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-114758/4-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-80679-11 MS	1007-VOC	Total/NA	Drinking Water	200.8	
550-80679-11 MSD	1007-VOC	Total/NA	Drinking Water	200.8	

### Analysis Batch: 115117

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-80679-1	1001-Main	Total/NA	Drinking Water	200.8	114756
550-80679-2	1001-Main	Total/NA	Drinking Water	200.8	114756
550-80679-3	1002-Math	Total/NA	Drinking Water	200.8	114756
550-80679-4	1002-Math	Total/NA	Drinking Water	200.8	114756
550-80679-5	1003-Gym A	Total/NA	Drinking Water	200.8	114756
550-80679-6	1004-Gym B	Total/NA	Drinking Water	200.8	114756
550-80679-7	1005-W/R	Total/NA	Drinking Water	200.8	114756
550-80679-8	1005-W/R	Total/NA	Drinking Water	200.8	114756
550-80679-9	1006-Fine Arts	Total/NA	Drinking Water	200.8	114756
550-80679-10	1007-VOC	Total/NA	Drinking Water	200.8	114756
MB 550-114756/1-A	Method Blank	Total/NA	Water	200.8	114756
LCS 550-114756/3-A	Lab Control Sample	Total/NA	Water	200.8	114756
LCSD 550-114756/4-A	Lab Control Sample Dup	Total/NA	Water	200.8	114756
550-80614-A-3-B MS	Matrix Spike	Total/NA	Water	200.8	114756

TestAmerica Phoenix

## QC Association Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

### Metals (Continued)

#### Analysis Batch: 115117 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-80614-A-3-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	114756

#### Analysis Batch: 115521

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-80679-11	1007-VOC	Total/NA	Drinking Water	200.8	114758
550-80679-12	1008-Portables	Total/NA	Drinking Water	200.8	114758
550-80679-13	1008-Portables	Total/NA	Drinking Water	200.8	114758
550-80679-14	1009-JTED	Total/NA	Drinking Water	200.8	114758
550-80679-15	1009-JTED	Total/NA	Drinking Water	200.8	114758
550-80679-16	1011-MPR	Total/NA	Drinking Water	200.8	114758
550-80679-17	1012-Book	Total/NA	Drinking Water	200.8	114758
550-80679-18	1013-Music	Total/NA	Drinking Water	200.8	114758
550-80679-19	1014-New R/R	Total/NA	Drinking Water	200.8	114758
550-80679-20	1003-Gym A	Total/NA	Drinking Water	200.8	114758
550-80679-21	1004-Gym B	Total/NA	Drinking Water	200.8	114758
MB 550-114758/1-A	Method Blank	Total/NA	Water	200.8	114758
LCSD 550-114758/4-A	Lab Control Sample Dup	Total/NA	Water	200.8	114758
550-80679-11 MS	1007-VOC	Total/NA	Drinking Water	200.8	114758
550-80679-11 MSD	1007-VOC	Total/NA	Drinking Water	200.8	114758

#### Analysis Batch: 115532

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
LCS 550-114758/3-A	Lab Control Sample	Total/NA	Water	200.8	114758

# Lab Chronicle

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Client Sample ID: 1001-Main

Date Collected: 04/05/17 07:01

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:19	TEK	TAL PHX

## Client Sample ID: 1001-Main

Date Collected: 04/05/17 07:02

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:21	TEK	TAL PHX

## Client Sample ID: 1002-Math

Date Collected: 04/05/17 07:05

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:22	TEK	TAL PHX

## Client Sample ID: 1002-Math

Date Collected: 04/05/17 07:06

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:24	TEK	TAL PHX

## Client Sample ID: 1003-Gym A

Date Collected: 04/05/17 07:08

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:26	TEK	TAL PHX

## Client Sample ID: 1004-Gym B

Date Collected: 04/05/17 07:12

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:28	TEK	TAL PHX

TestAmerica Phoenix

# Lab Chronicle

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

**Client Sample ID: 1005-W/R**

**Date Collected: 04/05/17 07:18**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-7**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:30	TEK	TAL PHX

**Client Sample ID: 1005-W/R**

**Date Collected: 04/05/17 07:20**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-8**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:32	TEK	TAL PHX

**Client Sample ID: 1006-Fine Arts**

**Date Collected: 04/05/17 06:55**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-9**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:34	TEK	TAL PHX

**Client Sample ID: 1007-VOC**

**Date Collected: 04/05/17 07:30**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-10**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114756	04/13/17 07:09	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115117	04/17/17 19:36	TEK	TAL PHX

**Client Sample ID: 1007-VOC**

**Date Collected: 04/05/17 07:31**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-11**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:08	TEK	TAL PHX

**Client Sample ID: 1008-Portables**

**Date Collected: 04/05/17 07:25**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-12**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:14	TEK	TAL PHX

TestAmerica Phoenix



# Lab Chronicle

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

## Client Sample ID: 1008-Portables

Date Collected: 04/05/17 07:26

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-13

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:16	TEK	TAL PHX

## Client Sample ID: 1009-JTED

Date Collected: 04/05/17 07:34

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-14

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:17	TEK	TAL PHX

## Client Sample ID: 1009-JTED

Date Collected: 04/05/17 07:35

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-15

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:19	TEK	TAL PHX

## Client Sample ID: 1011-MPR

Date Collected: 04/05/17 06:46

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-16

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:21	TEK	TAL PHX

## Client Sample ID: 1012-Book

Date Collected: 04/05/17 06:57

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-17

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:23	TEK	TAL PHX

## Client Sample ID: 1013-Music

Date Collected: 04/05/17 06:58

Date Received: 04/10/17 11:30

## Lab Sample ID: 550-80679-18

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:25	TEK	TAL PHX

TestAmerica Phoenix

## Lab Chronicle

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

**Client Sample ID: 1014-New R/R**

**Date Collected: 04/05/17 06:54**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-19**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:27	TEK	TAL PHX

**Client Sample ID: 1003-Gym A**

**Date Collected: 04/05/17 07:13**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-20**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:29	TEK	TAL PHX

**Client Sample ID: 1004-Gym B**

**Date Collected: 04/05/17 07:12**

**Date Received: 04/10/17 11:30**

**Lab Sample ID: 550-80679-21**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			114758	04/13/17 07:14	SGO	TAL PHX
Total/NA	Analysis	200.8		1	115521	04/22/17 19:35	TEK	TAL PHX

### Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

## Accreditation/Certification Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

### Laboratory: TestAmerica Phoenix

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	EPA Region	Identification Number	Expiration Date
Arizona	State Program	9	AZ0728	06-09-17

Analysis Method	Prep Method	Matrix	Analyte
-----------------	-------------	--------	---------

## Method Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW / Prescott Unified District

TestAmerica Job ID: 550-80679-1  
SDG: Prescott High School

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	TAL PHX

### Protocol References:

EPA = US Environmental Protection Agency


### Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

80619 Collection Log  
for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling


Name of School District	P450
School Name	High School
Building (name/number)	1001 - MAIN
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <u>Write this number on the sample container and on this sheet</u> )	
Date of Collection	4/5/17
Time of Collection	0701
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

550-80679 Chain of Custody



For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only  
Relinquished date and signature  11:30 4/10/17 (21.3") no ice


These samples were collected for lead screening purposes only and cannot be used for compliance.

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P430
School Name	HS
Building (name/number)	1001 MAIN
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-02 Date of Collection	4/5/17
Time of Collection	0702
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P43D
School Name	HS
Building (name/number)	1002 - MATH
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-03 Date of Collection	4/5/2017
Time of Collection	0705
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	RJ

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	PUSD
School Name	HS
Building (name/number)	1002 - MATH
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-04 Date of Collection	4/5/2017
Time of Collection	0700
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	RJ

Notes Sample collector:
-------------------------

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*




## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P430
School Name	HS
Building (name/number)	1003- GYM A
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
- 05	Date of Collection 4/5/2017
	Time of Collection 0708
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:
-------------------------

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P430
School Name	HS
Building (name/number)	1004- Gym B
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b>Write this number on the sample container and on this sheet</b> )	
Date of Collection	4/5/2017
Time of Collection	
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	P430
School Name	HS
Building (name/number)	1005-WIR
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0718
Printed Name of Sample Collector	RICK LONG
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	R430
School Name	NS
Building (name/number)	1005 - WTR
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0720
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	RZ

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	PHSD
School Name	HS
Building (name/number)	1006 - FIVE ARTS
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0655
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	DUSD
School Name	HS
Building (name/number)	1007-102
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-10 Date of Collection	4/5/2017
Time of Collection	0730
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	PUSD
School Name	HS
Building (name/number)	1057-VOC
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0731
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P430
School Name	NS
Building (name/number)	1008 - Portables
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-12 Date of Collection	4/5/2017
Time of Collection	0725
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*




## Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DUSD
School Name	HS
Building (name/number)	1008- Portables
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0720
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:
-------------------------

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	DUSD
School Name	HS
Building (name/number)	1009-JTED
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	
Sample Identification Number ( <b>Write this number on the sample container and on this sheet</b> )	
Date of Collection	4/5/2017
Time of Collection	0734
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	P430
School Name	NS
Building (name/number)	1009-5TSD
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-15 Date of Collection	4/5/2017
Time of Collection	0735
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	Rick Long

Notes Sample collector:
-------------------------

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	PUSD
School Name	HS
Building (name/number)	1011-MPR
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-16 Date of Collection	4/5/2017
Time of Collection	0640
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	PC130
School Name	HS
Building (name/number)	1012 - BOOK
Type of Fixture (tap, drinking fountain etc.)	DRINK FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0657
Printed Name of Sample Collector	RICK LONG
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	D430
School Name	HS
Building (name/number)	1013 - MUSIC
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-18 Date of Collection	4.5.2017
Time of Collection	0658
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_


*These samples were collected for lead screening purposes only and cannot be used for compliance.*

### Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	P450
School Name	HS
Building (name/number)	1014 - New RIR
Type of Fixture (tap, drinking fountain etc.)	OF
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4.5.2017
Time of Collection	0654
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	PHSD
School Name	HS
Building (name/number)	1003- GYM A
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
-20 Date of Collection	4/5/2017
Time of Collection	0713
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	RJ

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*



## Collection Log

*for experienced sample collectors*

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	P430
School Name	HS
Building (name/number)	1004- Gym B
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	
Sample Identification Number ( <b><u>Write this number on the sample container and on this sheet</u></b> )	
Date of Collection	4/5/2017
Time of Collection	0712
Printed Name of Sample Collector	Rick Long
Signature Sample Collector	[Signature]

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

## Login Sample Receipt Checklist

Client: Arizona Dept of Environmental Quality

Job Number: 550-80679-1  
SDG Number: Prescott High School

**Login Number: 80679**

**List Number: 1**

**Creator: Gravlin, Andrea**

**List Source: TestAmerica Phoenix**

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	False	No sample date and/or time on COC, logged in per container labels.
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	Check done at department level as required.