

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Phoenix

4625 East Cotton Ctr Blvd

Suite 189

Phoenix, AZ 85040

Tel: (602)437-3340

TestAmerica Job ID: 550-78333-1

TestAmerica SDG: Bradshaw Mountain High School West

Client Project/Site: Pb in DW / Humboldt Unified #22

For:

Arizona Dept of Environmental Quality

1110 W. Washington St., MC5415B-1

Phoenix, Arizona 85007

Attn: David A Burchard



Authorized for release by:

3/15/2017 1:09:12 PM

Susan Armijo, Project Manager II

(602)659-7664

susan.armijo@testamericainc.com

LINKS

Review your project
results through

Total Access

Have a Question?



Visit us at:

www.testamericainc.com

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1

2

3

4

5

6

7

8

9

10

11

12

13

14



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Sample Summary	5
Detection Summary	6
Client Sample Results	7
QC Sample Results	8
QC Association Summary	10
Lab Chronicle	11
Certification Summary	13
Method Summary	14
Chain of Custody	15
Receipt Checklists	21

Definitions/Glossary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Qualifiers

Metals

Qualifier	Qualifier Description
V1	CCV recovery was above method acceptance limits. The analyte was not detected in the sample.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Job ID: 550-78333-1

Laboratory: TestAmerica Phoenix

Narrative

Job Narrative 550-78333-1

Comments

No additional comments.

Receipt

The samples were received on 2/28/2017 10:45 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 18.3° C.


Receipt Exceptions

The following samples for metals were received unpreserved and were preserved upon receipt to the laboratory: #1 (550-78333-1), #2 (550-78333-2), #3 (550-78333-3), #4 (550-78333-4), #5 (550-78333-5), #6 (550-78333-6) and #7 (550-78333-7). Regulatory documents require a 24-hour waiting period from the time of the addition of the acid preservative to the time of digestion.

Metals

Method(s) 200.8: The continuing calibration verification (CCV) associated with batch 550-111961 recovered above the upper control limit for Lead (Pb). The samples associated with this CCV were non-detects for the affected analytes; therefore, the data have been reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.



Sample Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
550-78333-1	#1	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-2	#2	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-3	#3	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-4	#4	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-5	#5	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-6	#6	Drinking Water	02/27/17 11:00	02/28/17 10:45
550-78333-7	#7	Drinking Water	02/27/17 11:00	02/28/17 10:45

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Detection Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Client Sample ID: #1**Lab Sample ID: 550-78333-1**

No Detections.

Client Sample ID: #2**Lab Sample ID: 550-78333-2**

No Detections.

Client Sample ID: #3**Lab Sample ID: 550-78333-3**

No Detections.

Client Sample ID: #4**Lab Sample ID: 550-78333-4**

No Detections.

Client Sample ID: #5**Lab Sample ID: 550-78333-5**

No Detections.

Client Sample ID: #6**Lab Sample ID: 550-78333-6**


No Detections.

Client Sample ID: #7**Lab Sample ID: 550-78333-7**

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix



Client Sample Results

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Client Sample ID: #1

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-1

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:40	03/08/17 22:38	1

Client Sample ID: #2

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-2

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:40	03/08/17 22:40	1

Client Sample ID: #3

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-3

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:40	03/08/17 22:42	1

Client Sample ID: #4

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-4

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND	V1	0.50	ppb		03/07/17 09:55	03/08/17 22:58	1

Client Sample ID: #5

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-5

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND	V1	0.50	ppb		03/07/17 09:55	03/08/17 23:04	1

Client Sample ID: #6

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-6

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND	V1	0.50	ppb		03/07/17 09:55	03/08/17 23:06	1

Client Sample ID: #7

Date Collected: 02/27/17 11:00
Date Received: 02/28/17 10:45

Lab Sample ID: 550-78333-7

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:55	03/10/17 18:42	1

TestAmerica Phoenix

QC Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
 SDG: Bradshaw Mountain High School West

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 550-111730/1-A
Matrix: Water
Analysis Batch: 111960

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 111730

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:40	03/08/17 21:48	1

Lab Sample ID: LCS 550-111730/4-A
Matrix: Water
Analysis Batch: 111960

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 111730

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	105		ppb		105	85 - 115

Lab Sample ID: LCSD 550-111730/5-A
Matrix: Water
Analysis Batch: 111960

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 111730

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	100	105		ppb		105	85 - 115	0	20

Lab Sample ID: 550-78329-A-15-B MS
Matrix: Water
Analysis Batch: 111960

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 111730

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	ND		100	102		ppb		102	70 - 130

Lab Sample ID: 550-78329-A-15-C MSD
Matrix: Water
Analysis Batch: 111960

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 111730

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	ND		100	101		ppb		100	70 - 130	1	20

Lab Sample ID: MB 550-111734/1-A
Matrix: Water
Analysis Batch: 111961

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 111734

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND	V1	0.50	ppb		03/07/17 09:55	03/08/17 22:48	1

Lab Sample ID: MB 550-111734/1-A
Matrix: Water
Analysis Batch: 112143

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 111734

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/07/17 09:55	03/10/17 18:26	1

Lab Sample ID: LCS 550-111734/4-A
Matrix: Water
Analysis Batch: 111961

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	108	V1	ppb		108	85 - 115

TestAmerica Phoenix

QC Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
 SDG: Bradshaw Mountain High School West

Lab Sample ID: LCS 550-111734/4-A
Matrix: Water
Analysis Batch: 112143

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	92.9		ppb		93	85 - 115

Lab Sample ID: LCSD 550-111734/5-A
Matrix: Water
Analysis Batch: 111961

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	100	108	V1	ppb		108	85 - 115	0	20

Lab Sample ID: LCSD 550-111734/5-A
Matrix: Water
Analysis Batch: 112143

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	100	91.8		ppb		92	85 - 115	1	20

Lab Sample ID: 550-78333-4 MS
Matrix: Drinking Water
Analysis Batch: 111961

Client Sample ID: #4
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	ND	V1	100	103	V1	ppb		103	70 - 130

Lab Sample ID: 550-78333-4 MS
Matrix: Drinking Water
Analysis Batch: 112143

Client Sample ID: #4
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	ND		100	87.9		ppb		88	70 - 130

Lab Sample ID: 550-78333-4 MSD
Matrix: Drinking Water
Analysis Batch: 111961

Client Sample ID: #4
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	ND	V1	100	104	V1	ppb		104	70 - 130	0	20

Lab Sample ID: 550-78333-4 MSD
Matrix: Drinking Water
Analysis Batch: 112143

Client Sample ID: #4
Prep Type: Total/NA
Prep Batch: 111734

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	ND		100	87.4		ppb		87	70 - 130	1	20

QC Association Summary

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
 SDG: Bradshaw Mountain High School West

Metals

Prep Batch: 111730

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78333-1	#1	Total/NA	Drinking Water	200.8	
550-78333-2	#2	Total/NA	Drinking Water	200.8	
550-78333-3	#3	Total/NA	Drinking Water	200.8	
MB 550-111730/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-111730/4-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-111730/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-78329-A-15-B MS	Matrix Spike	Total/NA	Water	200.8	
550-78329-A-15-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	

Prep Batch: 111734

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78333-4	#4	Total/NA	Drinking Water	200.8	
550-78333-5	#5	Total/NA	Drinking Water	200.8	
550-78333-6	#6	Total/NA	Drinking Water	200.8	
550-78333-7	#7	Total/NA	Drinking Water	200.8	
MB 550-111734/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-111734/4-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-111734/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-78333-4 MS	#4	Total/NA	Drinking Water	200.8	
550-78333-4 MSD	#4	Total/NA	Drinking Water	200.8	

Analysis Batch: 111960

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78333-1	#1	Total/NA	Drinking Water	200.8	111730
550-78333-2	#2	Total/NA	Drinking Water	200.8	111730
550-78333-3	#3	Total/NA	Drinking Water	200.8	111730
MB 550-111730/1-A	Method Blank	Total/NA	Water	200.8	111730
LCS 550-111730/4-A	Lab Control Sample	Total/NA	Water	200.8	111730
LCSD 550-111730/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	111730
550-78329-A-15-B MS	Matrix Spike	Total/NA	Water	200.8	111730
550-78329-A-15-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	111730

Analysis Batch: 111961

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78333-4	#4	Total/NA	Drinking Water	200.8	111734
550-78333-5	#5	Total/NA	Drinking Water	200.8	111734
550-78333-6	#6	Total/NA	Drinking Water	200.8	111734
MB 550-111734/1-A	Method Blank	Total/NA	Water	200.8	111734
LCS 550-111734/4-A	Lab Control Sample	Total/NA	Water	200.8	111734
LCSD 550-111734/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	111734
550-78333-4 MS	#4	Total/NA	Drinking Water	200.8	111734
550-78333-4 MSD	#4	Total/NA	Drinking Water	200.8	111734

Analysis Batch: 112143

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78333-7	#7	Total/NA	Drinking Water	200.8	111734
MB 550-111734/1-A	Method Blank	Total/NA	Water	200.8	111734
LCS 550-111734/4-A	Lab Control Sample	Total/NA	Water	200.8	111734
LCSD 550-111734/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	111734
550-78333-4 MS	#4	Total/NA	Drinking Water	200.8	111734
550-78333-4 MSD	#4	Total/NA	Drinking Water	200.8	111734

TestAmerica Phoenix

Lab Chronicle

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
 SDG: Bradshaw Mountain High School West

Client Sample ID: #1

Lab Sample ID: 550-78333-1

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111730	03/07/17 09:40	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111960	03/08/17 22:38	KLH	TAL PHX

Client Sample ID: #2

Lab Sample ID: 550-78333-2

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111730	03/07/17 09:40	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111960	03/08/17 22:40	KLH	TAL PHX

Client Sample ID: #3

Lab Sample ID: 550-78333-3

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111730	03/07/17 09:40	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111960	03/08/17 22:42	KLH	TAL PHX

Client Sample ID: #4

Lab Sample ID: 550-78333-4

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111734	03/07/17 09:55	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111961	03/08/17 22:58	KLH	TAL PHX

Client Sample ID: #5

Lab Sample ID: 550-78333-5

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111734	03/07/17 09:55	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111961	03/08/17 23:04	KLH	TAL PHX

Client Sample ID: #6

Lab Sample ID: 550-78333-6

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111734	03/07/17 09:55	SGO	TAL PHX
Total/NA	Analysis	200.8		1	111961	03/08/17 23:06	KLH	TAL PHX

Lab Chronicle

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Client Sample ID: #7

Lab Sample ID: 550-78333-7

Date Collected: 02/27/17 11:00

Matrix: Drinking Water

Date Received: 02/28/17 10:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			111734	03/07/17 09:55	SGO	TAL PHX
Total/NA	Analysis	200.8		1	112143	03/10/17 18:42	TEK	TAL PHX

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Certification Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Laboratory: TestAmerica Phoenix

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
Arizona	State Program	9	AZ0728	06-09-17

Analysis Method	Prep Method	Matrix	Analyte
-----------------	-------------	--------	---------

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Humboldt Unified #22

TestAmerica Job ID: 550-78333-1
SDG: Bradshaw Mountain High School West

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	TAL PHX

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340



78333

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt UNIF.ES #22
School Name	BRADSHAW Mtn H.S. WEST
Building (name/number)	Build # 1001
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Office
Sample Identification Number (ensure this number is also on the sample container)	#1
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	KEN JOHNSON
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

550-78333 Chain of Custody



For relinquishing samples upon delivery to labs only

Relinquished date and signature FedEx → *[Signature]* 2/28/17 10:45 (18.300) No ICE

These samples were collected for lead screening purposes only and cannot be used for compliance.

①

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt Unified #22
School Name	BROADSHAW MIDDLE H.S. WEST
Building (name/number)	Building B 1002
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	LIBRARY
Sample Identification Number (ensure this number is also on the sample container)	42
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	KEN JOHNSON
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

②

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt Unified #22
School Name	Bradshaw Mt. H.S. West
Building (name/number)	1003 Bldg C
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	ENTRANCE
Sample Identification Number (ensure this number is also on the sample container)	#3
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	KEN JOHNSON
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

3

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt U.P.S. #22
School Name	BRADSDON Mtn H.S. W.P.S.
Building (name/number)	1004 Bldg D, E, F
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	By CLASSROOMS
Sample Identification Number (ensure this number is also on the sample container)	#4
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	Ken Johnson
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

④

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt Univ. FFS #22
School Name	BRADSHAW Mtn H. S. WEST
Building (name/number)	1005 Bldg A
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	HALLWAY
Sample Identification Number (ensure this number is also on the sample container)	#5 & #6
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	Ken Johnson
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

5

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Humboldt Unified #22
School Name	BRADENAWAY H.S.
Building (name/number)	1007 Bldg F
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	BAND ROOM
Sample Identification Number (ensure this number is also on the sample container)	47
Date of Collection	2/27/17
Time of Collection	11:00
Printed Name of Sample Collector	KEN JOHNSON
Signature Sample Collector	<i>Ken Johnson</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

6

Login Sample Receipt Checklist

Client: Arizona Dept of Environmental Quality

Job Number: 550-78333-1

SDG Number: Bradshaw Mountain High School West

Login Number: 78333

List Number: 1

Creator: Vilaboy, Monica

List Source: TestAmerica Phoenix

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	Check done at department level as required.

