

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)



Laboratory Report

for

Arizona Department of Environmental Quality
1110 West Washington Street
Phoenix, AZ 85007
Attention: David Burchard

Date of Issue
02/23/2017



Eurofins Eaton
Analytical

TDF: Thomas.D.French
Project Manager



Report: 640298
Project: DOUGLAS-USD
ADHS License #: AZ0778
Group: Douglas High School
PO#: PO#: ADEQ16-116686:3

* Accredited in accordance with TNI 2009 and ISO/IEC 17025:2005.

* Laboratory certifies that the test results meet all **TNI 2009 and ISO/IEC 17025:2005** requirements unless noted under the individual analysis.

* Following the cover page are State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.

* Test results relate only to the sample(s) tested.

* This report shall not be reproduced except in full, without the written approval of the laboratory.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Mississippi	Certified
-----	-----	Montana	Cert 0035
Arizona	AZ0778	Nebraska	Certified
Arkansas	Certified	Nevada	CA00006-2016
California-Monrovia-ELAP	2813	New Hampshire *	2959
California-Colton- ELAP	2812	New Jersey *	CA 008
California-Folsom- ELAP	2820	New Mexico	Certified
California-Fresno- ELAP	2966	New York *	11320
Colorado	Certified	North Carolina	06701
Connecticut	PH-0107	North Dakota	R-009
Delaware	CA 006	Oregon (Primary AB) *	ORELAP 4034
Florida *	E871024	Pennsylvania *	68-565
Georgia	947	Puerto Rico	Certified
Guam	16-003r	Rhode Island	LAO00326
Hawaii	Certified	South Carolina	87016
Idaho	Certified	South Dakota	Certified
Illinois *	200033	Tennessee	TN02839
Indiana	C-CA-01	Texas *	T104704230-15-9
Kansas *	E-10268	Utah *	CA000062016-10
Kentucky	90107	Vermont	VT0114
Louisiana *	LA16003	Virginia *	460260
Maine	CA0006	Washington	C838
Maryland	224	-----	-----
Commonwealth of Northern Marianas Is.	MP0004	-----	-----
Massachusetts	M-CA006	EPA Region 5	Certified
Michigan	9906	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO 17025 as verified by the ANSI-ASQ National Accreditation Board/ANAB.

Refer to Certificate and scope of accreditation (AT 1807) found at: <http://www.eatonanalytical.com>

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
1,4-Dioxane	EPA 522	x		x
2,3,7,8-TCDD	Modified EPA 1613B	x		x
Acrylamide	In House Method (2440)	x		x
Alkalinity	SM 2320B	x	x	x
Ammonia	EPA 350.1		x	x
Ammonia	SM 4500-NH3 H		x	x
Anions and DBPs by IC	EPA 300.0	x	x	x
Anions and DBPs by IC	EPA 300.1	x		x
Asbestos	EPA 100.2	x	x	
Bicarbonate Alkalinity as HCO ₃	SM 2320B	x	x	x
BOD / CBOD	SM 5210B		x	x
Bromate	In House Method (2447)	x		x
Carbamates	EPA 531.2	x		x
Carbonate as CO ₃	SM 2330B	x	x	x
Carbonyls	EPA 556	x		x
COD	EPA 410.4 / SM 5220D		x	
Chloramines	SM 4500-CL G	x	x	x
Chlorinated Acids	EPA 515.4	x		x
Chlorinated Acids	EPA 555	x		x
Chlorine Dioxide	SM 4500-CLO ₂ D	x		x
Chlorine -Total/Free/ Combined Residual	SM 4500-Cl G	x	x	x
Conductivity	EPA 120.1		x	
Conductivity	SM 2510B	x	x	x
Corrosivity (Langelier Index)	SM 2330B	x		x
Cryptosporidium	EPA 1623	x		x
Cyanide, Amenable	SM 4500-CN G	x	x	
Cyanide, Free	SM 4500CN F	x	x	x
Cyanide, Total	EPA 335.4	x	x	x
Cyanogen Chloride (screen)	In House Method (2470)	x		x
Diquat and Paraquat	EPA 549.2	x		x
DBP/HAA	SM 6251B	x		x
Dissolved Oxygen	SM 4500-O G		x	x
DOC	SM 5310C	x		x
E. Coli (MTF/EC+MUG)		x		x
E. Coli	CFR 141.21(f)(6)(i)	x		x
E. Coli	SM 9223		x	
E. Coli (Enumeration)	SM 9221B.1/ SM 9221F	x		x
E. Coli (Enumeration)	SM 9223B	x		x
EDB/DCBP	EPA 504.1	x		
EDB/DBCP and DBP	EPA 551.1	x		x
EDTA and NTA	In House Method (2454)	x		x
Endothall	EPA 548.1	x		x
Endothall	In-house Method (2445)	x		x
Enterococci	SM 9230B	x	x	
Fecal Coliform	SM 9221 E (MTF/EC)	x		
Fecal Coliform	SM 9221C, E (MTF/EC)		x	
Fecal Coliform (Enumeration)	SM 9221E (MTF/EC)	x		x
Fecal Coliform with Chlorine Present	SM 9221E		x	
Fecal Streptococci	SM 9230B	x	x	
Fluoride	SM 4500-F C	x	x	x
Giardia	EPA 1623	x		x
Glyphosate	EPA 547	x		x
Gross Alpha/Beta	EPA 900.0	x	x	x
Gross Alpha Coprecipitation	SM 7110 C	x	x	x
Hardness	SM 2340B	x	x	x
Heterotrophic Bacteria	In House Method (2439)	x		x
Heterotrophic Bacteria	SM 9215 B	x		x
Hexavalent Chromium	EPA 218.6	x	x	x

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
Hexavalent Chromium	EPA 218.7	x		x
Hexavalent Chromium	SM 3500-Cr B		x	
Hormones	EPA 539	x		x
Hydroxide as OH Calc.	SM 2330B	x		x
Kjeldahl Nitrogen	EPA 351.2		x	
Legionella	CDC Legionella	x		x
Mercury	EPA 245.1	x	x	x
Metals	EPA 200.7 / 200.8	x	x	x
Microcystin LR	ELISA (2360)	x		x
NDMA	EPA 521	x		x
NDMA	TQ In house method based on EPA 521 (2425)	x		x
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x	x
OCL, Pesticides/PCB	EPA 505	x		x
Ortho Phosphate	EPA 365.1	x	x	x
Ortho Phosphate	SM 4500P E			x
Ortho Phosphorous	SM 4500P E	x		
Oxyhalides Disinfection Byproducts	EPA 317.0	x		x
Perchlorate	EPA 331.0	x		x
Perchlorate (low and high)	EPA 314.0	x		x
Perfluorinated Alkyl Acids	EPA 537	x		x
pH	EPA 150.1	x		
pH	SM 4500-H+B	x	x	x
Phenylurea Pesticides/ Herbicides	In House Method, based on EPA 532 (2448)	x		x
Pseudomonas	IDEXX Pseudalert (2461)	x		x
Radium-226	GA Institute of Tech	x		x
Radium-228	GA Institute of Tech	x		x
Radon-222	SM 7500RN	x		x
Residue, Filterable	SM 2540C	x	x	x
Residue, Non-filterable	SM 2540D		x	
Residue, Total	SM 2540B		x	x
Residue, Volatile	EPA 160.4		x	
Semi-VOC	EPA 525.2	x		x
Semi-VOC	EPA 625		x	x
Silica	SM 4500-Si D	x	x	
Silica	SM 4500-SiO ₂ C	x	x	
Sulfide	SM 4500-S ⁻ D		x	
Sulfite	SM 4500-SO ₃ B	x	x	x
Surfactants	SM 5540C	x	x	x
Taste and Odor Analytes	SM 6040E	x		x
Total Coliform (P/A)	SM 9221 A, B	x		x
Total Coliform (Enumeration)	SM 9221 A, B, C	x		x
Total Coliform / E. coli	Colisure SM 9223	x		x
Total Coliform	SM 9221B		x	
Total Coliform with Chlorine Present	SM 9221B		x	
Total Coliform / E.coli (P/A and Enumeration)	SM 9223	x		x
TOC	SM 5310C	x	x	x
TOX	SM 5320B		x	
Total Phenols	EPA 420.1		x	
Total Phenols	EPA 420.4	x	x	x
Total Phosphorous	SM 4500 P E		x	
Turbidity	EPA 180.1	x	x	x
Turbidity	SM 2130B	x	x	
Uranium by ICP/MS	EPA 200.8	x		x
UV 254	SM 5910B	x		
VOC	EPA 524.2/EPA 524.3	x		x
VOC	EPA 624		x	x
VOC	EPA SW 846 8260	x		x
VOC	In House Method (2411)	x		x
Yeast and Mold	SM 9610	x		x

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
<u>201702210610</u>	1001-055	02/17/2017 0500
	Sample Type: Tap Facility ID: 1st Wing West Deccastore 1001 Sample Point ID: RM 101	
	@ICPMS	Freight - Outbound
		Freight - Return
<u>201702210611</u>	1001-056	02/17/2017 0501
	Sample Type: Drinking Fountain Facility ID: 1st Wing W-1001 Sample Point ID: 1st Wing West (ext.)	
	@ICPMS	
<u>201702210612</u>	1001-057	02/17/2017 0502
	Sample Type: Drinking Fountain Facility ID: 1st Wing East - 1002 Sample Point ID: 1002 East (ext.)	
	@ICPMS	
<u>201702210613</u>	1001-058	02/17/2017 0503
	Sample Type: Tap Facility ID: 1st Wing - 1002 Sample Point ID: Nurse RM	
	@ICPMS	
<u>201702210614</u>	1001-059	02/17/2017 0504
	Sample Type: Drinking Fountain Facility ID: 1st Wing - 1002 Sample Point ID: Principal's Lobby	
	@ICPMS	
<u>201702210615</u>	1001-060	02/17/2017 0505
	Sample Type: Tap Facility ID: 1st Wing - 1002 Sample Point ID: Teachers' Lounge	
	@ICPMS	
<u>201702210616</u>	1001-061	02/17/2017 0507
	Sample Type: Drinking Fountain Facility ID: 2nd Wing - 1003 Sample Point ID: 1003 West (ext.)	
	@ICPMS	
<u>201702210617</u>	1001-062	02/17/2017 0508

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
	Sample Type: Tap Facility ID: 2nd Wing - 1003 Sample Point ID: RM 201 @ICPMS	
<u>201702210618</u>	1001-063	02/17/2017 0509
	Sample Type: Drinking Fountain Facility ID: 2nd Wing - 1004 Sample Point ID: 1004 East (ext.) @ICPMS	
<u>201702210619</u>	1001-064	02/17/2017 0510
	Sample Type: Tap Facility ID: 2nd Wing - 1004 Sample Point ID: East Custodian RM @ICPMS	
<u>201702210620</u>	1001-065	02/17/2017 0512
	Sample Type: Tap Facility ID: 3rd Wing - 1005 Sample Point ID: RM 303 @ICPMS	
<u>201702210621</u>	1001-066	02/17/2017 0513
	Sample Type: Tap Facility ID: 3rd Wing - 1005 Sample Point ID: RM 305 @ICPMS	
<u>201702210622</u>	1001-067	02/17/2017 0514
	Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: South Wall, West Fixture @ICPMS	
<u>201702210623</u>	1001-068	02/17/2017 0515
	Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: South Wall, East Fixture @ICPMS	
<u>201702210624</u>	1001-069	02/17/2017 0516
	Sample Type: Drinking Fountain Facility ID: 3rd Wing - 1006 Sample Point ID: RM 302	

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
	@ICPMS	
<u>201702210625</u>	1001-070 Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: RM 308	02/17/2017 0518
	@ICPMS	
<u>201702210626</u>	1001-071 Sample Type: Drinking Fountain Facility ID: 3rd Wing - 1006 Sample Point ID: 1006 - East (ext.)	02/17/2017 0520
	@ICPMS	
<u>201702210627</u>	1001-072 Sample Type: Drinking Fountain Facility ID: 4th Wing - 1007 Sample Point ID: 1007 - West (ext.)	02/17/2017 0522
	@ICPMS	
<u>201702210628</u>	1001-073 Sample Type: Drinking Fountain Facility ID: 4th Wing - 1008 Sample Point ID: 1008 - East (ext.)	02/17/2017 0523
	@ICPMS	
<u>201702210629</u>	1001-074 Sample Type: Drinking Fountain Facility ID: 5th Wing - 1009 Sample Point ID: 1009 - East (ext.)	02/17/2017 0525
	@ICPMS	
<u>201702210630</u>	1001-075 Sample Type: Tap Facility ID: 5th Wing - 1009 Sample Point ID: East Custodian RM	02/17/2017 0526
	@ICPMS	
<u>201702210631</u>	1001-076 Sample Type: Drinking Fountain Facility ID: 6th Wing - 1011 Sample Point ID: 1011 East (ext.) by RM 608	02/17/2017 0528
	@ICPMS	
<u>201702210632</u>	1001-077	02/17/2017 0530

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
	Sample Type: Tap Facility ID: 6th Wing - 1011 Sample Point ID: RM 608, 1st Fixture by door @ICPMS	
<u>201702210633</u>	1001-078	02/17/2017 0535
	Sample Type: Drinking Fountain Facility ID: Auditorium - 1012 Sample Point ID: East Lobby @ICPMS	
<u>201702210634</u>	1001-079	02/17/2017 0536
	Sample Type: Drinking Fountain Facility ID: Auditorium - 1012 Sample Point ID: West Lobby @ICPMS	
<u>201702210635</u>	1001-080	02/17/2017 0539
	Sample Type: Tap Facility ID: Cafeteria - 1013 Sample Point ID: Kitchen (fix. by back entrance) @ICPMS	
<u>201702210636</u>	1001-081	02/17/2017 0541
	Sample Type: Drinking Fountain Facility ID: Cafeteria - 1013 Sample Point ID: N. Dining RM Wall @ICPMS	
<u>201702210637</u>	1001-082	02/17/2017 0543
	Sample Type: Drinking Fountain Facility ID: VOC Shops - 1014 Sample Point ID: Wst Wall - Early Chld hood dep @ICPMS	
<u>201702210638</u>	1001-083	02/17/2017 0544
	Sample Type: Drinking Fountain Facility ID: VOC Shops - 1014 Sample Point ID: Print Shop @ICPMS	
<u>201702210639</u>	1001-084	02/17/2017 0546
	Sample Type: Drinking Fountain Facility ID: Girls' P.E. - 1015 Sample Point ID: by Shower RM entrance	

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
	@ICPMS	
<u>201702210640</u>	1001-085	02/17/2017 0547
	Sample Type: Tap Facility ID: Girls' P.E. - 1015 Sample Point ID: West Fix. by showers	
	@ICPMS	
<u>201702210641</u>	1001-086	02/17/2017 0549
	Sample Type: Drinking Fountain Facility ID: Gymnasium - 1016 Sample Point ID: Lobby South	
	@ICPMS	
<u>201702210642</u>	1001-087	02/17/2017 0550
	Sample Type: Drinking Fountain Facility ID: Gymnasium - 1016 Sample Point ID: Lobby North	
	@ICPMS	
<u>201702210643</u>	1001-088	02/17/2017 0555
	Sample Type: Drinking Fountain Facility ID: Auto Shop - 1017 Sample Point ID: Entrance Hall	
	@ICPMS	
<u>201702210644</u>	1001-089	02/17/2017 0556
	Sample Type: Tap Facility ID: Auto Shop - 1017 Sample Point ID: Entrance Hall	
	@ICPMS	
<u>201702210645</u>	1001-090	02/17/2017 0600
	Sample Type: Drinking Fountain Facility ID: Library - 1018 Sample Point ID: by Restrooms	
	@ICPMS	
<u>201702210646</u>	1001-091	02/17/2017 0601
	Sample Type: Tap Facility ID: Library - 1018 Sample Point ID: Work Room	
	@ICPMS	
<u>201702210647</u>	1001-092	02/17/2017 0606

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
	Sample Type: Drinking Fountain Facility ID: Main Sto - 1019 Sample Point ID: Clock RM @ICPMS	
<u>201702210648</u>	1001-093	02/17/2017 0608
	Sample Type: Drinking Fountain Facility ID: Main Sto - 1019 Sample Point ID: Wood Shop @ICPMS	
<u>201702210649</u>	1001-094	02/17/2017 0612
	Sample Type: Drinking Fountain Facility ID: AG Shop - 1021 Sample Point ID: Welding Shop North @ICPMS	
<u>201702210650</u>	1001-095	02/17/2017 0617
	Sample Type: Drinking Fountain Facility ID: Wellness Center - 1022 Sample Point ID: Main Hall @ICPMS	
<u>201702210651</u>	1001-096	02/17/2017 0620
	Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: South West (ext.) @ICPMS	
<u>201702210652</u>	1001-097	02/17/2017 0622
	Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: East Wing @ICPMS	
<u>201702210653</u>	1001-098	02/17/2017 0624
	Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: West Wing @ICPMS	

Test Description

@ICPMS -- ICPMS Metals

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 640298
Project: DOUGLAS-USD
Sample Group: Douglas High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Alfredo Garcia

The following samples were received from you on **February 21, 2017** at **1130**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.

Sample #	Sample ID	Sample Date
----------	-----------	-------------


Collection Log

for experienced sample collectors


640298
MAR

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	1 st Wing WEST DECCASTORE 1001
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	RM 101
Sample Identification Number (Write the number here and on sample container)	1001-055
Date of Collection	2-17-2017
Time of Collection	5:00 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School (District 27)
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	1 st Wing W - 1001
Type of Fixture (tap, drinking fountain etc.)	D. F.
Location of Fixture (example, room number)	1 st Wing West (ext.)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1001-056
Date of Collection	2-17-2017
Time of Collection	5:01 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	1 st Wing East - 1002
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	1002 East (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1002-057
Date of Collection	2-17-2017
Time of Collection	5:02 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	3/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	1 st Wing - 1002
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	NURSE RM.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1002-058
Date of Collection	2-17-2017
Time of Collection	5:03 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	1 st Wing - 1002
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	Principal's Lobby
Sample Identification Number (Write the number here and on sample container)	1002-059
Date of Collection	2-17-2017
Time of Collection	5:04 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	1 st Wing - 1002
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	TEACHERS' Lounge.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1002-060
Date of Collection	2-17-2017
Time of Collection	5:05 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	2 nd Wing - 1003
Type of Fixture (tap, drinking fountain etc.)	D. F.
Location of Fixture (example, room number)	1003 west (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1003-061
Date of Collection	2-17-2017
Time of Collection	5:07 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17, 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	2 nd Wing - 1003
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	RM 201
Sample Identification Number (<u>Write the number here and on sample container</u>)	1003-062
Date of Collection	2-17-2017
Time of Collection	5:08 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/24/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	2 nd Wing - 1004
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	1004 EAST (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1004-063
Date of Collection	2-17-2017
Time of Collection	5:09 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 W3U
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	2nd Wing - 1004
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	East Custodian Rm.
Sample Identification Number (Write the number here and on sample container)	1004-064
Date of Collection	2-17-2017
Time of Collection	5:10 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 11:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	3rd Wing - 1005
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	Rm 303
Sample Identification Number (<u>Write the number here and on sample container</u>)	1005-065
Date of Collection	2-17-2017
Time of Collection	5:12 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	3 RD WING - 1005
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	RM 305
Sample Identification Number (<u>Write the number here and on sample container</u>)	1005-066
Date of Collection	2-17-2017
Time of Collection	5:13 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	3RD WING - 1006
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	South wall, west fixture
Sample Identification Number (<u>Write the number here and on sample container</u>)	1006-067
Date of Collection	2-17-2017
Time of Collection	5:14 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	3 RD Wing - 1006
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	South wall, East fixture
Sample Identification Number (<u>Write the number here and on sample container</u>)	1006 - 068
Date of Collection	2-17-2017
Time of Collection	5:15 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	3 rd Wing - 1006
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	Rm 302
Sample Identification Number (<u>Write the number here and on sample container</u>)	1006-069
Date of Collection	2-17-2017
Time of Collection	5:16 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	3 rd Wing - 1006
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	RM 308
Sample Identification Number (<u>Write the number here and on sample container</u>)	1006 - 070
Date of Collection	2-17-2017
Time of Collection	5:18 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 11:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	3 rd Wing - 1006
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	1006 EAST (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1006-071
Date of Collection	2-17-2017
Time of Collection	5:20 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	4 th Wing-1007
Type of Fixture (tap, drinking fountain etc.)	D. F
Location of Fixture (example, room number)	1007-West (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1007-072
Date of Collection	2-17-2017
Time of Collection	5:22 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	4 th Wing - 1008
Type of Fixture (tap, drinking fountain etc.)	D. F.
Location of Fixture (example, room number)	1008 EAST (ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1008-073
Date of Collection	2-17-2017
Time of Collection	5:23 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	5 th wing - 1009
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	1009 EAST (ext)
Sample Identification Number (Write the number here and on sample container)	1009-074
Date of Collection	2-17-2017
Time of Collection	5:25 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/20/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	5 th Wing - 1009
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	EAST CUSTODIAN RM.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1009-075
Date of Collection	2-17-2017
Time of Collection	5:26 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/13/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	6 th Wing - 1011
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	1011 East (ext) by Rm 608
Sample Identification Number (Write the number here and on sample container)	1011-076
Date of Collection	2-17-2017
Time of Collection	5:28 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	6 th WING - 1011
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	Rm 608, 1 st fixture by door
Sample Identification Number (<u>Write the number here and on sample container</u>)	1011-077
Date of Collection	2-17-2017
Time of Collection	5:30AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	Auditorium - 1012
Type of Fixture (tap, drinking fountain etc.)	D.F. case
Location of Fixture (example, room number)	EAST LOBBY
Sample Identification Number (<u>Write the number here and on sample container</u>)	1012-078
Date of Collection	2-17-2017
Time of Collection	5:35 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 W30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	Auditorium - 1012
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	West Lobby
Sample Identification Number (<u>Write the number here and on sample container</u>)	1012-079
Date of Collection	2-17-2017
Time of Collection	5:36 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	CAFETERIA - 1013
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	KITCHEN (fixture by Back entrance)
Sample Identification Number (Write the number here and on sample container)	1013-080
Date of Collection	2-17-2017
Time of Collection	5:39 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 6:50
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	CAFETERIA - 1013
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	N. DINING RM WALL
Sample Identification Number (<u>Write the number here and on sample container</u>)	1013-081
Date of Collection	2-17-2017
Time of Collection	5:41 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	VOC SHOPS - 1014
Type of Fixture (tap, drinking fountain etc.)	D. F.
Location of Fixture (example, room number)	WEST WALL - Early Childhood Development
Sample Identification Number (<u>Write the number here and on sample container</u>)	1014-082
Date of Collection	2-17-2017
Time of Collection	5:43 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1730
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	VOC SHOPS - 1014
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	PRINT Shop
Sample Identification Number (<u>Write the number here and on sample container</u>)	1014-083
Date of Collection	2-17-2017
Time of Collection	5:44 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 11:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	GIRLS' P.E. - 1015
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	by Shower Rm. Entrance
Sample Identification Number (<u>Write the number here and on sample container</u>)	1015 - 084
Date of Collection	2-17-2017
Time of Collection	5:46 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/19/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	GIRLS' P.E. - 1015
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	West fixture by Showers
Sample Identification Number (<u>Write the number here and on sample container</u>)	1015 - 085
Date of Collection	2-17-2017
Time of Collection	5:47 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	GYMNASIUM - 1016
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	LOBBY South.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1016-086
Date of Collection	2-17-2017
Time of Collection	5:49 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	GYMNASIUM - 1016
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	LOBBY NORTH
Sample Identification Number (<u>Write the number here and on sample container</u>)	1016-087
Date of Collection	2-17-2017
Time of Collection	5:50 A.M.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	7/2/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	AUTO SHOP - 1017
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	ENTRANCE HALL
Sample Identification Number (<u>Write the number here and on sample container</u>)	1017-088
Date of Collection	2-17-2017
Time of Collection	5:55 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	Auto Shop - 1017
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	ENTRANCE HALL
Sample Identification Number (<u>Write the number here and on sample container</u>)	1017-089
Date of Collection	2-17-2017
Time of Collection	5:56 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	LIBRARY - 1018
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	by RESTROOMS.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1018-090
Date of Collection	2-17-2017
Time of Collection	6:00 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	LIBRARY - 1018
Type of Fixture (tap, drinking fountain etc.)	TAP
Location of Fixture (example, room number)	WORK ROOM
Sample Identification Number (<u>Write the number here and on sample container</u>)	1018-091
Date of Collection	2-17-2017
Time of Collection	6:01 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	MAINT Sto - 1019
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	CLOCK RM.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1019-092
Date of Collection	2-17-2017
Time of Collection	6:06 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	MAINT ST - 1019
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	WOOD SHOP
Sample Identification Number (<u>Write the number here and on sample container</u>)	1019-093
Date of Collection	2-17-2017
Time of Collection	6:08 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS High School)
Building (name/number)	AGSHOP - 1021
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	Welding Shop NORTH.
Sample Identification Number (<u>Write the number here and on sample container</u>)	1021-094
Date of Collection	2-17-2017
Time of Collection	6:12 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS UNIFIED School District 27
School Name	DHS (DOUGLAS HIGH School)
Building (name/number)	WELLNESS CENTER - 1022
Type of Fixture (tap, drinking fountain etc.)	DF
Location of Fixture (example, room number)	MAIN HALL
Sample Identification Number (<u>Write the number here and on sample container</u>)	1022-095
Date of Collection	2-17-2017
Time of Collection	6:17 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	Field House - 1023
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	South West (Ext)
Sample Identification Number (<u>Write the number here and on sample container</u>)	1023-096
Date of Collection	2-17-2017
Time of Collection	6:20 AM
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/21/17 1130
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	Fieldhouse - 1023
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	EAST wing
Sample Identification Number (<u>Write the number here and on sample container</u>)	1023-097
Date of Collection	2-17-2017
Time of Collection	6:22 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	DOUGLAS Unified School District 27
School Name	DHS (DOUGLAS HIGH SCHOOL)
Building (name/number)	Fieldhouse - 1023
Type of Fixture (tap, drinking fountain etc.)	D.F.
Location of Fixture (example, room number)	West Wing
Sample Identification Number (<u>Write the number here and on sample container</u>)	1023-098
Date of Collection	2-17-2017
Time of Collection	6:24 AM.
Printed Name of Sample Collector	Alfredo Garcia
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	2/17/17 1:30
Signature	
Notes:	

For relinquishing samples upon delivery to labs only










Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

DHS

1500 15th street


Mult. STO
1019
MAIN+ STO
1019



FLOOR PLAN


08/11/2019 10:10:10

DHS AG Shop



FLOOR PLAN

AGSHOPSKETUPAPQ52413



FIRE ALARM

AGSHOPSKETUPAPQ52413

Douglas Unified District	Douglas High School	1001	1 st Wing West • DECA • DF (EXT)	Classrooms	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1002	1 st Wing Admin East • DECA • DF EXT • T.Wing • NURSE Principal • NURSE Assistant	Admin & Ancillary	1949	P O Box 1237	Douglas	85607-	Cochise	4
Douglas Unified District	Douglas High School	1003	2 nd Wing West • DE	Classrooms	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1004	2 nd Wing East • DE	Classrooms	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1005	3 rd Wing West • TAP RM 303 • TAP RM 305	Classrooms	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1006	3 rd Wing East • TAP • SP. Needs 308 • Cul Arts (302) TAP X3 • DE SP. Needs (306) 2F	Classrooms	1949	P O Box 1237	Douglas	85607-	Cochise	35
Douglas Unified District	Douglas High School	1007	4 th Wing West • DE • SOURCE	Classrooms	1959	P O Box 1237	Douglas	85607-	Cochise	1
Douglas Unified District	Douglas High School	1008	4 th Wing East • DE • SOURCE	Classrooms	1959	P O Box 1237	Douglas	85607-	Cochise	1
Douglas Unified District	Douglas High School	1009	5 th Wing 1 st Floor • DE East • TAP East Cust. Rm.	Classrooms	1964	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1010	5 th Wing 2 nd Floor • NO SOURCE	Classrooms	1964	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	1011	6 th Wing • DE by 608 • RM 608 TAP	Classrooms	1995	P O Box 1237	Douglas	85607-	Cochise	4

23

Douglas Unified District	Douglas High School	✓ 1012	Auditorium Lobby • DFE L • DF W	Auditorium & Music	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1013	Cafeteria. • Kit • TAP M • dining DF	Cafeteria	1959	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1014	VOC Shops • ECD DF N • PRINT DF	Shops	1959	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1015	GIRLS PE • DF O • TAP	Girls PE	1982	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1016	GYM • LOBBY DF S P • LOBBY DF N	Gymnasium	1949	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1017	Auto Shop • DF Q • TAP	Auto Shop	1976	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1018	LIBRARY • DF R • W.R.	Library	1995	P O Box 1237	Douglas	85607-	Cochise	± 2
Douglas Unified District	Douglas High School	✓ 1019	WAREHOUSE • DF S • woodshop DF	Maintenance storage	1954	P O Box 1237	Douglas	85607-	Cochise	2
Douglas Unified District	Douglas High School	✓ 1020	MODULAR 400 No SOURCE Modular	Office and classroom	1982	P O Box 1237	Douglas	85607-	Cochise	0
Douglas Unified District	Douglas High School	✓ 1021	• DF AG Shop	Classroom	2004	P O Box 1237	Douglas	85607-	Cochise	1
Douglas Unified District	Douglas High School	✓ 1022	• DF • TAP UNISEX RR Wellness Center	Offices and classrooms	1965	P O Box 1237	Douglas	85607-	Cochise	1

Douglas Unified District	Douglas High School	1023	DF East DF West FIELDHOUSE	STUDENT LOCKERS/SHOWERS	2011	P O Box 1237	Douglas	85607-	Cochise	3 1
Douglas Unified District	Douglas High School	1024	NO SOURCE Modular (Room 601/603)	Classroom	1982	P O Box 1237	Douglas	85607-	Cochise	2 0
Total Containers										44

3

40

44

40

23
19
44

3

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
(626) 386-1100 FAX (626) 386-1101

Kit #: 158129



Created By: Thomas.D.French - [TDF]
Deliver By: 02/06/2017

STG: Bottle Orders

Ice Type: W

Note: Sampler Please return this paper with your samples

Client ID: ADEQ-LEAD
Project Code: DOUGLAS-USD Bottle Orders
Group Name: Douglas High School
PO#/JOB#:

Ship Sample Kits to Douglas Unified School District #27 1132 E. 12th Street Douglas, AZ 85607 Attn: Cesar B. Soto/Business Mgr. District Facilities Phone: (520) 364-2447
--

Send Report to Arizona Department of Environmental Quality 1110 West Washington Street Phoenix, AZ 85007 Attn: David Burchard Phone: (602) 771-4298
--

Billing Address Arizona Department of Environmental Quality 1110 West Washington Street Phoenix, AZ 85007 Attn: ADEQ Phone: (602) 771-1936

of

Sample Tests

44

@ICPMS

Bottle Qty - Type [preservative information]

1 - 250ml poly [no preservative]

UN DOT #

Comments

Douglas High School - Include Douglas Unified School District #27 Full Sample list, 44 Lead Sampling Plan Records, packing instructions for return shipment to Eurofins Eaton Analytical, Inc. 750 Royal Oaks Drive, Suite C, Monrovia, CA 91016. 44 sample containers. Return Shipment Fed EX

Sampler - please refer to Sampling Plan Records for instructions on completing paperwork and what to include with return shipment of the samples.

Code

Status

Date Shipped

Via

Tracking #

of Coolers

Prepared By

INTERNAL CHAIN OF CUSTODY RECORD



Eaton Analytical

EEA Folder Number:

SAMPLES REC'D DAY OF COLLECTION? ☐

SAMPLE TEMP RECEIVED:

IR Gun ID = 518 (Observation = 17.7 °C) (Corr. Factor = -0.1 °C) (Final = 17.6 °C)

TYPE OF ICE: Real ☐ Synthetic ☒ No Ice ☒ CONDITION OF ICE: Frozen ☐ Partially Frozen ☐ Thawed ☒ N/A ☒

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx ☒ UPS / DHL / Area Fast / Top Line / Other: _____

TRK# 7204 7946 1443

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

4) UCMR3: 524.3: (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)
(non-GLEC) 522: (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

≤ 10°C if received within 48 hours of sample collection (not the same business day); ≤ 6°C if received after 48 hours of sample collection. Measure temperature for each method above.

5) LT2: Giardia /Cryptosporidium: <20 °C, not frozen (received after 8 hours of sample collection)

E. Coli: < 10°C, not frozen (if received after 2 hours of sample collection)

Giardia/Crypto: (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

E.Coli: (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

6) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

RECEIVED BY: <u>[Signature]</u>	PRINT NAME: <u>TE</u>	COMPANY/TITLE: Eurofins Eaton Analytical	DATE: <u>2/21/17</u>	TIME: <u>1130</u>
---------------------------------	-----------------------	--	----------------------	-------------------

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Comments
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Hits
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
02/22/2017 22:16	Lead Total ICAP/MS	<u>201702210615</u> <u>1001-060</u>	0.74	15	ug/L	0.5
02/22/2017 22:17	Lead Total ICAP/MS	<u>201702210617</u> <u>1001-062</u>	2.1	15	ug/L	0.5
02/22/2017 22:20	Lead Total ICAP/MS	<u>201702210620</u> <u>1001-065</u>	3.4	15	ug/L	0.5
02/22/2017 22:24	Lead Total ICAP/MS	<u>201702210621</u> <u>1001-066</u>	2.6	15	ug/L	0.5
02/22/2017 22:25	Lead Total ICAP/MS	<u>201702210622</u> <u>1001-067</u>	1.3	15	ug/L	0.5
02/22/2017 22:25	Lead Total ICAP/MS	<u>201702210623</u> <u>1001-068</u>	1.9	15	ug/L	0.5
02/22/2017 22:26	Lead Total ICAP/MS	<u>201702210624</u> <u>1001-069</u>	1.7	15	ug/L	0.5
02/22/2017 22:28	Lead Total ICAP/MS	<u>201702210626</u> <u>1001-071</u>	0.68	15	ug/L	0.5
02/22/2017 22:36	Lead Total ICAP/MS	<u>201702210630</u> <u>1001-075</u>	0.89	15	ug/L	0.5
02/22/2017 22:42	Lead Total ICAP/MS	<u>201702210634</u> <u>1001-079</u>	3.7	15	ug/L	0.5
02/22/2017 22:53	Lead Total ICAP/MS	<u>201702210644</u> <u>1001-089</u>	1.2	15	ug/L	0.5
02/22/2017 22:54	Lead Total ICAP/MS	<u>201702210646</u> <u>1001-091</u>	7.9	15	ug/L	0.5

SUMMARY OF POSITIVE DATA ONLY

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
1001-055 (201702210610)						Sampled on 02/17/2017 0500			
Sample Type: Tap Facility ID: 1st Wing West Deccastore 1001 Sample Point ID: RM 101									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:09	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-056 (201702210611)						Sampled on 02/17/2017 0501			
Sample Type: Drinking Fountain Facility ID: 1st Wing W-1001 Sample Point ID: 1st Wing West (ext.)									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:13	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-057 (201702210612)						Sampled on 02/17/2017 0502			
Sample Type: Drinking Fountain Facility ID: 1st Wing East - 1002 Sample Point ID: 1002 East (ext.)									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:14	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-058 (201702210613)						Sampled on 02/17/2017 0503			
Sample Type: Tap Facility ID: 1st Wing - 1002 Sample Point ID: Nurse RM									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:14	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-059 (201702210614)						Sampled on 02/17/2017 0504			
Sample Type: Drinking Fountain Facility ID: 1st Wing - 1002 Sample Point ID: Principal's Lobby									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:15	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-060 (201702210615)						Sampled on 02/17/2017 0505			
Sample Type: Tap Facility ID: 1st Wing - 1002 Sample Point ID: Teachers' Lounge									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:16	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	0.74	ug/L	0.5	1
1001-061 (201702210616)						Sampled on 02/17/2017 0507			
Sample Type: Drinking Fountain Facility ID: 2nd Wing - 1003 Sample Point ID: 1003 West (ext.)									
EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:17	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-062 (201702210617)						Sampled on 02/17/2017 0508			

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Tap Facility ID: 2nd Wing - 1003 Sample Point ID: RM 201 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:17	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	2.1	ug/L	0.5	1
1001-063 (201702210618)						Sampled on 02/17/2017 0509			
Sample Type: Drinking Fountain Facility ID: 2nd Wing - 1004 Sample Point ID: 1004 East (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:18	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-064 (201702210619)						Sampled on 02/17/2017 0510			
Sample Type: Tap Facility ID: 2nd Wing - 1004 Sample Point ID: East Custodian RM EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:19	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1001-065 (201702210620)						Sampled on 02/17/2017 0512			
Sample Type: Tap Facility ID: 3rd Wing - 1005 Sample Point ID: RM 303 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:20	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	3.4	ug/L	0.5	1
1001-066 (201702210621)						Sampled on 02/17/2017 0513			
Sample Type: Tap Facility ID: 3rd Wing - 1005 Sample Point ID: RM 305 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:24	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	2.6	ug/L	0.5	1
1001-067 (201702210622)						Sampled on 02/17/2017 0514			
Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: South Wall, West Fixture EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:25	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	1.3	ug/L	0.5	1
1001-068 (201702210623)						Sampled on 02/17/2017 0515			
Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: South Wall, East Fixture EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:25	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	1.9	ug/L	0.5	1
1001-069 (201702210624)						Sampled on 02/17/2017 0516			

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: 3rd Wing - 1006 Sample Point ID: RM 302 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:26	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	1.7	ug/L	0.5	1
<u>1001-070 (201702210625)</u>						Sampled on 02/17/2017 0518			
Sample Type: Tap Facility ID: 3rd Wing - 1006 Sample Point ID: RM 308 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:27	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-071 (201702210626)</u>						Sampled on 02/17/2017 0520			
Sample Type: Drinking Fountain Facility ID: 3rd Wing - 1006 Sample Point ID: 1006 - East (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:28	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	0.68	ug/L	0.5	1
<u>1001-072 (201702210627)</u>						Sampled on 02/17/2017 0522			
Sample Type: Drinking Fountain Facility ID: 4th Wing - 1007 Sample Point ID: 1007 - West (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:28	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-073 (201702210628)</u>						Sampled on 02/17/2017 0523			
Sample Type: Drinking Fountain Facility ID: 4th Wing - 1008 Sample Point ID: 1008 - East (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:29	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-074 (201702210629)</u>						Sampled on 02/17/2017 0525			
Sample Type: Drinking Fountain Facility ID: 5th Wing - 1009 Sample Point ID: 1009 - East (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:32	972305	972352	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-075 (201702210630)</u>						Sampled on 02/17/2017 0526			
Sample Type: Tap Facility ID: 5th Wing - 1009 Sample Point ID: East Custodian RM EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:36	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	0.89	ug/L	0.5	1
<u>1001-076 (201702210631)</u>						Sampled on 02/17/2017 0528			

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: 6th Wing - 1011 Sample Point ID: 1011 East (ext.) by RM 608 EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:38	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-077 (201702210632)</u>						Sampled on 02/17/2017 0530			
Sample Type: Tap Facility ID: 6th Wing - 1011 Sample Point ID: RM 608, 1st Fixture by door EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:39	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-078 (201702210633)</u>						Sampled on 02/17/2017 0535			
Sample Type: Drinking Fountain Facility ID: Auditorium - 1012 Sample Point ID: East Lobby EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:41	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-079 (201702210634)</u>						Sampled on 02/17/2017 0536			
Sample Type: Drinking Fountain Facility ID: Auditorium - 1012 Sample Point ID: West Lobby EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:42	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	3.7	ug/L	0.5	1
<u>1001-080 (201702210635)</u>						Sampled on 02/17/2017 0539			
Sample Type: Tap Facility ID: Cafeteria - 1013 Sample Point ID: Kitchen (fix. by back entrance) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:43	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-081 (201702210636)</u>						Sampled on 02/17/2017 0541			
Sample Type: Drinking Fountain Facility ID: Cafeteria - 1013 Sample Point ID: N. Dining RM Wall EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:43	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-082 (201702210637)</u>						Sampled on 02/17/2017 0543			
Sample Type: Drinking Fountain Facility ID: VOC Shops - 1014 Sample Point ID: Wst Wall - Early Chld hood dep EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:44	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-083 (201702210638)</u>						Sampled on 02/17/2017 0544			

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: VOC Shops - 1014 Sample Point ID: Print Shop EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:45	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-084 (201702210639)</u>						Sampled on 02/17/2017 0546			
Sample Type: Drinking Fountain Facility ID: Girls' P.E. - 1015 Sample Point ID: by Shower RM entrance EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:46	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-085 (201702210640)</u>						Sampled on 02/17/2017 0547			
Sample Type: Tap Facility ID: Girls' P.E. - 1015 Sample Point ID: West Fix. by showers EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:47	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-086 (201702210641)</u>						Sampled on 02/17/2017 0549			
Sample Type: Drinking Fountain Facility ID: Gymnasium - 1016 Sample Point ID: Lobby South EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:51	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-087 (201702210642)</u>						Sampled on 02/17/2017 0550			
Sample Type: Drinking Fountain Facility ID: Gymnasium - 1016 Sample Point ID: Lobby North EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:51	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-088 (201702210643)</u>						Sampled on 02/17/2017 0555			
Sample Type: Drinking Fountain Facility ID: Auto Shop - 1017 Sample Point ID: Entrance Hall EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:52	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-089 (201702210644)</u>						Sampled on 02/17/2017 0556			
Sample Type: Tap Facility ID: Auto Shop - 1017 Sample Point ID: Entrance Hall EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:53	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	1.2	ug/L	0.5	1
<u>1001-090 (201702210645)</u>						Sampled on 02/17/2017 0600			

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: Library - 1018 Sample Point ID: by Restrooms EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:54	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-091 (201702210646)</u>						Sampled on 02/17/2017 0601			
Sample Type: Tap Facility ID: Library - 1018 Sample Point ID: Work Room EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:54	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	7.9	ug/L	0.5	1
<u>1001-092 (201702210647)</u>						Sampled on 02/17/2017 0606			
Sample Type: Drinking Fountain Facility ID: Main Sto - 1019 Sample Point ID: Clock RM EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:55	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-093 (201702210648)</u>						Sampled on 02/17/2017 0608			
Sample Type: Drinking Fountain Facility ID: Main Sto - 1019 Sample Point ID: Wood Shop EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:56	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-094 (201702210649)</u>						Sampled on 02/17/2017 0612			
Sample Type: Drinking Fountain Facility ID: AG Shop - 1021 Sample Point ID: Welding Shop North EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 22:57	972305	972353	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-095 (201702210650)</u>						Sampled on 02/17/2017 0617			
Sample Type: Drinking Fountain Facility ID: Wellness Center - 1022 Sample Point ID: Main Hall EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 23:02	972305	972354	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-096 (201702210651)</u>						Sampled on 02/17/2017 0620			
Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: South West (ext.) EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 23:05	972305	972354	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-097 (201702210652)</u>						Sampled on 02/17/2017 0622			

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 640298

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
02/21/2017 1130

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: East Wing EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 23:05	972305	972354	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1001-098 (201702210653)</u> Sample Type: Drinking Fountain Facility ID: Field House - 1023 Sample Point ID: West Wing EPA 200.8 - ICPMS Metals									
02/21/17	02/22/17 23:06	972305	972354	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1

Sampled on 02/17/2017 0624

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality

ICPMS Metals

Prep Batch: 972305 Analytical Batch: 972352

201702210610	1001-055
201702210611	1001-056
201702210612	1001-057
201702210613	1001-058
201702210614	1001-059
201702210615	1001-060
201702210616	1001-061
201702210617	1001-062
201702210618	1001-063
201702210619	1001-064
201702210620	1001-065
201702210621	1001-066
201702210622	1001-067
201702210623	1001-068
201702210624	1001-069
201702210625	1001-070
201702210626	1001-071
201702210627	1001-072
201702210628	1001-073
201702210629	1001-074

Analysis Date: 02/22/2017[illegible]

ICPMS Metals

Prep Batch: 972305 Analytical Batch: 972353

201702210630	1001-075
201702210631	1001-076
201702210632	1001-077
201702210633	1001-078
201702210634	1001-079
201702210635	1001-080
201702210636	1001-081
201702210637	1001-082
201702210638	1001-083
201702210639	1001-084
201702210640	1001-085
201702210641	1001-086
201702210642	1001-087
201702210643	1001-088
201702210644	1001-089
201702210645	1001-090
201702210646	1001-091
201702210647	1001-092
201702210648	1001-093
201702210649	1001-094

Analysis Date: 02/22/2017[illegible]

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality

ICPMS Metals

Prep Batch: 972305 Analytical Batch: 972354

201702210650	1001-095
201702210651	1001-096
201702210652	1001-097
201702210653	1001-098

Analysis Date: 02/22/2017

Analyzed by: CYP
Analyzed by: CYP
Analyzed by: CYP
Analyzed by: CYP

Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
ICPMS Metals by EPA 200.8									
Analytical Batch: 972352					Analysis Date: 02/22/2017				
LCS1	Lead Total ICAP/MS		20	20.1	ug/L	100	(85-115)		
LCS2	Lead Total ICAP/MS		20	20.1	ug/L	100	(85-115)	20	0.0
MBLK	Lead Total ICAP/MS			<0.5	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.501	ug/L	100	(50-150)		
MS_201702210610	Lead Total ICAP/MS	ND	20	21.4	ug/L	107	(70-130)		
MS2_201702210620	Lead Total ICAP/MS	3.4	20	24.7	ug/L	106	(70-130)		
MSD_201702210610	Lead Total ICAP/MS	ND	20	21.4	ug/L	107	(70-130)	20	0.0
MSD2_201702210620	Lead Total ICAP/MS	3.4	20	24.8	ug/L	107	(70-130)	20	0.81
ICPMS Metals by EPA 200.8									
Analytical Batch: 972353					Analysis Date: 02/22/2017				
LCS1	Lead Total ICAP/MS		20	20.1	ug/L	101	(85-115)		
LCS2	Lead Total ICAP/MS		20	20.4	ug/L	102	(85-115)	20	1.5
MBLK	Lead Total ICAP/MS			<0.5	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.501	ug/L	100	(50-150)		
MS_201702210630	Lead Total ICAP/MS	0.89	20	22.6	ug/L	108	(70-130)		
MS2_201702210640	Lead Total ICAP/MS	ND	20	21.2	ug/L	106	(70-130)		
MSD_201702210630	Lead Total ICAP/MS	0.89	20	21.6	ug/L	104	(70-130)	20	4.5
MSD2_201702210640	Lead Total ICAP/MS	ND	20	21.4	ug/L	107	(70-130)	20	0.94
ICPMS Metals by EPA 200.8									
Analytical Batch: 972354					Analysis Date: 02/22/2017				
LCS1	Lead Total ICAP/MS		20	20.4	ug/L	102	(85-115)		
LCS2	Lead Total ICAP/MS		20	20.3	ug/L	102	(85-115)	20	0.49
MBLK	Lead Total ICAP/MS			<0.5	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.512	ug/L	102	(50-150)		
MS_201702210650	Lead Total ICAP/MS	ND	20	21.6	ug/L	108	(70-130)		
MSD_201702210650	Lead Total ICAP/MS	ND	20	21.7	ug/L	109	(70-130)	20	0.46

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.