

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Phoenix

4625 East Cotton Ctr Blvd

Suite 189

Phoenix, AZ 85040

Tel: (602)437-3340

TestAmerica Job ID: 550-79969-1

TestAmerica Sample Delivery Group: Page High School

Client Project/Site: Pb in DW / Page Unified School District

For:

Arizona Dept of Environmental Quality

1110 W. Washington St., MC5415B-1

Phoenix, Arizona 85007

Attn: David A Burchard



Authorized for release by:

4/10/2017 3:09:36 PM

Jannel Franklin, Project Manager I

(732)593-2551

jannel.franklin@testamericainc.com

Designee for

Susan Armijo, Project Manager II

(602)659-7664

susan.armijo@testamericainc.com

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

1

2

3

4

5

6

7

8

9

10

11

12

13

14



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Sample Summary	5
Detection Summary	6
Client Sample Results	8
QC Sample Results	11
QC Association Summary	12
Lab Chronicle	13
Certification Summary	17
Method Summary	18
Chain of Custody	19
Receipt Checklists	49

Definitions/Glossary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Job ID: 550-79969-1

Laboratory: TestAmerica Phoenix

Narrative

Job Narrative 550-79969-1

Comments

No additional comments.

Receipt

The samples were received on 3/29/2017 10:30 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 19.5° C.

Receipt Exceptions

The following samples for metals were received unpreserved and were preserved upon receipt to the laboratory: A Building drinking fountain (550-79969-1), PHS 1002-2 (550-79969-2), PHS 1003-1 Kitchen Sink (550-79969-3), PHS 1003-2 Westside (550-79969-4), Main Building East PHS 1004-1 (550-79969-5), North Computer Lab - PHS 1004-2 (550-79969-6), West Girls RR bottle #3 PHS 1004-3 (550-79969-7), PHS 1005-1 (550-79969-8), PHS 1006-1 VOC trades (550-79969-9), PHS - 1007-1 (550-79969-10), PHS 1007-2 East Gym H.S. (550-79969-11), PHS 1008-4 Pool Gym (550-79969-12), PHS 1010-1 Lobby CAB (550-79969-13), PHS 1010-2 Band Room drinking fountain (550-79969-14), PHS - 1011-1 Gym Lobby (550-79969-15), PHS 1015-1 (550-79969-16), DO-1 (550-79969-17), DO-2 (550-79969-18) and PHS 1018-1 (550-79969-19). Regulatory documents require a 24-hour waiting period from the time of the addition of the acid preservative to the time of digestion.

Metals

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

Sample Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
550-79969-1	A Building drinking fountain	Drinking Water	03/23/17 10:37	03/29/17 10:30
550-79969-2	PHS 1002-2	Drinking Water	03/23/17 10:38	03/29/17 10:30
550-79969-3	PHS 1003-1 Kitchen Sink	Drinking Water	03/23/17 10:53	03/29/17 10:30
550-79969-4	PHS 1003-2 Westside	Drinking Water	03/23/17 10:49	03/29/17 10:30
550-79969-5	Main Building East PHS 1004-1	Drinking Water	03/23/17 10:08	03/29/17 10:30
550-79969-6	North Computer Lab - PHS 1004-2	Drinking Water	03/23/17 10:11	03/29/17 10:30
550-79969-7	West Girls RR bottle #3 PHS 1004-3	Drinking Water	03/23/17 10:15	03/29/17 10:30
550-79969-8	PHS 1005-1	Drinking Water	03/23/17 10:26	03/29/17 10:30
550-79969-9	PHS 1006-1 VOC trades	Drinking Water	03/23/17 10:30	03/29/17 10:30
550-79969-10	PHS - 1007-1	Drinking Water	03/23/17 11:06	03/29/17 10:30
550-79969-11	PHS 1007-2 East Gym H.S.	Drinking Water	03/23/17 11:07	03/29/17 10:30
550-79969-12	PHS 1008-4 Pool Gym	Drinking Water	03/23/17 11:10	03/29/17 10:30
550-79969-13	PHS 1010-1 Lobby CAB	Drinking Water	03/23/17 10:43	03/29/17 10:30
550-79969-14	PHS 1010-2 Band Room drinking fountain	Drinking Water	03/23/17 10:43	03/29/17 10:30
550-79969-15	PHS - 1011-1 Gym Lobby	Drinking Water	03/23/17 11:08	03/29/17 10:30
550-79969-16	PHS 1015-1	Drinking Water	03/23/17 10:26	03/29/17 10:30
550-79969-17	DO-1	Drinking Water	03/23/17 08:05	03/29/17 10:30
550-79969-18	DO-2	Drinking Water	03/23/17 08:10	03/29/17 10:30
550-79969-19	PHS 1018-1	Drinking Water	03/23/17 10:31	03/29/17 10:30

Detection Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Client Sample ID: A Building drinking fountain

Lab Sample ID: 550-79969-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.53		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1002-2

Lab Sample ID: 550-79969-2

No Detections.

Client Sample ID: PHS 1003-1 Kitchen Sink

Lab Sample ID: 550-79969-3

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.9		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1003-2 Westside

Lab Sample ID: 550-79969-4

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.7		0.50	ppb	1		200.8	Total/NA

Client Sample ID: Main Building East PHS 1004-1

Lab Sample ID: 550-79969-5

No Detections.

Client Sample ID: North Computer Lab - PHS 1004-2

Lab Sample ID: 550-79969-6

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.76		0.50	ppb	1		200.8	Total/NA

Client Sample ID: West Girls RR bottle #3 PHS 1004-3

Lab Sample ID: 550-79969-7

No Detections.

Client Sample ID: PHS 1005-1

Lab Sample ID: 550-79969-8

No Detections.

Client Sample ID: PHS 1006-1 VOC trades

Lab Sample ID: 550-79969-9

No Detections.

Client Sample ID: PHS - 1007-1

Lab Sample ID: 550-79969-10

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.84		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1007-2 East Gym H.S.

Lab Sample ID: 550-79969-11

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.82		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1008-4 Pool Gym

Lab Sample ID: 550-79969-12

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.7		0.50	ppb	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

Detection Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Client Sample ID: PHS 1010-1 Lobby CAB

Lab Sample ID: 550-79969-13

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.4		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1010-2 Band Room drinking fountain

Lab Sample ID: 550-79969-14

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.8		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS - 1011-1 Gym Lobby

Lab Sample ID: 550-79969-15

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.0		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1015-1

Lab Sample ID: 550-79969-16

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.2		0.50	ppb	1		200.8	Total/NA

Client Sample ID: DO-1

Lab Sample ID: 550-79969-17

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.0		0.50	ppb	1		200.8	Total/NA

Client Sample ID: DO-2

Lab Sample ID: 550-79969-18

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.81		0.50	ppb	1		200.8	Total/NA

Client Sample ID: PHS 1018-1

Lab Sample ID: 550-79969-19

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

Client Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Client Sample ID: A Building drinking fountain

Date Collected: 03/23/17 10:37

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-1

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.53		0.50	ppb		03/30/17 19:51	04/08/17 14:10	1

Client Sample ID: PHS 1002-2

Date Collected: 03/23/17 10:38

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-2

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:15	1

Client Sample ID: PHS 1003-1 Kitchen Sink

Date Collected: 03/23/17 10:53

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-3

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.9		0.50	ppb		03/30/17 19:51	04/08/17 14:17	1

Client Sample ID: PHS 1003-2 Westside

Date Collected: 03/23/17 10:49

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-4

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.7		0.50	ppb		03/30/17 19:51	04/08/17 14:19	1

Client Sample ID: Main Building East PHS 1004-1

Date Collected: 03/23/17 10:08

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-5

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:21	1

Client Sample ID: North Computer Lab - PHS 1004-2

Date Collected: 03/23/17 10:11

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-6

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.76		0.50	ppb		03/30/17 19:51	04/08/17 14:23	1

Client Sample ID: West Girls RR bottle #3 PHS 1004-3

Date Collected: 03/23/17 10:15

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-7

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:25	1

Client Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Client Sample ID: PHS 1005-1

Date Collected: 03/23/17 10:26

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-8

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:27	1

Client Sample ID: PHS 1006-1 VOC trades

Date Collected: 03/23/17 10:30

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-9

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:29	1

Client Sample ID: PHS - 1007-1

Date Collected: 03/23/17 11:06

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-10

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.84		0.50	ppb		03/30/17 19:51	04/08/17 14:34	1

Client Sample ID: PHS 1007-2 East Gym H.S.

Date Collected: 03/23/17 11:07

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-11

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.82		0.50	ppb		03/30/17 19:51	04/08/17 14:36	1

Client Sample ID: PHS 1008-4 Pool Gym

Date Collected: 03/23/17 11:10

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-12

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.7		0.50	ppb		03/30/17 19:51	04/08/17 14:38	1

Client Sample ID: PHS 1010-1 Lobby CAB

Date Collected: 03/23/17 10:43

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-13

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.4		0.50	ppb		03/30/17 19:51	04/08/17 14:40	1

Client Sample ID: PHS 1010-2 Band Room drinking fountain

Date Collected: 03/23/17 10:43

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-14

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.8		0.50	ppb		03/30/17 19:51	04/08/17 14:42	1

Client Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Client Sample ID: PHS - 1011-1 Gym Lobby

Date Collected: 03/23/17 11:08

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-15

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.0		0.50	ppb		03/30/17 19:51	04/08/17 14:44	1

Client Sample ID: PHS 1015-1

Date Collected: 03/23/17 10:26

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-16

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.2		0.50	ppb		03/30/17 19:51	04/08/17 14:46	1

Client Sample ID: DO-1

Date Collected: 03/23/17 08:05

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-17

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.0		0.50	ppb		03/30/17 19:51	04/08/17 14:48	1

Client Sample ID: DO-2

Date Collected: 03/23/17 08:10

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-18

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.81		0.50	ppb		03/30/17 19:51	04/08/17 14:50	1

Client Sample ID: PHS 1018-1

Date Collected: 03/23/17 10:31

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-19

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:52	1

QC Sample Results

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 550-113729/1-A
Matrix: Water
Analysis Batch: 114406

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 113729

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:51	04/08/17 14:00	1

Lab Sample ID: LCS 550-113729/4-A
Matrix: Water
Analysis Batch: 114406

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 113729

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Lead	100	98.4		ppb		98	85 - 115

Lab Sample ID: LCSD 550-113729/5-A
Matrix: Water
Analysis Batch: 114406

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 113729

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	Limits	RPD	Limit
Lead	100	94.4		ppb		94	85 - 115	4	20

Lab Sample ID: 550-79969-1 MS
Matrix: Drinking Water
Analysis Batch: 114406

Client Sample ID: A Building drinking fountain
Prep Type: Total/NA
Prep Batch: 113729

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Lead	0.53		100	90.5		ppb		90	70 - 130

Lab Sample ID: 550-79969-1 MSD
Matrix: Drinking Water
Analysis Batch: 114406

Client Sample ID: A Building drinking fountain
Prep Type: Total/NA
Prep Batch: 113729

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	Limits	RPD	Limit
Lead	0.53		100	89.3		ppb		89	70 - 130	1	20

QC Association Summary

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Metals

Prep Batch: 113729

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-79969-1	A Building drinking fountain	Total/NA	Drinking Water	200.8	
550-79969-2	PHS 1002-2	Total/NA	Drinking Water	200.8	
550-79969-3	PHS 1003-1 Kitchen Sink	Total/NA	Drinking Water	200.8	
550-79969-4	PHS 1003-2 Westside	Total/NA	Drinking Water	200.8	
550-79969-5	Main Building East PHS 1004-1	Total/NA	Drinking Water	200.8	
550-79969-6	North Computer Lab - PHS 1004-2	Total/NA	Drinking Water	200.8	
550-79969-7	West Girls RR bottle #3 PHS 1004-3	Total/NA	Drinking Water	200.8	
550-79969-8	PHS 1005-1	Total/NA	Drinking Water	200.8	
550-79969-9	PHS 1006-1 VOC trades	Total/NA	Drinking Water	200.8	
550-79969-10	PHS - 1007-1	Total/NA	Drinking Water	200.8	
550-79969-11	PHS 1007-2 East Gym H.S.	Total/NA	Drinking Water	200.8	
550-79969-12	PHS 1008-4 Pool Gym	Total/NA	Drinking Water	200.8	
550-79969-13	PHS 1010-1 Lobby CAB	Total/NA	Drinking Water	200.8	
550-79969-14	PHS 1010-2 Band Room drinking fountain	Total/NA	Drinking Water	200.8	
550-79969-15	PHS - 1011-1 Gym Lobby	Total/NA	Drinking Water	200.8	
550-79969-16	PHS 1015-1	Total/NA	Drinking Water	200.8	
550-79969-17	DO-1	Total/NA	Drinking Water	200.8	
550-79969-18	DO-2	Total/NA	Drinking Water	200.8	
550-79969-19	PHS 1018-1	Total/NA	Drinking Water	200.8	
MB 550-113729/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-113729/4-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-113729/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-79969-1 MS	A Building drinking fountain	Total/NA	Drinking Water	200.8	
550-79969-1 MSD	A Building drinking fountain	Total/NA	Drinking Water	200.8	

Analysis Batch: 114406

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-79969-1	A Building drinking fountain	Total/NA	Drinking Water	200.8	113729
550-79969-2	PHS 1002-2	Total/NA	Drinking Water	200.8	113729
550-79969-3	PHS 1003-1 Kitchen Sink	Total/NA	Drinking Water	200.8	113729
550-79969-4	PHS 1003-2 Westside	Total/NA	Drinking Water	200.8	113729
550-79969-5	Main Building East PHS 1004-1	Total/NA	Drinking Water	200.8	113729
550-79969-6	North Computer Lab - PHS 1004-2	Total/NA	Drinking Water	200.8	113729
550-79969-7	West Girls RR bottle #3 PHS 1004-3	Total/NA	Drinking Water	200.8	113729
550-79969-8	PHS 1005-1	Total/NA	Drinking Water	200.8	113729
550-79969-9	PHS 1006-1 VOC trades	Total/NA	Drinking Water	200.8	113729
550-79969-10	PHS - 1007-1	Total/NA	Drinking Water	200.8	113729
550-79969-11	PHS 1007-2 East Gym H.S.	Total/NA	Drinking Water	200.8	113729
550-79969-12	PHS 1008-4 Pool Gym	Total/NA	Drinking Water	200.8	113729
550-79969-13	PHS 1010-1 Lobby CAB	Total/NA	Drinking Water	200.8	113729
550-79969-14	PHS 1010-2 Band Room drinking fountain	Total/NA	Drinking Water	200.8	113729
550-79969-15	PHS - 1011-1 Gym Lobby	Total/NA	Drinking Water	200.8	113729
550-79969-16	PHS 1015-1	Total/NA	Drinking Water	200.8	113729
550-79969-17	DO-1	Total/NA	Drinking Water	200.8	113729
550-79969-18	DO-2	Total/NA	Drinking Water	200.8	113729
550-79969-19	PHS 1018-1	Total/NA	Drinking Water	200.8	113729
MB 550-113729/1-A	Method Blank	Total/NA	Water	200.8	113729
LCS 550-113729/4-A	Lab Control Sample	Total/NA	Water	200.8	113729
LCSD 550-113729/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	113729
550-79969-1 MS	A Building drinking fountain	Total/NA	Drinking Water	200.8	113729
550-79969-1 MSD	A Building drinking fountain	Total/NA	Drinking Water	200.8	113729

TestAmerica Phoenix

Lab Chronicle

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Client Sample ID: A Building drinking fountain

Date Collected: 03/23/17 10:37

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:10	TEK	TAL PHX

Client Sample ID: PHS 1002-2

Date Collected: 03/23/17 10:38

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:15	TEK	TAL PHX

Client Sample ID: PHS 1003-1 Kitchen Sink

Date Collected: 03/23/17 10:53

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:17	TEK	TAL PHX

Client Sample ID: PHS 1003-2 Westside

Date Collected: 03/23/17 10:49

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:19	TEK	TAL PHX

Client Sample ID: Main Building East PHS 1004-1

Date Collected: 03/23/17 10:08

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:21	TEK	TAL PHX

Client Sample ID: North Computer Lab - PHS 1004-2

Date Collected: 03/23/17 10:11

Date Received: 03/29/17 10:30

Lab Sample ID: 550-79969-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:23	TEK	TAL PHX

TestAmerica Phoenix

Lab Chronicle

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Client Sample ID: West Girls RR bottle #3 PHS 1004-3

Lab Sample ID: 550-79969-7

Date Collected: 03/23/17 10:15

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:25	TEK	TAL PHX

Client Sample ID: PHS 1005-1

Lab Sample ID: 550-79969-8

Date Collected: 03/23/17 10:26

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:27	TEK	TAL PHX

Client Sample ID: PHS 1006-1 VOC trades

Lab Sample ID: 550-79969-9

Date Collected: 03/23/17 10:30

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:29	TEK	TAL PHX

Client Sample ID: PHS - 1007-1

Lab Sample ID: 550-79969-10

Date Collected: 03/23/17 11:06

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:34	TEK	TAL PHX

Client Sample ID: PHS 1007-2 East Gym H.S.

Lab Sample ID: 550-79969-11

Date Collected: 03/23/17 11:07

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:36	TEK	TAL PHX

Client Sample ID: PHS 1008-4 Pool Gym

Lab Sample ID: 550-79969-12

Date Collected: 03/23/17 11:10

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:38	TEK	TAL PHX

Lab Chronicle

Client: Arizona Dept of Environmental Quality
 Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
 SDG: Page High School

Client Sample ID: PHS 1010-1 Lobby CAB

Lab Sample ID: 550-79969-13

Date Collected: 03/23/17 10:43

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:40	TEK	TAL PHX

Client Sample ID: PHS 1010-2 Band Room drinking fountain

Lab Sample ID: 550-79969-14

Date Collected: 03/23/17 10:43

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:42	TEK	TAL PHX

Client Sample ID: PHS - 1011-1 Gym Lobby

Lab Sample ID: 550-79969-15

Date Collected: 03/23/17 11:08

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:44	TEK	TAL PHX

Client Sample ID: PHS 1015-1

Lab Sample ID: 550-79969-16

Date Collected: 03/23/17 10:26

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:46	TEK	TAL PHX

Client Sample ID: DO-1

Lab Sample ID: 550-79969-17

Date Collected: 03/23/17 08:05

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:48	TEK	TAL PHX

Client Sample ID: DO-2

Lab Sample ID: 550-79969-18

Date Collected: 03/23/17 08:10

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:50	TEK	TAL PHX

Lab Chronicle

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Client Sample ID: PHS 1018-1

Lab Sample ID: 550-79969-19

Date Collected: 03/23/17 10:31

Matrix: Drinking Water

Date Received: 03/29/17 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113729	03/30/17 19:51	SLS	TAL PHX
Total/NA	Analysis	200.8		1	114406	04/08/17 14:52	TEK	TAL PHX

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Accreditation/Certification Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School

Laboratory: TestAmerica Phoenix

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	EPA Region	Identification Number	Expiration Date
Arizona	State Program	9	AZ0728	06-09-17

Analysis Method	Prep Method	Matrix	Analyte
-----------------	-------------	--------	---------

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW / Page Unified School District

TestAmerica Job ID: 550-79969-1
SDG: Page High School


Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	TAL PHX

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340



79969

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School District #8
School Name	Page High School
Building (name/number)	A Building - 1002-1
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	A building
Sample Identification Number (ensure this number is also on the sample container)	A Building drinking fountain
Date of Collection	3/23/17
Time of Collection	10:37am
Printed Name of Sample Collector	marcus daw
Signature Sample Collector	<i>[Signature]</i>

Notes Sample collector:

"classroom A" aka "A-Building"

550-79969 Chain of Custody



For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	3-29-17 1030
Signature	<i>[Signature]</i>
Notes:	19.5% NO LEAD FE Gr

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

①

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	A building - 1002-2
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	A Building DF
Sample Identification Number (ensure this number is also on the sample container)	PHS 1002-2
Date of Collection	3/23/17
Time of Collection	10:38 am
Printed Name of Sample Collector	MARIS DAW
Signature Sample Collector	

Notes Sample collector:

"Classroom A" aka "A-Building"

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

2

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Cafeteria-1003-1
Type of Fixture (tap, drinking fountain etc.)	Kitchen sink
Location of Fixture (example, room number)	Cafeteria
Sample Identification Number (ensure this number is also on the sample container)	PHS 1003-1 Kitchen sink
Date of Collection	3/23/17
Time of Collection	10:53 AM
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	M-D-

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

3

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist # 8
School Name	Page High School
Building (name/number)	Cafeteria 1003-2
Type of Fixture (tap, drinking fountain etc.)	Westside - Drinking fountain
Location of Fixture (example, room number)	Cafeteria
Sample Identification Number (ensure this number is also on the sample container)	PHS 1003-2 Westside
Date of Collection	3/23/17
Time of Collection	10:49 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.


4

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	(PHS) main 1004-1
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	PHS main
Sample Identification Number (ensure this number is also on the sample container)	main Building East - PHS 1004-1
Date of Collection	3/23/17
Time of Collection	10:08 am
Printed Name of Sample Collector	marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.



Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Main Bldg - 1004-2
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	PHS main
Sample Identification Number (ensure this number is also on the sample container)	North Computer Lab - PHS 1004-2
Date of Collection	3/23/17
Time of Collection	10:11am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	M-D-

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.




Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist. #8
School Name	Page High School
Building (name/number)	Main Bldg - 1004-3
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	PHS main
Sample Identification Number (ensure this number is also on the sample container)	West Girls RR bottle #3 PHS-1004-3
Date of Collection	3/23/17
Time of Collection	10:15am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

(7)


1
2
3
4
5
6
7
8
9
10
11
12
13
14

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Auto Body - 1005-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	PHS 1005-1
Date of Collection	3/23/17
Time of Collection	10:26 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

8




Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	VOC Trades 1006-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	PHS 1006-1 VOC trades
Date of Collection	3/23/17
Time of Collection	10:30 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.


9

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	HS Physical Ed - 1007-1
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	West Gym H.S. Boys lockers
Sample Identification Number (ensure this number is also on the sample container)	PHS-1007-1
Date of Collection	3/23/17
Time of Collection	11:06 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.


10

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Dodge High School
Building (name/number)	HS Physical Ed 1007-2
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	East Gym H.S.
Sample Identification Number (ensure this number is also on the sample container)	PHS 10072 EAST GYM H.S.
Date of Collection	3/23/17
Time of Collection	11:07 am
Printed Name of Sample Collector	MARCUS DAW
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

11

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Gym 2 1008-1
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	Pool Gym
Sample Identification Number (ensure this number is also on the sample container)	PHS 1008-4 POOL GYM
Date of Collection	3/23/17
Time of Collection	11:10 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

12

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Auditorium - 1010-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Lobby CAB
Sample Identification Number (ensure this number is also on the sample container)	PHS 1010-1 Lobby CAB
Date of Collection	3/23/17
Time of Collection	10:43 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

CAB = Cultural Arts Building

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

13

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Auditorium 1010-2
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	Band Room
Sample Identification Number (ensure this number is also on the sample container)	PHS 1010-2 Band Room drinking fountain
Date of Collection	3/23/17
Time of Collection	10:43 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

14

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Weight Room -1011-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Gym Lobby
Sample Identification Number (ensure this number is also on the sample container)	PHS 1011-1 Gym Lobby
Date of Collection	3/23/17
Time of Collection	11:08 am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	M-D-

Notes Sample collector:
 There were no water fountains in weight room

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.




Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Visitor Fieldhouse - 1015-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	East inside RR Visitor field house
Sample Identification Number (ensure this number is also on the sample container)	PHS 1015-1
Date of Collection	3/22/17
Time of Collection	10:26am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

"Metal" ^{aka} "Visitor Fieldhouse"

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.



Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist # 8
School Name	Page High School
Building (name/number)	1017 - District Office
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Business office
Sample Identification Number (ensure this number is also on the sample container)	DO-1
Date of Collection	3/23/17
Time of Collection	8:05 AM
Printed Name of Sample Collector	Kris Swatski
Signature Sample Collector	<i>Kris Swatski</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

(17)

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	1017- District office
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Board Room Side
Sample Identification Number (ensure this number is also on the sample container)	DO-25
Date of Collection	3/23/17
Time of Collection	8:10 AM
Printed Name of Sample Collector	Kris Swatski
Signature Sample Collector	<i>Kris Swatski</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.


18

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	New Fieldhouse 1018-1
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	New Fieldhouse
Sample Identification Number (ensure this number is also on the sample container)	PHS 1018-1
Date of Collection	3/23/17
Time of Collection	10:31am
Printed Name of Sample Collector	Marcus Daw
Signature Sample Collector	

Notes Sample collector:

Dist. Maint = New Fieldhouse

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

(19)

Bottle Order Information

Bottle Order: Lead in Drinking Water
 Bottle Order #: 9509
 Request From Client: 2/14/2017
 Date Order Posted: 2/14/2017 8:55:48AM
 Order Status: Ready To Process
 Prepared By: Amanda Barkley
 Deliver By Date: 2/21/2017 11:59:00PM
 Lab Project Number: 55009093

Order Completion Information

Creator: Amanda Barkley
 Filled by: *A*
 Sent Date: 02-21-17
 Sent Via:
 Tracking #:

Sets	Bottles/Set	Qty	Bottle Type Description	Preservative	Method	Matrix	Sample Type	Comments	Lot #
34	1	34	Plastic 250ml - unpreserved	None	200.8 - Lead	Water	Normal		034736018

Notes to Field Staff:

Health and Safety Notes: Preservative Comment

Relinquished By	Company	Date	Time	Received By	Company	Seal #:
Relinquished By	Company	Date	Time	Received By	Company	Seal #: Seal #: Seal #: Seal #:

Please notify us immediately if an error is found in shipment


1
2
3
4
5
6
7
8
9
10
11
12
13
14

Page Unified School District Warehouse

ATTN: Kris Swatski

817 Aqua Street

Page, AZ 86040



Page Unified District	Page High School	1001	Classroom D	HS Classroom	1961	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1002	Classroom A	HS Classroom	1961	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1003	Cafeteria	Cafeteria	1961	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1004	Main Bldg.	Classrooms	1971	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1005	Auto Body	Classroom	1976	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1006	Voc. Trades	Classrooms	1988	P O Box 1927	Page	86040-1927	Coconino	1
Page Unified District	Page High School	1007	HS Physical Ed.	Classroom	1990	P O Box 1927	Page	86040-1927	Coconino	1
Page Unified District	Page High School	1008	(PAUL GYM) Gym 2	Classroom	1979	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1009	Pool NA	Classroom	1979	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1010	Auditorium	Auditorium/Classroom	1976	P O Box 1927	Page	86040-1927	Coconino	2

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist # 8
School Name	Page High School
Building (name/number)	1001 classroom D
Type of Fixture (tap, drinking fountain etc.)	N/A
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector:
 sample sent w/ Desertview Elem.

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	1009 POOL
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	N/A
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector:
 NO water in pool
 has been closed for years

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.



Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist # 8
School Name	Page High School
Building (name/number)	1012 Greenhouse
Type of Fixture (tap, drinking fountain etc.)	N/A
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector:


greenhouse closed

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.



Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	1013- Portable 1
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector:	N/A there is no portable
-------------------------	-----------------------------

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	1014 - Portable II
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector: *N/A*
there is no portable #

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	1016 - Storage
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector: *N/A
no drinking water @
Storage*

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	Page Unified School Dist #8
School Name	Page High School
Building (name/number)	Boiler room e gym Pool 1019
Type of Fixture (tap, drinking fountain etc.)	N/A
Location of Fixture (example, room number)	
Sample Identification Number (ensure this number is also on the sample container)	
Date of Collection	
Time of Collection	
Printed Name of Sample Collector	
Signature Sample Collector	

Notes Sample collector:

N/A


For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.





Page Unified District	Page High School	1011	Weight Room	1	Classroom	1983	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1012	Greenhouse	NP	Greenhouse	1993	P O Box 1927	Page	86040-1927	Coconino	1
Page Unified District	Page High School	1013	Portable I	NP	Classroom	1984	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1014	Portable II	NP	Classroom	1984	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1015	Visitor Fieldhouse	1	Shower/Locker	1983	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1016	Storage	NP	Storage	1983	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1017	Supt. Board	2	Dist. Admin.	1976	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1018	Dist. Maint.	1	Dist. Maint.	1982	P O Box 1927	Page	86040-1927	Coconino	2
Page Unified District	Page High School	1019	Boiler room at Gym pool	NP	heating pool	2006	P O Box 1927	Page	86040-1927	Coconino	1
Total Containers											34

NP total 19

Login Sample Receipt Checklist

Client: Arizona Dept of Environmental Quality

Job Number: 550-79969-1
SDG Number: Page High School

Login Number: 79969
List Number: 1
Creator: Vilaboy, Monica

List Source: TestAmerica Phoenix

Question	Answer	Comment
Radioactivity wasn't checked or is \leq background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <math><6\text{mm}</math> (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	Check done at department level as required.

