

# TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

## ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Phoenix

4625 East Cotton Ctr Blvd

Suite 189

Phoenix, AZ 85040

Tel: (602)437-3340

TestAmerica Job ID: 550-79660-1

TestAmerica SDG: Julia Randall Elementary School

Client Project/Site: Pb in DW/Payson Unified School District

For:

Arizona Dept of Environmental Quality

1110 W. Washington St., MC5415B-1

Phoenix, Arizona 85007

Attn: David A Burchard



Authorized for release by:

4/6/2017 3:30:36 PM

Jannel Franklin, Project Manager I

(732)593-2551

[jannel.franklin@testamericainc.com](mailto:jannel.franklin@testamericainc.com)

Designee for

Susan Armijo, Project Manager II

(602)659-7664

[susan.armijo@testamericainc.com](mailto:susan.armijo@testamericainc.com)

### LINKS

Review your project  
results through

TotalAccess

Have a Question?



Visit us at:

[www.testamericainc.com](http://www.testamericainc.com)

*This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.*

*Results relate only to the items tested and the sample(s) as received by the laboratory.*

# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Sample Summary . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
QC Sample Results . . . . .	8
QC Association Summary . . . . .	9
Lab Chronicle . . . . .	10
Certification Summary . . . . .	11
Method Summary . . . . .	12
Chain of Custody . . . . .	13
Receipt Checklists . . . . .	18



## Definitions/Glossary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

### Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

## Case Narrative

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

**Job ID: 550-79660-1**

**Laboratory: TestAmerica Phoenix**

### Narrative

#### Job Narrative 550-79660-1

### Comments

No additional comments.

### Receipt

The sample was received on 3/22/2017 10:00 AM; the sample arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 22.2° C.

### Receipt Exceptions

The following samples for metals were received unpreserved and were preserved upon receipt to the laboratory: A-01 (550-79660-1). Regulatory documents require a 24-hour waiting period from the time of the addition of the acid preservative to the time of digestion.

### Metals

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

## Sample Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
550-79660-1	A-01	Drinking Water	03/20/17 06:35	03/22/17 10:00

## Detection Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

**Client Sample ID: A-01**

**Lab Sample ID: 550-79660-1**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.7		0.50	ppb	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

## Client Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

**Client Sample ID: A-01**

**Date Collected: 03/20/17 06:35**

**Date Received: 03/22/17 10:00**

**Lab Sample ID: 550-79660-1**

**Matrix: Drinking Water**

**Method: 200.8 - Metals (ICP/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.7		0.50	ppb		03/30/17 19:15	04/04/17 19:01	1

# QC Sample Results

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

## Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 550-113721/1-A  
Matrix: Water  
Analysis Batch: 114062

Client Sample ID: Method Blank  
Prep Type: Total/NA  
Prep Batch: 113721

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/30/17 19:15	04/04/17 18:20	1

Lab Sample ID: LCS 550-113721/4-A  
Matrix: Water  
Analysis Batch: 114062

Client Sample ID: Lab Control Sample  
Prep Type: Total/NA  
Prep Batch: 113721

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	96.7		ppb		97	85 - 115

Lab Sample ID: LCSD 550-113721/5-A  
Matrix: Water  
Analysis Batch: 114062

Client Sample ID: Lab Control Sample Dup  
Prep Type: Total/NA  
Prep Batch: 113721

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	100	96.3		ppb		96	85 - 115	0	20

Lab Sample ID: 550-79655-A-7-B MS  
Matrix: Water  
Analysis Batch: 114062

Client Sample ID: Matrix Spike  
Prep Type: Total/NA  
Prep Batch: 113721

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	ND		100	84.5		ppb		85	70 - 130

Lab Sample ID: 550-79655-A-7-C MSD  
Matrix: Water  
Analysis Batch: 114062

Client Sample ID: Matrix Spike Duplicate  
Prep Type: Total/NA  
Prep Batch: 113721

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Lead	ND		100	81.4		ppb		81	70 - 130	4	20



## QC Association Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

### Metals

#### Prep Batch: 113721

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-79660-1	A-01	Total/NA	Drinking Water	200.8	
MB 550-113721/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-113721/4-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-113721/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-79655-A-7-B MS	Matrix Spike	Total/NA	Water	200.8	
550-79655-A-7-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	

#### Analysis Batch: 114062

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-79660-1	A-01	Total/NA	Drinking Water	200.8	113721
MB 550-113721/1-A	Method Blank	Total/NA	Water	200.8	113721
LCS 550-113721/4-A	Lab Control Sample	Total/NA	Water	200.8	113721
LCSD 550-113721/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	113721
550-79655-A-7-B MS	Matrix Spike	Total/NA	Water	200.8	113721
550-79655-A-7-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	113721

## Lab Chronicle

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

**Client Sample ID: A-01**

**Date Collected: 03/20/17 06:35**

**Date Received: 03/22/17 10:00**

**Lab Sample ID: 550-79660-1**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			113721	03/30/17 19:15	EXZ	TAL PHX
Total/NA	Analysis	200.8		1	114062	04/04/17 19:01	TEK	TAL PHX

### Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

## Accreditation/Certification Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

### Laboratory: TestAmerica Phoenix

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	EPA Region	Identification Number	Expiration Date
Arizona	State Program	9	AZ0728	06-09-17

Analysis Method	Prep Method	Matrix	Analyte
-----------------	-------------	--------	---------

## Method Summary

Client: Arizona Dept of Environmental Quality  
Project/Site: Pb in DW/Payson Unified School District

TestAmerica Job ID: 550-79660-1  
SDG: Julia Randall Elementary School

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	TAL PHX

### Protocol References:

EPA = US Environmental Protection Agency

### Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

79660

### Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	PAYSON UNIFIED SCHOOL DISTRICT #10
School Name	JULIA RANDALL ELEMENTARY
Building (name/number)	A-BUILDING (GYM)
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	GYM
Sample Identification Number ( <u>Write this number on the sample container and on this sheet</u> )	A-01
Date of Collection	3/20/17
Time of Collection	4:35 AM
Printed Name of Sample Collector	ROBERT WILCOX
Signature Sample Collector	<i>Robert Wilcox</i>

550-79660 Chain of Custody



Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	3-22-17 1000
Signature	<i>[Signature]</i>
Notes: 22.2°C NO TCB FB Cr	

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

①

These samples were collected for lead screening purposes only and cannot be used for compliance.

To whom it may concern,

We have only provided one sample for Julia Randall Elementary as this was all that was called for out of the new building, all other samples requested no longer exist on the premises. Let us know if there is any other necessary testing that needs to be done at that particular site.

Thank you,

A handwritten signature in black ink, appearing to read 'Robbie Wilcox', written in a cursive style.

Robbie Wilcox  
Maintenance Coordinator  
Payson Unified School District #10

# Bottle Order Information

Bottle Order: Lead in Drinking Water  
 Bottle Order #: 9554  
 Request From Client: 2/16/2017  
 Date Order Posted: 2/17/2017 9:33:13AM  
 Order Status: Ready To Process  
 Prepared By: Amanda Barkley  
 Deliver By Date: 2/24/2017 11:59:00PM  
 Lab Project Number: 55009093

# Order Completion Information

Creator: Amanda Barkley  
 Filled by:  
 Sent Date:  
 Sent Via:  
 Tracking #:

*[Handwritten signature]*  
02-22-17

Sets	Bottles/Set	Qty	Bottle Type Description	Preservative	Method	Matrix	Sample Type	Comments	Lot #
17	1	17	Plastic 250ml - unpreserved	None	200.8 - Lead	Water	Normal	03/39018	

# Notes to Field Staff:

Health and Safety Notes:

Preservative

Comment

Relinquished By	Company	Date	Time	Received By	Company	Seal #
Relinquished By	Company	Date	Time	Received By	Company	Seal #
Relinquished By	Company	Date	Time	Received By	Company	Seal #
Relinquished By	Company	Date	Time	Received By	Company	Seal #

Please notify us immediately if an error is found in shipment

Payson Unified District	Randall Elementary School	1001	Portable I	Classroom	1984	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1002	Portable II	Classroom	1984	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1003	100 Building	Classroom	1971	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1004	200 Building	Classroom	1974	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1005	Library	Library	1983	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1006	Computer lab	Classroom	1967	P O Box 919	Payson	85547-0919	Gila	2
Payson Unified District	Julia Randall Elementary School	1007	Storage	Storage	1981	P O Box 919	Payson	85547-0919	Gila	2

\*NOTE - DNE/DOES NOT EXIST



	School								
Payson Unified District	Julia Randall Elementary School	1008	Cafeteria, Office, Classrooms	Classrooms	1985	P O Box 919	Payson	85547-0919	Gila
Payson Unified District	Julia Randall Elementary School	1010	<sup>DINKIN FOUNDATION</sup> Replacement school	Classrooms, etc.	2009	P O Box 919	Payson	85547-0919	Gila
<b>Total Containers</b>									

2 DNE

17

Payson Unified District

ATTN: Leslie Herdeman/Maintenance Department

409 W. Longhorn Road

Payson, AZ 85541

## Login Sample Receipt Checklist

Client: Arizona Dept of Environmental Quality

Job Number: 550-79660-1

SDG Number: Julia Randall Elementary School

**Login Number: 79660**

**List Number: 1**

**Creator: Vilaboy, Monica**

**List Source: TestAmerica Phoenix**

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	Check done at department level as required.