

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Phoenix

4625 East Cotton Ctr Blvd

Suite 189

Phoenix, AZ 85040

Tel: (602)437-3340

TestAmerica Job ID: 550-78971-1

TestAmerica Sample Delivery Group: Kingman High School

Client Project/Site: Pb in DW/Kingman Unified District

For:

Arizona Dept of Environmental Quality

1110 W. Washington St., MC5415B-1

Phoenix, Arizona 85007

Attn: David A Burchard



Authorized for release by:

3/23/2017 11:06:44 AM

Susan Armijo, Project Manager II

(602)659-7664

susan.armijo@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Sample Summary	5
Detection Summary	6
Client Sample Results	7
QC Sample Results	9
QC Association Summary	10
Lab Chronicle	11
Certification Summary	13
Method Summary	14
Chain of Custody	15
Receipt Checklists	23



Definitions/Glossary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Job ID: 550-78971-1

Laboratory: TestAmerica Phoenix

Narrative

Job Narrative 550-78971-1

Comments

No additional comments.

Receipt

The samples were received on 3/9/2017 10:00 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 20.0° C.

Receipt Exceptions

The following samples for metals were received unpreserved and were preserved upon receipt to the laboratory: Drinking A 150 (550-78971-1), Drinking B129 (550-78971-2), Drinking C113 (550-78971-3), Drinking D104 (550-78971-4), Drinking G110 (550-78971-5), F108 JC (550-78971-6), SINK E103 (550-78971-7) and Drinking H117 (550-78971-8). Regulatory documents require a 24-hour waiting period from the time of the addition of the acid preservative to the time of digestion.

Metals

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

Sample Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
550-78971-1	Drinking A 150	Drinking Water	03/07/17 01:00	03/09/17 10:00
550-78971-2	Drinking B129	Drinking Water	03/07/17 01:10	03/09/17 10:00
550-78971-3	Drinking C113	Drinking Water	03/07/17 01:20	03/09/17 10:00
550-78971-4	Drinking D104	Drinking Water	03/07/17 01:30	03/09/17 10:00
550-78971-5	Drinking G110	Drinking Water	03/07/17 01:40	03/09/17 10:00
550-78971-6	F108 JC	Drinking Water	03/07/17 01:50	03/09/17 10:00
550-78971-7	SINK E103	Drinking Water	03/07/17 02:00	03/09/17 10:00
550-78971-8	Drinking H117	Drinking Water	03/07/17 02:10	03/09/17 10:00

Detection Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Client Sample ID: Drinking A 150

Lab Sample ID: 550-78971-1

No Detections.

Client Sample ID: Drinking B129

Lab Sample ID: 550-78971-2

No Detections.

Client Sample ID: Drinking C113

Lab Sample ID: 550-78971-3

No Detections.

Client Sample ID: Drinking D104

Lab Sample ID: 550-78971-4

No Detections.

Client Sample ID: Drinking G110

Lab Sample ID: 550-78971-5

No Detections.

Client Sample ID: F108 JC

Lab Sample ID: 550-78971-6

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.95		0.50	ppb	1		200.8	Total/NA

Client Sample ID: SINK E103

Lab Sample ID: 550-78971-7

No Detections.

Client Sample ID: Drinking H117

Lab Sample ID: 550-78971-8

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Phoenix

Client Sample Results

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Client Sample ID: Drinking A 150

Date Collected: 03/07/17 01:00

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-1

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:07	1

Client Sample ID: Drinking B129

Date Collected: 03/07/17 01:10

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-2

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:09	1

Client Sample ID: Drinking C113

Date Collected: 03/07/17 01:20

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-3

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:11	1

Client Sample ID: Drinking D104

Date Collected: 03/07/17 01:30

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-4

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:13	1

Client Sample ID: Drinking G110

Date Collected: 03/07/17 01:40

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-5

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:15	1

Client Sample ID: F108 JC

Date Collected: 03/07/17 01:50

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-6

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.95		0.50	ppb		03/21/17 13:02	03/21/17 15:17	1

Client Sample ID: SINK E103

Date Collected: 03/07/17 02:00

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-7

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:19	1

TestAmerica Phoenix

Client Sample Results

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Client Sample ID: Drinking H117

Date Collected: 03/07/17 02:10

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-8

Matrix: Drinking Water

Method: 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 15:21	1

QC Sample Results

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 550-112867/1-A
Matrix: Water
Analysis Batch: 112929

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 112867

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	ND		0.50	ppb		03/21/17 13:02	03/21/17 14:52	1

Lab Sample ID: LCS 550-112867/4-A
Matrix: Water
Analysis Batch: 112929

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 112867

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	100	91.2		ppb		91	85 - 115

Lab Sample ID: LCSD 550-112867/5-A
Matrix: Water
Analysis Batch: 112929

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 112867

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	100	92.6		ppb		93	85 - 115	2	20

Lab Sample ID: 550-78970-A-8-B MS
Matrix: Water
Analysis Batch: 112929

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 112867

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Lead	ND		100	90.0		ppb		90	70 - 130

Lab Sample ID: 550-78970-A-8-C MSD
Matrix: Water
Analysis Batch: 112929

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 112867

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Lead	ND		100	91.3		ppb		91	70 - 130	1	20

QC Association Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Metals

Prep Batch: 112867

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78971-1	Drinking A 150	Total/NA	Drinking Water	200.8	
550-78971-2	Drinking B129	Total/NA	Drinking Water	200.8	
550-78971-3	Drinking C113	Total/NA	Drinking Water	200.8	
550-78971-4	Drinking D104	Total/NA	Drinking Water	200.8	
550-78971-5	Drinking G110	Total/NA	Drinking Water	200.8	
550-78971-6	F108 JC	Total/NA	Drinking Water	200.8	
550-78971-7	SINK E103	Total/NA	Drinking Water	200.8	
550-78971-8	Drinking H117	Total/NA	Drinking Water	200.8	
MB 550-112867/1-A	Method Blank	Total/NA	Water	200.8	
LCS 550-112867/4-A	Lab Control Sample	Total/NA	Water	200.8	
LCSD 550-112867/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	
550-78970-A-8-B MS	Matrix Spike	Total/NA	Water	200.8	
550-78970-A-8-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	

Analysis Batch: 112929

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
550-78971-1	Drinking A 150	Total/NA	Drinking Water	200.8	112867
550-78971-2	Drinking B129	Total/NA	Drinking Water	200.8	112867
550-78971-3	Drinking C113	Total/NA	Drinking Water	200.8	112867
550-78971-4	Drinking D104	Total/NA	Drinking Water	200.8	112867
550-78971-5	Drinking G110	Total/NA	Drinking Water	200.8	112867
550-78971-6	F108 JC	Total/NA	Drinking Water	200.8	112867
550-78971-7	SINK E103	Total/NA	Drinking Water	200.8	112867
550-78971-8	Drinking H117	Total/NA	Drinking Water	200.8	112867
MB 550-112867/1-A	Method Blank	Total/NA	Water	200.8	112867
LCS 550-112867/4-A	Lab Control Sample	Total/NA	Water	200.8	112867
LCSD 550-112867/5-A	Lab Control Sample Dup	Total/NA	Water	200.8	112867
550-78970-A-8-B MS	Matrix Spike	Total/NA	Water	200.8	112867
550-78970-A-8-C MSD	Matrix Spike Duplicate	Total/NA	Water	200.8	112867

Lab Chronicle

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Client Sample ID: Drinking A 150

Date Collected: 03/07/17 01:00

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:07	TEK	TAL PHX

Client Sample ID: Drinking B129

Date Collected: 03/07/17 01:10

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:09	TEK	TAL PHX

Client Sample ID: Drinking C113

Date Collected: 03/07/17 01:20

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:11	TEK	TAL PHX

Client Sample ID: Drinking D104

Date Collected: 03/07/17 01:30

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:13	TEK	TAL PHX

Client Sample ID: Drinking G110

Date Collected: 03/07/17 01:40

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:15	TEK	TAL PHX

Client Sample ID: F108 JC

Date Collected: 03/07/17 01:50

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:17	TEK	TAL PHX

TestAmerica Phoenix

Lab Chronicle

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Client Sample ID: SINK E103

Date Collected: 03/07/17 02:00

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-7

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:19	TEK	TAL PHX

Client Sample ID: Drinking H117

Date Collected: 03/07/17 02:10

Date Received: 03/09/17 10:00

Lab Sample ID: 550-78971-8

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	200.8			112867	03/21/17 13:02	TEK	TAL PHX
Total/NA	Analysis	200.8		1	112929	03/21/17 15:21	TEK	TAL PHX

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

Certification Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Laboratory: TestAmerica Phoenix

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
Arizona	State Program	9	AZ0728	06-09-17

Analysis Method	Prep Method	Matrix	Analyte
-----------------	-------------	--------	---------

Method Summary

Client: Arizona Dept of Environmental Quality
Project/Site: Pb in DW/Kingman Unified District

TestAmerica Job ID: 550-78971-1
SDG: Kingman High School

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	TAL PHX

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL PHX = TestAmerica Phoenix, 4625 East Cotton Ctr Blvd, Suite 189, Phoenix, AZ 85040, TEL (602)437-3340

78971


ADEQ Public School Drinking Water Lead Screening Program
Sampling Plan & Collection Log

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School	550-78971 Chain of Custody 
School Name	KINGMAN HIGH SCHOOL	
Building (name/number)	A	
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN	
Location of Fixture (example, room number)	100 HALL	
Sample Identification Number (Write this number on the sample container and on this sheet)	DRINKING A 150	
Date of Collection	3/7/17	
Time of Collection	1:00 AM	
Printed Name of Sample Collector	Arnold Mc Coy	
Signature Sample Collector	<i>Arnold Mc Coy</i>	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

Felix  3/9/17 11:00 AM *re 2000*

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	B
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	DRINKING B129
Date of Collection	3/7/17
Time of Collection	1:10 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	C
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	DRINKING C113
Date of Collection	3/7/17
Time of Collection	1:20 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	D
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (Write this number on the sample container and on this sheet)	DRINKING D104
Date of Collection	3/7/17
Time of Collection	1:30 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeg.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	G
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (Write this number on the sample container and on this sheet)	DRINKING G110
Date of Collection	3/7/17
Time of Collection	1:40 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	F
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	F108 JC
Date of Collection	3/7/17
Time of Collection	1:50 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	E
Type of Fixture (tap, drinking fountain etc.)	SINK
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	SINK E103
Date of Collection	3/7/17
Time of Collection	2:00 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Collection Log

For experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Kingman Unified School
School Name	KINGMAN HIGH SCHOOL
Building (name/number)	H
Type of Fixture (tap, drinking fountain etc.)	DRINKING FOUNTAIN
Location of Fixture (example, room number)	100 HALL
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	DRINKING H117
Date of Collection	3/7/17
Time of Collection	2:10 AM
Printed Name of Sample Collector	Arnold Mc Coy
Signature Sample Collector	<i>Arnold Mc Coy</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature

These samples were collected for lead screening purposes only and cannot be used for compliance.

<http://www.azdeq.gov/LeadScreeningProg>

Login Sample Receipt Checklist

Client: Arizona Dept of Environmental Quality

Job Number: 550-78971-1
SDG Number: Kingman High School

Login Number: 78971

List Number: 1

Creator: Gravlin, Andrea

List Source: TestAmerica Phoenix

Question	Answer	Comment
Radioactivity wasn't checked or is \leq background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	Check done at department level as required.