

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: (626) 386-1100  
Fax: (626) 386-1101  
1 800 566 LABS (1 800 566 5227)



AT-1807

## Laboratory Report

for

Arizona Department of Environmental Quality  
1110 West Washington Street  
Phoenix, AZ 85007  
Attention: David Burchard

Date of Issue  
03/30/2017



Eurofins Eaton  
Analytical

TDF: Thomas.D.French  
Project Manager



Report: 647443  
Project: MV-ED  
ADHS License #: AZ0778  
Group: Mohave Valley Junior  
PO#: High School  
PO#: ADEQ16-116686:3

\* Accredited in accordance with TNI 2009 and ISO/IEC 17025:2005.

\* Laboratory certifies that the test results meet all **TNI 2009 and ISO/IEC 17025:2005** requirements unless noted under the individual analysis.

\* Following the cover page are State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.

\* Test results relate only to the sample(s) tested.

\* This report shall not be reproduced except in full, without the written approval of the laboratory.

### STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Mississippi	Certified
Arizona	AZ0778	Montana	Cert 0035
Arkansas	Certified	Nebraska	Certified
California-Monrovia-ELAP	2813	Nevada	CA00006-2016
California-Colton- ELAP	2812	New Hampshire *	2959
California-Folsom- ELAP	2820	New Jersey *	CA 008
California-Fresno- ELAP	2966	New Mexico	Certified
Colorado	Certified	New York *	11320
Connecticut	PH-0107	North Carolina	06701
Delaware	CA 006	North Dakota	R-009
Florida *	E871024	Oregon (Primary AB) *	ORELAP 4034
Georgia	947	Pennsylvania *	68-565
Guam	16-003r	Puerto Rico	Certified
Hawaii	Certified	Rhode Island	LAO00326
Idaho	Certified	South Carolina	87016
Illinois *	200033	South Dakota	Certified
Indiana	C-CA-01	Tennessee	TN02839
Kansas *	E-10268	Texas *	T104704230-15-9
Kentucky	90107	Utah *	CA000062016-10
Louisiana *	LA16003	Vermont	VT0114
Maine	CA0006	Virginia *	460260
Maryland	224	Washington	C838
Commonwealth of Northern Marianas Is.	MP0004	Wyoming	Certified
Massachusetts	M-CA006	EPA Region 5	Certified
Michigan	9906	Los Angeles County Sanitation Districts	10264

\* NELAP/TNI Recognized Accreditation Bodies

ISO 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO 17025 as verified by the ANSI-ASQ National Accreditation Board/ANAB. Refer to Certificate and scope of accreditation (AT 1807) found at: <http://www.eatonanalytical.com>

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
1,4-Dioxane	EPA 522	x		x
2,3,7,8-TCDD	Modified EPA 1613B	x		x
Acrylamide	In House Method (2440)	x		x
Alkalinity	SM 2320B	x	x	x
Ammonia	EPA 350.1		x	x
Ammonia	SM 4500-NH3 H		x	x
Anions and DBPs by IC	EPA 300.0	x	x	x
Anions and DBPs by IC	EPA 300.1	x		x
Asbestos	EPA 100.2	x	x	
Bicarbonate Alkalinity as HCO3	SM 2320B	x	x	x
BOD / CBOD	SM 5210B		x	x
Bromate	In House Method (2447)	x		x
Carbamates	EPA 531.2	x		x
Carbonate as CO3	SM 2330B	x	x	x
Carbonyls	EPA 556	x		x
COD	EPA 410.4 / SM 5220D		x	
Chloramines	SM 4500-CL G	x	x	x
Chlorinated Acids	EPA 515.4	x		x
Chlorinated Acids	EPA 555	x		x
Chlorine Dioxide	SM 4500-CLO2 D	x		x
Chlorine -Total/Free/ Combined Residual	SM 4500-Cl G	x	x	x
Conductivity	EPA 120.1		x	
Conductivity	SM 2510B	x	x	x
Corrosivity (Langelier Index)	SM 2330B	x		x
Cryptosporidium	EPA 1623	x		x
Cyanide, Amenable	SM 4500-CN G	x	x	
Cyanide, Free	SM 4500CN F	x	x	x
Cyanide, Total	EPA 335.4	x	x	x
Cyanogen Chloride (screen)	In House Method (2470)	x		x
Diquat and Paraquat	EPA 549.2	x		x
DBP/HAA	SM 6251B	x		x
Dissolved Oxygen	SM 4500-O G		x	x
DOC	SM 5310C	x		x
E. Coli (MTF/EC+MUG)		x		x
E. Coli	CFR 141.21(f)(6)(i)	x		x
E. Coli	SM 9223		x	
E. Coli (Enumeration)	SM 9221B.1/ SM 9221F	x		x
E. Coli (Enumeration)	SM 9223B	x		x
EDB/DCBP	EPA 504.1	x		
EDB/DCBP and DBP	EPA 551.1	x		x
EDTA and NTA	In House Method (2454)	x		x
Endothall	EPA 548.1	x		x
Endothall	In-house Method (2445)	x		x
Enterococci	SM 9230B	x	x	
Fecal Coliform	SM 9221 E (MTF/EC)	x		
Fecal Coliform	SM 9221C, E (MTF/EC)		x	
Fecal Coliform (Enumeration)	SM 9221E (MTF/EC)	x		x
Fecal Coliform with Chlorine Present	SM 9221E		x	
Fecal Streptococci	SM 9230B	x	x	
Fluoride	SM 4500-F C	x	x	x
Giardia	EPA 1623	x		x
Glyphosate	EPA 547	x		x
Gross Alpha/Beta	EPA 900.0	x	x	x
Gross Alpha Coprecipitation	SM 7110 C	x	x	x
Hardness	SM 2340B	x	x	x
Heterotrophic Bacteria	In House Method (2439)	x		x
Heterotrophic Bacteria	SM 9215 B	x		x
Hexavalent Chromium	EPA 218.6	x	x	x

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
Hexavalent Chromium	EPA 218.7	x		x
Hexavalent Chromium	SM 3500-Cr B		x	
Hormones	EPA 539	x		x
Hydroxide as OH Calc.	SM 2330B	x		x
Kjeldahl Nitrogen	EPA 351.2		x	
Legionella	CDC Legionella	x		x
Mercury	EPA 245.1	x	x	x
Metals	EPA 200.7 / 200.8	x	x	x
Microcystin LR	ELISA (2360)	x		x
NDMA	EPA 521	x		x
NDMA	TQ In house method based on EPA 521 (2425)	x		x
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x	x
OCL, Pesticides/PCB	EPA 505	x		x
Ortho Phosphate	EPA 365.1	x	x	x
Ortho Phosphate	SM 4500P E			x
Ortho Phosphorous	SM 4500P E	x		
Oxyhalides Disinfection Byproducts	EPA 317.0	x		x
Perchlorate	EPA 331.0	x		x
Perchlorate (low and high)	EPA 314.0	x		x
Perfluorinated Alkyl Acids	EPA 537	x		x
pH	EPA 150.1	x		
pH	SM 4500-H+B	x	x	x
Phenylurea Pesticides/ Herbicides	In House Method, based on EPA 532 (2448)	x		x
Pseudomonas	IDEXX Pseudalert (2461)	x		x
Radium-226	GA Institute of Tech	x		x
Radium-228	GA Institute of Tech	x		x
Radon-222	SM 7500RN	x		x
Residue, Filterable	SM 2540C	x	x	x
Residue, Non-filterable	SM 2540D		x	
Residue, Total	SM 2540B		x	x
Residue, Volatile	EPA 160.4		x	
Semi-VOC	EPA 525.2	x		x
Semi-VOC	EPA 625		x	x
Silica	SM 4500-Si D	x	x	
Silica	SM 4500-SiO2 C	x	x	
Sulfide	SM 4500-S <sup>-</sup> D		x	
Sulfite	SM 4500-SO <sup>3</sup> B	x	x	x
Surfactants	SM 5540C	x	x	x
Taste and Odor Analytes	SM 6040E	x		x
Total Coliform (P/A)	SM 9221 A, B	x		x
Total Coliform (Enumeration)	SM 9221 A, B, C	x		x
Total Coliform / E. coli	Colisure SM 9223	x		x
Total Coliform	SM 9221B		x	
Total Coliform with Chlorine Present	SM 9221B		x	
Total Coliform / E.coli (P/A and Enumeration)	SM 9223	x		x
TOC	SM 5310C	x	x	x
TOX	SM 5320B		x	
Total Phenols	EPA 420.1		x	
Total Phenols	EPA 420.4	x	x	x
Total Phosphorous	SM 4500 P E		x	
Turbidity	EPA 180.1	x	x	x
Turbidity	SM 2130B	x	x	
Uranium by ICP/MS	EPA 200.8	x		x
UV 254	SM 5910B	x		
VOC	EPA 524.2/EPA 524.3	x		x
VOC	EPA 624		x	x
VOC	EPA SW 846 8260	x		x
VOC	In House Method (2411)	x		x
Yeast and Mold	SM 9610	x		x



#1 Sample

647443

### Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Mohave Valley El. school Dist #16
School Name	Mohave Valley Jr. High
Building (name/number)	Building A, Admin/ Lib/ Classroom 1003
Type of Fixture (tap, drinking fountain etc.)	Kitchen Sink
Location of Fixture (example, room number)	Kitchen
Sample Identification Number ( <b>Write this number on the sample container and on this sheet</b> )	Prod # 00066315 Container # 001583
Date of Collection	3-20-17
Time of Collection	12:00
Printed Name of Sample Collector	David Beard
Signature Sample Collector	David Beard

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	3-27-17 12:30
Signature	MKV
Notes:	20.0 °C

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

# 2 Sample

### Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Mohave Valley El. School District 16
School Name	Mohave Valley Jr. High
Building (name/number)	Building B, Gym, Cafe, chrm., 1002
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number ( <b>Write this number on the sample container and on this sheet</b> )	Prod. # 000 66315
Date of Collection	3-20-17
Time of Collection	12:00
Printed Name of Sample Collector	David Berang
Signature Sample Collector	<i>David Berang</i>

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	3-27-17 12:30
Signature	<i>MRU</i>
Notes:	20.0°C

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

#3 Sample

### Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	Mohave Valley El. Sch. Dist. #16
School Name	MVJH
Building (name/number)	Building A, admin, Lib, cbrm 1001
Type of Fixture (tap, drinking fountain etc.)	Drinking Fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number ( <b>Write this number on the sample container and on this sheet</b> )	Prod # 0006615 Container # 00 1565
Date of Collection	3-20-17
Time of Collection	12:00
Printed Name of Sample Collector	David Bernard
Signature Sample Collector	David Bernard

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	MAV 3-27-17 12:30
Signature	MAV
Notes:	20.0 °C

For relinquishing samples upon delivery to labs only

Relinquished date and signature \_\_\_\_\_

*These samples were collected for lead screening purposes only and cannot be used for compliance.*

# INTERNAL CHAIN OF CUSTODY RECORD

**EUROFINS** | Eaton Analytical

BEA Folder Number:  

SAMPLES REC'D DAY OF COLLECTION?  

SAMPLE TEMP. RECEIVED: (Observation = 20.2 °C) (Corr. Factor = -2 °C) (Final = 20.0 °C) Thawed N/A

IR Gun ID = 569A (Observation = 20.2 °C) (Corr. Factor = -2 °C) (Final = 20.0 °C) Partially Frozen   Thawed N/A

CONDITION OF ICE: Frozen   Partially Frozen   Thawed N/A

TYPE OF ICE: Real   Synthetic   No Ice  / Other:  

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other:  

Compliance Acceptance Criteria:


- 1) Chemistry: >0, ≤ 6°C, not frozen (NELAP) (if received after 24 hrs of sample collection, within 8 hours)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)
 

1 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C)	2 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C)
3 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C)	4 = (Observation = <u> </u> °C) (Corr. Factor = <u> </u> °C) (Final = <u> </u> °C)
- 4) UCMR3: 524.3: (Observation =   °C) (Corr. Factor =   °C) (Final =   °C)  
 (non-GLEC) 522: (Observation =   °C) (Corr. Factor =   °C) (Final =   °C)  
≤ 10°C if received within 48 hours of sample collection (not the same business day); ≤ 6°C if received after 48 hours of sample collection. Measure temperature for each method above.
- 5) LT2: Giardia / Cryptosporidium: < 20 °C, not frozen (received after 8 hours of sample collection)  
 E. Coli: < 10°C, not frozen (if received after 2 hours of sample collection)  
 Giardia/Crypto: (Observation =   °C) (Corr. Factor =   °C) (Final =   °C)  
 E. Coli: (Observation =   °C) (Corr. Factor =   °C) (Final =   °C)
- 6) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

Note: If samples are out of temperature range, let the ASMP know. ASMP will determine whether to proceed with analysis or not.

RECEIVED BY: <u>MAV</u>	SIGNATURE: <u>MAV</u>	COMPANY TITLE: <u>Eurofins Eaton Analytical</u>	DATE: <u>3-27-17</u>	TIME: <u>12:30</u>
-------------------------	-----------------------	---	----------------------	--------------------

Mohave Valley Elementary District	Mohave Valley Junior High School	1001	Building A	Admin/Lib/Clsrm	1998	P O Box 5070	Mohave Valley	86446-	Mohave	1
Mohave Valley Elementary District	Mohave Valley Junior High School	1002	Building B	gym/caf/clsrn	1998	P O Box 5070	Mohave Valley	86446-	Mohave	1
Mohave Valley Elementary District	Mohave Valley Junior High School	1003	Building C	resident/Sec.	1998	P O Box 5070	Mohave Valley	86446-	Mohave	1
<b>Total Containers</b>										<b>3</b>



Tel: (626) 386-1100  
Fax: (626) 386-1101  
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality  
David Burchard  
1110 West Washington Street  
Phoenix, AZ 85007

---

Tel: (626) 386-1100  
Fax: (626) 386-1101  
1 800 566 LABS (1 800 566 5227)

Laboratory Hits  
Report: 647443

**Arizona Department of Environmental Quality**  
David Burchard  
1110 West Washington Street  
Phoenix, AZ 85007

Samples Received on:  
03/27/2017 1230

---

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
----------	---------	-----------	--------	-------------	-------	-----

---

Tel: (626) 386-1100  
 Fax: (626) 386-1101  
 1 800 566 LABS (1 800 566 5227)

Laboratory Data  
 Report: 647443

**Arizona Department of Environmental Quality**  
 David Burchard  
 1110 West Washington Street  
 Phoenix, AZ 85007

Samples Received on:  
 03/27/2017 1230

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
<b>001583 (201703270458)</b>						<b>Sampled on 03/20/2017 1200</b>			
Sample Type: Kitchen Sink									
Facility ID: Building A Admin Clsrm 1003									
Sample Point ID: Kitchen									
<b>EPA 200.8 - ICPMS Metals</b>									
03/28/17	03/29/17 12:43	981512	981980	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<b>001561 (201703270459)</b>						<b>Sampled on 03/20/2017 1200</b>			
Sample Type: Drinking Fountain									
Facility ID: Building B, Gym Cafe Clsrm1002									
Sample Point ID: Hallway									
<b>EPA 200.8 - ICPMS Metals</b>									
03/28/17	03/29/17 12:44	981512	981980	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<b>001565 (201703270460)</b>						<b>Sampled on 03/20/2017 1200</b>			
Sample Type: Drinking Fountain									
Facility ID: Building A Admin Lib Clsrm 100									
<b>EPA 200.8 - ICPMS Metals</b>									
03/28/17	03/29/17 12:46	981512	981980	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1

Rounding on totals after summation.  
 (c) - indicates calculated results

Tel: (626) 386-1100  
Fax: (626) 386-1101  
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality

---

**ICPMS Metals****Prep Batch: 981512 Analytical Batch: 981980**

201703270458	001583
201703270459	001561
201703270460	001565

**Analysis Date: 03/29/2017**

Analyzed by: DTN  
Analyzed by: DTN  
Analyzed by: DTN

Tel: (626) 386-1100  
Fax: (626) 386-1101  
1 800 566 LABS (1 800 566 5227)

Arizona Department of Environmental Quality

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
<b>ICPMS Metals by EPA 200.8</b>									
<b>Analytical Batch: 981980</b>					<b>Analysis Date: 03/29/2017</b>				
LCS1	Lead Total ICAP/MS		20	19.6	ug/L	98	(85-115)		
LCS2	Lead Total ICAP/MS		20	19.9	ug/L	99	(85-115)	20	1.0
MBLK	Lead Total ICAP/MS			<0.25	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.369	ug/L	74	(50-150)		
MS_201703270571	Lead Total ICAP/MS	ND	20	18.2	ug/L	91	(70-130)		
MSD_201703270571	Lead Total ICAP/MS	ND	20	18.1	ug/L	90	(70-130)	20	0.55

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.