

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)



AT-1807

Laboratory Report

for

Arizona Department of Environmental Quality
1110 West Washington Street
Phoenix, AZ 85007
Attention: David Burchard

Date of Issue
06/28/2017


Eurofins Eaton
Analytical, Inc.

TDF: Thomas.D.French
Project Manager

Report:667980
Project:KAYENTA-UD
ADHS License #:AZ0778
Group:Monument Valley High School
PO#:PO#: ADEQ16-116686:3



* Accredited in accordance with TNI 2009 and ISO/IEC 17025:2005.

* Laboratory certifies that the test results meet all **TNI 2009 and ISO/IEC 17025:2005** requirements unless noted under the individual analysis.

* Following the cover page are State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.

* Test results relate only to the sample(s) tested.

* This report shall not be reproduced except in full, without the written approval of the laboratory.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Mississippi	Certified
Arizona	AZ0778	Montana	Cert 0035
Arkansas	Certified	Nebraska	Certified
California-Monrovia-ELAP	2813	Nevada	CA00006-2016
California-Colton- ELAP	2812	New Hampshire *	2959
California-Folsom- ELAP	2820	New Jersey *	CA 008
California-Fresno- ELAP	2966	New Mexico	Certified
Colorado	Certified	New York *	11320
Connecticut	PH-0107	North Carolina	06701
Delaware	CA 006	North Dakota	R-009
Florida *	E871024	Oregon (Primary AB) *	ORELAP 4034
Georgia	947	Pennsylvania *	68-565
Guam	16-003r	Puerto Rico	Certified
Hawaii	Certified	Rhode Island	LAO00326
Idaho	Certified	South Carolina	87016
Illinois *	200033	South Dakota	Certified
Indiana	C-CA-01	Tennessee	TN02839
Kansas *	E-10268	Texas *	T104704230-15-9
Kentucky	90107	Utah *	CA000062016-10
Louisiana *	LA16003	Vermont	VT0114
Maine	CA0006	Virginia *	460260
Maryland	224	Washington	C838
Commonwealth of Northern Marianas Is.	MP0004	Wyoming	Certified
Massachusetts	M-CA006	EPA Region 5	Certified
Michigan	9906	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO 17025 as verified by the ANSI-ASQ National Accreditation Board/ANAB.

Refer to Certificate and scope of accreditation (AT 1807) found at: <http://www.eatonanalytical.com>

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
1,4-Dioxane	EPA 522	x		x
2,3,7,8-TCDD	Modified EPA 1613B	x		x
Acrylamide	In House Method (2440)	x		x
Alkalinity	SM 2320B	x	x	x
Ammonia	EPA 350.1		x	x
Ammonia	SM 4500-NH3 H		x	x
Anions and DBPs by IC	EPA 300.0	x	x	x
Anions and DBPs by IC	EPA 300.1	x		x
Asbestos	EPA 100.2	x	x	
Bicarbonate Alkalinity as HCO3	SM 2320B	x	x	x
BOD / CBOD	SM 5210B		x	x
Bromate	In House Method (2447)	x		x
Carbamates	EPA 531.2	x		x
Carbonate as CO3	SM 2330B	x	x	x
Carbonyls	EPA 556	x		x
COD	EPA 410.4 / SM 5220D		x	
Chloramines	SM 4500-CL G	x	x	x
Chlorinated Acids	EPA 515.4	x		x
Chlorinated Acids	EPA 555	x		x
Chlorine Dioxide	SM 4500-CLO2 D	x		x
Chlorine -Total/Free/ Combined Residual	SM 4500-Cl G	x	x	x
Conductivity	EPA 120.1		x	
Conductivity	SM 2510B	x	x	x
Corrosivity (Langelier Index)	SM 2330B	x		x
Cryptosporidium	EPA 1623	x		x
Cyanide, Amenable	SM 4500-CN G	x	x	
Cyanide, Free	SM 4500CN F	x	x	x
Cyanide, Total	EPA 335.4	x	x	x
Cyanogen Chloride (screen)	In House Method (2470)	x		x
Diquat and Paraquat	EPA 549.2	x		x
DBP/HAA	SM 6251B	x		x
Dissolved Oxygen	SM 4500-O G		x	x
DOC	SM 5310C	x		x
E. Coli (MTF/EC+MUG)		x		x
E. Coli	CFR 141.21(f)(6)(i)	x		x
E. Coli	SM 9223		x	
E. Coli (Enumeration)	SM 9221B.1/ SM 9221F	x		x
E. Coli (Enumeration)	SM 9223B	x		x
EDB/DCBP	EPA 504.1	x		
EDB/DCBP and DBP	EPA 551.1	x		x
EDTA and NTA	In House Method (2454)	x		x
Endothall	EPA 548.1	x		x
Endothall	In-house Method (2445)	x		x
Enterococci	SM 9230B	x	x	
Fecal Coliform	SM 9221 E (MTF/EC)	x		
Fecal Coliform	SM 9221C, E (MTF/EC)		x	
Fecal Coliform (Enumeration)	SM 9221E (MTF/EC)	x		x
Fecal Coliform with Chlorine Present	SM 9221E		x	
Fecal Streptococci	SM 9230B	x	x	
Fluoride	SM 4500-F C	x	x	x
Giardia	EPA 1623	x		x
Glyphosate	EPA 547	x		x
Gross Alpha/Beta	EPA 900.0	x	x	x
Gross Alpha Coprecipitation	SM 7110 C	x	x	x
Hardness	SM 2340B	x	x	x
Heterotrophic Bacteria	In House Method (2439)	x		x
Heterotrophic Bacteria	SM 9215 B	x		x
Hexavalent Chromium	EPA 218.6	x	x	x

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
Hexavalent Chromium	EPA 218.7	x		x
Hexavalent Chromium	SM 3500-Cr B		x	
Hormones	EPA 539	x		x
Hydroxide as OH Calc.	SM 2330B	x		x
Kjeldahl Nitrogen	EPA 351.2		x	
Legionella	CDC Legionella	x		x
Mercury	EPA 245.1	x	x	x
Metals	EPA 200.7 / 200.8	x	x	x
Microcystin LR	ELISA (2360)	x		x
NDMA	EPA 521	x		x
NDMA	TQ In house method based on EPA 521 (2425)	x		x
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x	x
OCL, Pesticides/PCB	EPA 505	x		x
Ortho Phosphate	EPA 365.1	x	x	x
Ortho Phosphate	SM 4500P E			x
Ortho Phosphorous	SM 4500P E	x		
Oxyhalides Disinfection Byproducts	EPA 317.0	x		x
Perchlorate	EPA 331.0	x		x
Perchlorate (low and high)	EPA 314.0	x		x
Perfluorinated Alkyl Acids	EPA 537	x		x
pH	EPA 150.1	x		
pH	SM 4500-H+B	x	x	x
Phenylurea Pesticides/ Herbicides	In House Method, based on EPA 532 (2448)	x		x
Pseudomonas	IDEXX Pseudalert (2461)	x		x
Radium-226	GA Institute of Tech	x		x
Radium-228	GA Institute of Tech	x		x
Radon-222	SM 7500RN	x		x
Residue, Filterable	SM 2540C	x	x	x
Residue, Non-filterable	SM 2540D		x	
Residue, Total	SM 2540B		x	x
Residue, Volatile	EPA 160.4		x	
Semi-VOC	EPA 525.2	x		x
Semi-VOC	EPA 625		x	x
Silica	SM 4500-Si D	x	x	
Silica	SM 4500-SiO2 C	x	x	
Sulfide	SM 4500-S ⁻ D		x	
Sulfite	SM 4500-SO ³ B	x	x	x
Surfactants	SM 5540C	x	x	x
Taste and Odor Analytes	SM 6040E	x		x
Total Coliform (P/A)	SM 9221 A, B	x		x
Total Coliform (Enumeration)	SM 9221 A, B, C	x		x
Total Coliform / E. coli	Colisure SM 9223	x		x
Total Coliform	SM 9221B		x	
Total Coliform with Chlorine Present	SM 9221B		x	
Total Coliform / E.coli (P/A and Enumeration)	SM 9223	x		x
TOC	SM 5310C	x	x	x
TOX	SM 5320B		x	
Total Phenols	EPA 420.1		x	
Total Phenols	EPA 420.4	x	x	x
Total Phosphorous	SM 4500 P E		x	
Turbidity	EPA 180.1	x	x	x
Turbidity	SM 2130B	x	x	
Uranium by ICP/MS	EPA 200.8	x		x
UV 254	SM 5910B	x		
VOC	EPA 524.2/EPA 524.3	x		x
VOC	EPA 624		x	x
VOC	EPA SW 846 8260	x		x
VOC	In House Method (2411)	x		x
Yeast and Mold	SM 9610	x		x

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 667980
Project: KAYENTA-UD
Sample Group: Monument Valley High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Chris Claw

The following samples were received from you on **June 21, 2017 at 1035**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, Inc..

Sample #	Sample ID	Sample Date
<u>201706210777</u>	4001 A	06/16/2017 0826
	Sample Type: Not Provided Facility ID: Admin Sample Point ID: Not Provided	
	@ICPMS	Freight - Outbound
		Freight - Return
<u>201706210778</u>	4001 B	06/16/2017 0829
	Sample Type: Not Provided Facility ID: Admin Sample Point ID: Not Provided	
	@ICPMS	
<u>201706210779</u>	4002 A	06/16/2017 0830
	Sample Type: Drinking Fountain Facility ID: Counsel Sample Point ID: Ofc	
	@ICPMS	
<u>201706210780</u>	4002 B	06/16/2017 0833
	Sample Type: Drinking Fountain Facility ID: Counsel Sample Point ID: Ofc	
	@ICPMS	
<u>201706210781</u>	4003	06/16/2017 0840
	Sample Type: Drinking Fountain Facility ID: 2nd Floor Hall Sample Point ID: Hallway	
	@ICPMS	
<u>201706210782</u>	4004 A	06/16/2017 0843
	Sample Type: Not Provided Facility ID: Gym Lobby Sample Point ID: Not Provided	
	@ICPMS	
<u>201706210783</u>	4004 B	06/16/2017 0845
	Sample Type: Drinking Fountain Facility ID: Boys Locker Sample Point ID: Not Provided	
	@ICPMS	
<u>201706210784</u>	4004 C	06/16/2017 0000

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 667980
Project: KAYENTA-UD
Sample Group: Monument Valley High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Chris Claw

The following samples were received from you on **June 21, 2017 at 1035**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, Inc..

Sample #	Sample ID	Sample Date
	Sample Type: Not Provided Facility ID: Mezz Sample Point ID: Not Provided @ICPMS	
201706210785	4005	06/16/2017 0850
	Sample Type: Drinking Fountain Facility ID: Gym Lobby Sample Point ID: Not Provided @ICPMS	
201706210787	4006 A	06/16/2017 0905
	Sample Type: Drinking Fountain Facility ID: Courtyard Sample Point ID: Not Provided @ICPMS	
201706210788	4006 B	06/16/2017 0907
	Sample Type: Drinking Fountain Facility ID: Courtyard Sample Point ID: Not Provided @ICPMS	
201706210789	4007 A	06/16/2017 0915
	Sample Type: Drinking Fountain Facility ID: West Wing Sample Point ID: Not Provided @ICPMS	
201706210790	4007 B	06/16/2017 0916
	Sample Type: Drinking Fountain Facility ID: West Wing Sample Point ID: Not Provided @ICPMS	
201706210791	4008 A	06/16/2017 0920
	Sample Type: Drinking Fountain Facility ID: East Wing Sample Point ID: Hallway @ICPMS	
201706210792	4008 B	06/16/2017 0921
	Sample Type: Drinking Fountain Facility ID: East Wing Sample Point ID: Hallway	

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
 1110 West Washington Street
 Phoenix, AZ 85007

Attn: David Burchard
 Phone: (602) 771-4298

Client ID: ADEQ-LEAD
 Folder #: 667980
 Project: KAYENTA-UD
 Sample Group: Monument Valley High School

Project Manager: Thomas.D.French
 Phone: (480) 778-1558
 PO #: ADEQ16-116686:3
 Sampler: Chris Claw

The following samples were received from you on **June 21, 2017 at 1035**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, Inc..

Sample #	Sample ID	Sample Date
	@ICPMS	
201706210793	4010	06/16/2017 0925
	Sample Type: Sink Facility ID: Counseling Sample Point ID: Ofc	
	@ICPMS	
201706210794	4012 A	06/16/2017 0935
	Sample Type: Sink Facility ID: Storage Sample Point ID: Not Provided	
	@ICPMS	
201706210795	4012 B	06/16/2017 0936
	Sample Type: Sink Facility ID: Storage Sample Point ID: Not Provided	
	@ICPMS	
201706210796	4013 A	06/16/2017 0936
	Sample Type: Sink Facility ID: Concession Sample Point ID: Not Provided	
	@ICPMS	
201706210797	4013 B	06/16/2017 0936
	Sample Type: Sink Facility ID: Concession Sample Point ID: Not Provided	
	@ICPMS	
201706210798	Nash Center 1014	06/16/2017 0938
	Sample Type: Drinking Fountain Facility ID: Nash Center Sample Point ID: Not Provided	
	@ICPMS	
201706210799	1015	06/16/2017 0939
	Sample Type: Drinking Fountain Facility ID: SAC Sample Point ID: Hallway	
	@ICPMS	
201706210800	1016	06/16/2017 0941

Acknowledgement of Samples Received

Addr: **Arizona Department of Environmental Quality**
1110 West Washington Street
Phoenix, AZ 85007

Attn: David Burchard
Phone: (602) 771-4298

Client ID: ADEQ-LEAD
Folder #: 667980
Project: KAYENTA-UD
Sample Group: Monument Valley High School

Project Manager: Thomas.D.French
Phone: (480) 778-1558
PO #: ADEQ16-116686:3
Sampler: Chris Claw

The following samples were received from you on **June 21, 2017 at 1035**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, Inc..

Sample #	Sample ID	Sample Date
	Sample Type: Drinking Fountain Facility ID: SAC Sample Point ID: Hallway @ICPMS	
<u>201706210801</u>	1017	06/16/2017 0945
	Sample Type: Drinking Fountain Facility ID: Sci Wing Sample Point ID: Hallway @ICPMS	
<u>201706210802</u>	New SAC Addition 1019	06/16/2017 0950
	Sample Type: Sink Facility ID: Classroom Sample Point ID: Not Provided @ICPMS	
<u>201706210832</u>	1020	06/16/2017 0951
	Sample Type: Sink Facility ID: Nurses Ofc Sample Point ID: Not Provided @ICPMS	

Test Description


@ICPMS -- ICPMS Metals

667980


Collection Log
for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MHS
Building (name/number)	Admin Bldg.
Type of Fixture (tap, drinking fountain etc.)	Drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4001
Date of Collection	6/16
Time of Collection	8:26 am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:	Rec'd 2 bottles same ID for 4001 NO time or date on label 4001 Admin
-------------------------	--

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Admin
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	Kitchen
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4001
Date of Collection	6/16
Time of Collection	829am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles, same ID. for 4001
No time or date on label
4001 Admin

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	WWHS
Building (name/number)	Counsel
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	etc
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4002
Date of Collection	6/16
Time of Collection	8:30 am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

rec'd 2 bottles
same ID
no time or date on label
4002 Counsel

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	ICUSD
School Name	MVHS
Building (name/number)	Counsel
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	etc
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4002
Date of Collection	6/1/6
Time of Collection	833am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles
same ID.
No time or date on label
4002 Counsel

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	2nd floor Hall
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4003
Date of Collection	6/16
Time of Collection	8:40am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

classroom has no sinks or water.

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Boys Locker Rm
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4004
Date of Collection	6/16
Time of Collection	843
Printed Name of Sample Collector	Chris Clark
Signature Sample Collector	

Notes Sample collector:

Rec'd 1 bottle with Gym lobby as site name. (4004)
Rec'd 1 bottle with Boys Locker → Rec'd 2 COCs with same info. (4004)
Rec'd 1 bottle with Mezz as site name. (4004)
COC doesn't match bottle.
no time or date on label

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-26-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KMSD
School Name	WWHS
Building (name/number)	Boys Locker Rm
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (Write this number on the sample container and on this sheet)	4004
Date of Collection	6/16
Time of Collection	8:45
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

rec'd 1 bottle with gym lobby as site name (4004)

Notes Sample collector: rec'd 1 bottle with Boys Locker as site (4004)
description → rec'd 2 cocs with same info
rec'd 1 bottle with Mezz as site name (4004)
rmy too coc doesn't match.
No time or date on label

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and no flushing was done prior to sampling

Name of School District	KUSD
School Name	MWHS
Building (name/number)	Gym Lobby
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4005
Date of Collection	6/16
Time of Collection	8:50am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector: Rec'd 2 COCs only 1 bottle.
with the same info.
No time or date on label 4005 Gym Lobby

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Gym Lobby
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (Write this number on the sample container and on this sheet)	4005
Date of Collection	6/16
Time of Collection	853 am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector: rec'd 1 bottle and 2 cocs with same description.
no time or date on label 4005 Gym lobby

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.


- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MWHS
Building (name/number)	Courtyard / Ctr.
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4006
Date of Collection	6/16
Time of Collection	9:05 am
Printed Name of Sample Collector	Chris Chen
Signature Sample Collector	

Notes Sample collector:

rec'd 2 bottles same information

NO time or date on label 4006 Courtyard

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-26-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.


- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Court yard
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4006
Date of Collection	6/16
Time of Collection	6:07am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles same info.

No time or date on label 4006 courtyard

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 10:55
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	West wing
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4007
Date of Collection	6/16
Time of Collection	9:45 am
Printed Name of Sample Collector	Chavez
Signature Sample Collector	

Notes Sample collector: Rec'd 2 bottles same info. 4007 west wing
No time or date on label

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KMSD
School Name	MWH
Building (name/number)	West wing
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hall way
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4007
Date of Collection	6/16
Time of Collection	9:00am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:	Rec'd 2 bottles same info. 4007 west wing No time or date on label
-------------------------	--

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	WVHS
Building (name/number)	East wing
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4008
Date of Collection	6/16
Time of Collection	9:20 am
Printed Name of Sample Collector	Chris Chen
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles same info.
No time or date on label | 4008 East wing

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.


- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	WHS
Building (name/number)	East wing
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4008
Date of Collection	6/16
Time of Collection	9:21am
Printed Name of Sample Collector	Chris Chen
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles same info

No time or date on label | 4008 East wing

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 10:55
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MWH3
Building (name/number)	Counseling
Type of Fixture (tap, drinking fountain etc.)	sink
Location of Fixture (example, room number)	etc
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4010
Date of Collection	6/16
Time of Collection	9:25 am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	HVHS
Building (name/number)	Storage
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4012
Date of Collection	6/18
Time of Collection	14:35
Printed Name of Sample Collector	Chris Chen
Signature Sample Collector	

Notes Sample collector:	rec'd 2 bottles same info. No time or date on label 4012 storage
-------------------------	---

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Storage
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4012
Date of Collection	6/16
Time of Collection	9:35 am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:

Rec'd 2 bottles same info
No time or date on label 4012 storage

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 10:35
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that water had not been used at this tap or other taps in the area for 6 hours and no flushing was done prior to sampling

Name of School District	KUSD
School Name	HWHS
Building (name/number)	Concession
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4013
Date of Collection	6/16
Time of Collection	9:36
Printed Name of Sample Collector	Theresa Chan
Signature Sample Collector	

Notes Sample collector:	rec'd 2 bottles same info. No time or date on label/ 4013 Concession
-------------------------	---

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	Concession
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4013
Date of Collection	6/16
Time of Collection	9:36am
Printed Name of Sample Collector	Chris Chan
Signature Sample Collector	

Notes Sample collector:

rec'd 2 bottles same info.
No time or date on label 4013 concession

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	NMHS
Building (name/number)	Nash Center
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	Nash Center (1014)
Date of Collection	6/14
Time of Collection	9:38 am
Printed Name of Sample Collector	Chris Cla
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	RUSD
School Name	MWHS
Building (name/number)	SAC
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	1015
Date of Collection	6/16
Time of Collection	9:39
Printed Name of Sample Collector	Chris Clew
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	LENSD
School Name	WWHS
Building (name/number)	SAC
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	1016
Date of Collection	6/16
Time of Collection	9:10am
Printed Name of Sample Collector	Theresa Clew
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-20-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MWHS
Building (name/number)	Sci wing
Type of Fixture (tap, drinking fountain etc.)	drinking fountain
Location of Fixture (example, room number)	Hallway
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	1017
Date of Collection	6/16
Time of Collection	9:45 am
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KMSD
School Name	MMHS
Building (name/number)	Classroom
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	New SAC Addition (1019)
Date of Collection	6/16
Time of Collection	9:50
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log


for experienced sample collectors

Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☒ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KASH
School Name	WWHS
Building (name/number)	Nurses ofc
Type of Fixture (tap, drinking fountain etc.)	Sink
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	1020
Date of Collection	6/16
Time of Collection	9:51 am
Printed Name of Sample Collector	Chris Clew
Signature Sample Collector	

Notes Sample collector:

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-19 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	MVHS
Building (name/number)	
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4003
Date of Collection	6/16
Time of Collection	
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

only one water fountain
2nd floor classrooms

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-12 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	WCHSD
School Name	MVHS
Building (name/number)	
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	4010
Date of Collection	6/16
Time of Collection	
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

*no other water fountains or drinking
fountains*

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	LUSD
School Name	MWH3
Building (name/number)	
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	Alternative Ed. (1018)
Date of Collection	6/16
Time of Collection	
Printed Name of Sample Collector	Olivia Claw
Signature Sample Collector	

Notes Sample collector:

No Alternative Ed @ H.S.

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____


These samples were collected for lead screening purposes only and cannot be used for compliance.

Collection Log

for experienced sample collectors


Complete copy of this form for each sample collected. Your drinking water sample cannot be tested unless a copy of this form is shipped to the lab with each sample. In addition, the sample cannot be tested unless the Sample Identification Number matches the number on the sample container label.

- ☐ Check this box to confirm that **water had not been used at this tap or other taps in the area for 6 hours** and **no** flushing was done prior to sampling

Name of School District	KUSD
School Name	NWHS
Building (name/number)	
Type of Fixture (tap, drinking fountain etc.)	
Location of Fixture (example, room number)	
Sample Identification Number (<u>Write this number on the sample container and on this sheet</u>)	(1018) Alternative Ed.
Date of Collection	6/16
Time of Collection	
Printed Name of Sample Collector	Chris Claw
Signature Sample Collector	

Notes Sample collector:

~~from~~ No Alternative Ed @ High Sch.

For Lab use only	
Analyze this drinking water sample for lead	
Date and Time Lab received	6-21-17 1035
Signature	
Notes:	

For relinquishing samples upon delivery to labs only

Relinquished date and signature _____

These samples were collected for lead screening purposes only and cannot be used for compliance.

TRK# 7324 9820 4234

eurolins | Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 667980

SAMPLES REC'D DAY OF COLLECTION? ☐

SAMPLE TEMP RECEIVED:

IR Gun ID = 589A (Observation = 31.2 °C) (Corr.Factor = -2 °C) (Final = 31 °C)

CONDITION OF ICE: Frozen _____ Partially Frozen _____ Thawed _____ N/A X
 TYPE OF ICE: Real _____ Synthetic X No Ice _____ ON TRAC _____ DCS _____

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: ON TRAC

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: <10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: <10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

4) UCMR3: 524.3: (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)
 (non-GLEC)

522: (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

≤10°C if received within 48 hours of sample collection (not the same business day); ≤6°C if received after 48 hours of sample collection. Measure temperature for each method above.

5) LT2: Giardia /Cryptosporidium: <20 °C, not frozen (received after 8 hours of sample collection)

E. Coli: <10°C, not frozen (if received after 2 hours of sample collection)

Giardia/Crypto: (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

E. Coli: (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

6) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SIGNATURE

RECEIVED BY:

CHRIS GREEN

TIME

DATE

COMPANY/TITLE

Eutrolins Eaton Analytical

6.21.17


1035

Kayenta Unified District	Monument Valley High School	1001	✓ 4001	Classroom, Gym, Admin	1976	Navajo	2
Kayenta Unified District	Monument Valley High School	1002	✓ 4002	Admin, Counsel	1982	Navajo	2
Kayenta Unified District	Monument Valley High School	1003	✓ 4003	<i>only</i> 2nd Floor Classrooms <i>drinking fountain</i>	1978	Navajo	2
Kayenta Unified District	Monument Valley High School	1004	✓ 4004	Boys Laker, Mezz	1977	Navajo	2
Kayenta Unified District	Monument Valley High School	1005	✓ 4005	Gym Lobby, RR	1976	Navajo	2
Kayenta Unified District	Monument Valley High School	1006	✓ 4006	Courtyard, MP	1984	Navajo	2
Kayenta Unified District	Monument Valley High School	1007	✓ 4007	West Wing, Classroom	1983	Navajo	2
Kayenta Unified District	Monument Valley High School	1008	✓ 4008	East Wing, Classroom	1983	Navajo	2
Kayenta Unified District	Monument Valley High School	1010	✓ 4010	Counseling	1985	Navajo	2
Kayenta Unified District	Monument Valley High School	1012	✓ 4012	Field House, Storage	1985	Navajo	2

Kayenta Unified District	Monument Valley High School	1013	✓ 4013	RR, Concession	1986	Navajo	2
Kayenta Unified District	Monument Valley High School	1014	✓ Nash Center	Student Activity	1988	Navajo	1
Kayenta Unified District	Monument Valley High School	1015	✓ 1015	1015	2002	Navajo	1
Kayenta Unified District	Monument Valley High School	1016	✓ 1016	1016	2002	Navajo	1
Kayenta Unified District	Monument Valley High School	1017	✓ 1017	Science Wing	1999	Navajo	1
Kayenta Unified District	Monument Valley High School	1018	Alternative Education	(none) Alternative Education	1985	Navajo	2
Kayenta Unified District	Monument Valley High School	1019	✓ New SAC Addition	Classrooms, locker room	2005	Navajo	1
Kayenta Unified District	Monument Valley High School	1020	✓ 1020	Nurses Office	2008	Navajo	1

Total Containers

30



MVHS Classroom Plan

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/21/2017 1035

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
06/22/2017 19:47	Lead Total ICAP/MS	<u>4001 A</u>	0.59	15	ug/L	0.5
06/22/2017 19:47	Lead Total ICAP/MS	<u>4001 B</u>	1.5	15	ug/L	0.5
06/22/2017 19:48	Lead Total ICAP/MS	<u>4002 A</u>	0.76	15	ug/L	0.5
06/22/2017 19:53	Lead Total ICAP/MS	<u>4003</u>	0.86	15	ug/L	0.5
06/22/2017 20:05	Lead Total ICAP/MS	<u>4008 A</u>	0.67	15	ug/L	0.5
06/22/2017 20:09	Lead Total ICAP/MS	<u>4010</u>	1.3	15	ug/L	0.5
06/27/2017 12:39	Lead Total ICAP/MS	<u>4012 B</u>	5.6	15	ug/L	0.5
06/22/2017 20:11	Lead Total ICAP/MS	<u>4013 A</u>	11	15	ug/L	0.5
06/22/2017 20:12	Lead Total ICAP/MS	<u>4013 B</u>	0.51	15	ug/L	0.5
06/22/2017 20:20	Lead Total ICAP/MS	<u>1020</u>	0.84	15	ug/L	0.5

SUMMARY OF POSITIVE DATA ONLY

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/21/2017 1035

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
4001 A (201706210777)						Sampled on 06/16/2017 0826			
Sample Type: Not Provided Facility ID: Admin Sample Point ID: Not Provided									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:47	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	0.59	ug/L	0.5	1
4001 B (201706210778)						Sampled on 06/16/2017 0829			
Sample Type: Not Provided Facility ID: Admin Sample Point ID: Not Provided									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:47	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	1.5	ug/L	0.5	1
4002 A (201706210779)						Sampled on 06/16/2017 0830			
Sample Type: Drinking Fountain Facility ID: Counsel Sample Point ID: Ofc									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:48	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	0.76	ug/L	0.5	1
4002 B (201706210780)						Sampled on 06/16/2017 0833			
Sample Type: Drinking Fountain Facility ID: Counsel Sample Point ID: Ofc									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:50	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
4003 (201706210781)						Sampled on 06/16/2017 0840			
Sample Type: Drinking Fountain Facility ID: 2nd Floor Hall Sample Point ID: Hallway									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:53	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	0.86	ug/L	0.5	1
4004 A (201706210782)						Sampled on 06/16/2017 0843			
Sample Type: Not Provided Facility ID: Gym Lobby Sample Point ID: Not Provided									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:54	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
4004 B (201706210783)						Sampled on 06/16/2017 0845			
Sample Type: Drinking Fountain Facility ID: Boys Locker Sample Point ID: Not Provided									
EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:54	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
4004 C (201706210784)						Sampled on 06/16/2017 0000			

Rounding on totals after summation.
(c) - indicates calculated results

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/21/2017 1035

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Not Provided Facility ID: Mezz Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:55	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4005 (201706210785)</u>						Sampled on 06/16/2017 0850			
Sample Type: Drinking Fountain Facility ID: Gym Lobby Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:56	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4006 A (201706210787)</u>						Sampled on 06/16/2017 0905			
Sample Type: Drinking Fountain Facility ID: Courtyard Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:57	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4006 B (201706210788)</u>						Sampled on 06/16/2017 0907			
Sample Type: Drinking Fountain Facility ID: Courtyard Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 19:58	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4007 A (201706210789)</u>						Sampled on 06/16/2017 0915			
Sample Type: Drinking Fountain Facility ID: West Wing Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:00	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4007 B (201706210790)</u>						Sampled on 06/16/2017 0916			
Sample Type: Drinking Fountain Facility ID: West Wing Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:01	1005257	1005362	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>4008 A (201706210791)</u>						Sampled on 06/16/2017 0920			
Sample Type: Drinking Fountain Facility ID: East Wing Sample Point ID: Hallway EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:05	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	0.67	ug/L	0.5	1
<u>4008 B (201706210792)</u>						Sampled on 06/16/2017 0921			

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/21/2017 1035

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: East Wing Sample Point ID: Hallway EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:07	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
4010 (201706210793)						Sampled on 06/16/2017 0925			
Sample Type: Sink Facility ID: Counseling Sample Point ID: Ofc EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:09	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	1.3	ug/L	0.5	1
4012 A (201706210794)						Sampled on 06/16/2017 0935			
Sample Type: Sink Facility ID: Storage Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:10	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
4012 B (201706210795)						Sampled on 06/16/2017 0936			
Sample Type: Sink Facility ID: Storage Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/27/17 12:39	1005257	1005629	(EPA 200.8)	Lead Total ICAP/MS	5.6	ug/L	0.5	1
4013 A (201706210796)						Sampled on 06/16/2017 0936			
Sample Type: Sink Facility ID: Concession Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:11	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	11	ug/L	0.5	1
4013 B (201706210797)						Sampled on 06/16/2017 0936			
Sample Type: Sink Facility ID: Concession Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:12	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	0.51	ug/L	0.5	1
Nash Center 1014 (201706210798)						Sampled on 06/16/2017 0938			
Sample Type: Drinking Fountain Facility ID: Nash Center Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:13	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
1015 (201706210799)						Sampled on 06/16/2017 0939			

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality
David Burchard
1110 West Washington Street
Phoenix, AZ 85007

Samples Received on:
06/21/2017 1035

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: Drinking Fountain Facility ID: SAC Sample Point ID: Hallway EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:13	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1016 (201706210800)</u>						Sampled on 06/16/2017 0941			
Sample Type: Drinking Fountain Facility ID: SAC Sample Point ID: Hallway EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:14	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1017 (201706210801)</u>						Sampled on 06/16/2017 0945			
Sample Type: Drinking Fountain Facility ID: Sci Wing Sample Point ID: Hallway EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:15	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>New SAC Addition 1019 (201706210802)</u>						Sampled on 06/16/2017 0950			
Sample Type: Sink Facility ID: Classroom Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:16	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	ND	ug/L	0.5	1
<u>1020 (201706210832)</u>						Sampled on 06/16/2017 0951			
Sample Type: Sink Facility ID: Nurses Ofc Sample Point ID: Not Provided EPA 200.8 - ICPMS Metals									
06/22/17	06/22/17 20:20	1005257	1005363	(EPA 200.8)	Lead Total ICAP/MS	0.84	ug/L	0.5	1

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory QC Summary

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality

ICPMS Metals

Prep Batch: 1005257 Analytical Batch: 1005362

201706210777	4001 A
201706210778	4001 B
201706210779	4002 A
201706210780	4002 B
201706210781	4003
201706210782	4004 A
201706210783	4004 B
201706210784	4004 C
201706210785	4005
201706210787	4006 A
201706210788	4006 B
201706210789	4007 A
201706210790	4007 B

Analysis Date: 06/22/2017

Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS

ICPMS Metals

Prep Batch: 1005257 Analytical Batch: 1005363

201706210791	4008 A
201706210792	4008 B
201706210793	4010
201706210794	4012 A
201706210796	4013 A
201706210797	4013 B
201706210798	Nash Center 1014
201706210799	1015
201706210800	1016
201706210801	1017
201706210802	New SAC Addition 1019
201706210832	1020

Analysis Date: 06/22/2017

Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS
Analyzed by: AZS

ICPMS Metals

Prep Batch: 1005257 Analytical Batch: 1005629

201706210795	4012 B
--------------	--------

Analysis Date: 06/27/2017

Analyzed by: NEM

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 667980
Project: KAYENTA-UD
Group: Monument Valley High School

Arizona Department of Environmental Quality

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
ICPMS Metals by EPA 200.8									
Analytical Batch: 1005362					Analysis Date: 06/22/2017				
LCS1	Lead Total ICAP/MS		20	19.7	ug/L	99	(85-115)		
LCS2	Lead Total ICAP/MS		20	19.6	ug/L	98	(85-115)	20	0.51
MBLK	Lead Total ICAP/MS			<0.25	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.551	ug/L	110	(50-150)		
MS_201706210734	Lead Total ICAP/MS	ND	20	23.3	ug/L	116	(70-130)		
MS2_201706210780	Lead Total ICAP/MS	ND	20	22.4	ug/L	112	(70-130)		
MSD_201706210734	Lead Total ICAP/MS	ND	20	22.9	ug/L	114	(70-130)	20	1.7
MSD2_201706210780	Lead Total ICAP/MS	ND	20	23.1	ug/L	116	(70-130)	20	2.6
ICPMS Metals by EPA 200.8									
Analytical Batch: 1005363					Analysis Date: 06/22/2017				
LCS1	Lead Total ICAP/MS		20	19.5	ug/L	98	(85-115)		
LCS2	Lead Total ICAP/MS		20	19.7	ug/L	99	(85-115)	20	1.0
MBLK	Lead Total ICAP/MS			<0.25	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.543	ug/L	109	(50-150)		
MS_201706210791	Lead Total ICAP/MS	0.67	20	23.8	ug/L	116	(70-130)		
MS2_201706210802	Lead Total ICAP/MS	ND	20	23.0	ug/L	115	(70-130)		
MSD_201706210791	Lead Total ICAP/MS	0.67	20	23.6	ug/L	114	(70-130)	20	0.84
MSD2_201706210802	Lead Total ICAP/MS	ND	20	22.7	ug/L	114	(70-130)	20	1.8
ICPMS Metals by EPA 200.8									
Analytical Batch: 1005629					Analysis Date: 06/27/2017				
LCS1	Lead Total ICAP/MS		20	20.0	ug/L	100	(85-115)		
LCS2	Lead Total ICAP/MS		20	21.4	ug/L	107	(85-115)	20	6.8
MBLK	Lead Total ICAP/MS			<0.25	ug/L				
MRL_CHK	Lead Total ICAP/MS		0.5	0.490	ug/L	98	(50-150)		
MS_201704100347	Lead Total ICAP/MS	ND	20	20.1	ug/L	100	(70-130)		
MS2_201706070535	Lead Total ICAP/MS	ND	20	19.8	ug/L	99	(70-130)		
MSD_201704100347	Lead Total ICAP/MS	ND	20	21.0	ug/L	105	(70-130)	20	4.4
MSD2_201706070535	Lead Total ICAP/MS	ND	20	20.0	ug/L	100	(70-130)	20	1.5

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.